

METERED-DOSE INHALERS • DRY POWDER INHALERS
NEBULISERS • SOFT MIST INHALERS • NASAL PRODUCTS





About Us

Driving Results for Over 75 Years

Copley Scientific, founded in 1946 and headquartered in Nottingham, UK, remains a family-owned and managed company. With a rich history spanning nearly eight decades, we have solidified our position as the leading global manufacturer of inhaler test equipment. Additionally, we are well-recognised as a reliable provider of high quality test instrumentation for other pharmaceutical dosage forms, including tablets, capsules, creams, ointments and powders.

We continue to work closely with industry groups and leading experts to bring relevant new products to market, with all equipment backed by expert training and support.

Committed to excellence, we aim to deliver exemplary service for an outstanding customer experience.

We deliver pharmaceutical testing equipment with the necessary accuracy and reproducibility hardwired into its design by adopting the same Quality by Design (QbD) principles that our customers rely on to control product performance. Continuous improvement is a core element of this approach and we strive to exceed the expectations of the industry, not only by enhancing equipment performance but also through unrivalled service.

These commitments are exemplified by our investment in the ISO 9001:2015 Quality

Management System for which we have certification to the latest standard for all aspects of our business, including equipment design.

Copley customers benefit from:

- High quality pharmaceutical testing equipment, designed, manufactured and tested in the UK
- Product support from our friendly and experienced technical support team
- · First-class training and education



Our vision is to help scientists around the world improve the quality of people's lives

Copyright

This edition of the Copley Scientific Limited brochure is copyright 2024. All rights reserved.

No portion of this brochure may be reproduced without the permission of Copley Scientific Limited. Copley Scientific Limited reserves the right to make changes without notice to any products herein to improve reliability or design.

Copley Scientific Limited does not assume any liability arising out of the application or use of

any product described herein. Neither does it convey any licence under its patent rights nor the rights of others.

Any third party documentation or organisation mentioned herein is referential only and implies no company or product endorsement from, or affiliation with, Copley Scientific Limited.

The Copley Promise





Innovative

Novel solutions that maximise understanding and productivity



Compliant

Certified to the standards defined by global regulators and pharmacopoeias



Trusted

Quality products with accuracy, robustness and reliability built-in

Contents

Introduction	2
About Us ISO 9001: 2015 Quality Management	2
System	2
The Copley Promise	3
Table of Contents	4
Orally Inhaled & Nasal Drug Products (OINDPs)	6
Orally Inhaled Drug Products	7
Metered-Dose Inhalers (MDIs)	7
Dry Powder Inhalers (DPIs)	8
Nebulisers	9
Soft Mist Inhalers (SMIs)	10
Nasal Drug Products	10
Applications of OINDPs	11
Organisations and their Roles	12
-	
Regulatory Bodies in the UK, European	
	12
Union, China, Japan and USA	12
	12 14
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy -	14
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias	
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy -	14
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.)	14 15 15
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopeia (USP)	14 15 15 15
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopoeia (USP) Chinese Pharmacopoeia (ChP)	14 15 15 15 16
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopeia (USP) Chinese Pharmacopoeia (ChP) Japanese Pharmacopoeia (JP) Device Safety, Quality and Efficacy - International Standards	14 15 15 15 16 16
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopoeia (USP) Chinese Pharmacopoeia (ChP) Japanese Pharmacopoeia (JP) Device Safety, Quality and Efficacy - International Standards Organisation (ISO)	14 15 15 15 16 16
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopoeia (USP) Chinese Pharmacopoeia (ChP) Japanese Pharmacopoeia (JP) Device Safety, Quality and Efficacy - International Standards Organisation (ISO) Expert Groups	14 15 15 15 16 16
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopoeia (USP) Chinese Pharmacopoeia (ChP) Japanese Pharmacopoeia (JP) Device Safety, Quality and Efficacy - International Standards Organisation (ISO)	14 15 15 15 16 16
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopoeia (USP) Chinese Pharmacopoeia (ChP) Japanese Pharmacopoeia (JP) Device Safety, Quality and Efficacy - International Standards Organisation (ISO) Expert Groups European Pharmaceutical Aerosol	14 15 15 15 16 16 16
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopeia (USP) Chinese Pharmacopoeia (ChP) Japanese Pharmacopoeia (JP) Device Safety, Quality and Efficacy - International Standards Organisation (ISO) Expert Groups European Pharmaceutical Aerosol Group (EPAG)	14 15 15 15 16 16
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopeia (USP) Chinese Pharmacopoeia (ChP) Japanese Pharmacopoeia (JP) Device Safety, Quality and Efficacy - International Standards Organisation (ISO) Expert Groups European Pharmaceutical Aerosol Group (EPAG) International Pharmaceutical Consortium on Regulation and	14 15 15 15 16 16 16
Union, China, Japan and USA International Regulation and Harmonisation Drug Safety, Quality and Efficacy - The Pharmacopoeias European Pharmacopoeia (Ph. Eur.) United States Pharmacopoeia (USP) Chinese Pharmacopoeia (ChP) Japanese Pharmacopoeia (JP) Device Safety, Quality and Efficacy - International Standards Organisation (ISO) Expert Groups European Pharmaceutical Aerosol Group (EPAG) International Pharmaceutical Consortium on Regulation and Science (IPAC-RS)	14 15 15 15 16 16 16

Delivered Dose Uniformity	18
Background	18
DDU Over the Entire Contents	19
Collection Devices for DDU Testing	20
For MDIs, DPIs, BAIs and SMIs	20
- Dose Collection	20
- Waste Shot Collection	24
For MDIs with Spacers/VHCs	25
For Nasal Sprays, Nasal Aerosols and Nasal Powders	26
- Dose Collection	26
- Waste Shot Collection	28
USP Monographs	30
Delivered Dose Uniformity Testing of:	
Metered Dose Inhalers (MDIs)	32
MDIs with a Spacer/VHC	40
Dry Powder Inhalers (DPIs)	46
Nebulisers	52
Soft Mist Inhalers (SMIs)	58
Nasal Sprays	64
Nasal Aerosols	68
Nasal Powders	72

Aerodynamic Particle Size Distribution	n 78
Background	78
An Introduction to Cascade Impaction	79
How Does a Cascade Impactor Work?	82
Types of Cascade Impactor	84
Next Generation Impactor (NGI)	84
Andersen Cascade Impactor (ACI)	90
Multi-Stage Liquid Impinger (MSLI)	96
Glass Twin Impinger (GTI)	100
Aerodynamic Particle Size	
Measurement of:	
Metered Dose Inhalers (MDIs)	104
MDIs with a Spacer/VHC	111
Dry Powder Inhalers (DPIs)	118
Nebulisers	124
Soft Mist Inhalers (SMIs)	130
Nasal Sprays	136
Nasal Aerosols	141
Nasal Powders	147

Ancillaries	154
Breathing Simulators	156
Breathing Simulator Model BRS 100i	158
Breathing Simulator Model BRS 200i	162
Breathing Simulator Model BRS 300i	167
Flow Controllers	172
Breath Actuation Controller BAC 100i	176
Critical Flow Controller TPK 100i	180
Flow Rate Measurement	184
Flow Rate Sensor FRS	186
Flow Meter DFM 2000	187
Vacuum Pumps	188
Low Capacity Pump LCP6	190
High Capacity Pump HCP6	191
Super Capacity Pump SCP6	192
Environmental Control	194
EnviroMate™	196
Anti-Static Grounding Kit	200
Electrostatic Eliminator	200
Digital Static Meter	200
NGI Cooler™	202
nhaler Testing Workstation™ ITW	204
Spare/Additional Tubing	207
Quick-Release Connectors	207
Glass Expansion Chambers	208
Mouthpiece & Nosepiece Adapters	211
Inhalytix [®]	214
Improving IVIVCs	222
Background	222
DDU and APSD Testing	226
Realistic Breathing Profiles	226
Mixing Inlet	227
Realistic Throat and Nasal Models	228
Alberta Idealised Throat (AIT)	228
Alberta Idealised Nasal Inlet (AINI)	230
mproving IVIVCs: Example Test	

Improving IVIVCs	222
Background	222
DDU and APSD Testing	226
Realistic Breathing Profiles	226
Mixing Inlet	227
Realistic Throat and Nasal Models	228
Alberta Idealised Throat (AIT)	228
Alberta Idealised Nasal Inlet (AINI)	230
mproving IVIVCs: Example Test System for Realistic DDU	231
mproving IVIVCs: Example Test System for Realistic APSD	233
Dissolution Testing	238
Facemask Testing	244
Face Models	245
Test Systems for Assessing Facemask Performance	246
Facemask Testing Apparatus (FMA) for MDIs with a Spacer/VHC	246
Facemask Testing Stand (FMS) For Nebulisers	250

		St. Married St.	
Qualification/Servicing & Training	310		
Qualification Services	312		
Impactor Qualification	312		
Stage and Components Mensuration	312		
Data Interpretation	313		
Impactor Performance Restoration In-House and On-Site Equipment	314		
Servicing and Calibration Qualification Tools and Documents	316		
IQ/OQ Documentation	317		
Qualification Tools	318		
Product Protection Plans	319		
Support	320		
Design Support	320		
Training Services Inhaler Testing Academy®	321 321		
Index	322		
COPLEY			

Morphology	254
Cold Freon® Effect	255
Spray Force Tester SFT 1000	256
Plume Temperature Tester PTT 1000	258

Special Applications	260
Abbreviated Impactor	
Measurement (AIM)	261
AIM in QC	261
AIM in R&D	262
Fast Screen Andersen (FSA)	263
Reduced NGI (rNGI)	265
Fast Screening Impactor (FSI)	266
Volume and Resistance	
Compensator VRC	268
Generic Drug Development	270
Fluticasone Propionate/Salmeterol	
Aerosols & Powders	271
Albuterol Inhalation Aerosols	274
Device Robustness/Inhaler Misuse	276
Patient Exhalation Simulator PES	276

Automation	278
Automated Shake, Fire and Flow Control for MDIs, Nasal Sprays	
and Aerosols	282
Vertus® III & Vertus III+	284
DecaVertus® III	290
Automated Drug Recovery for DDU Testing	
DUSA Shaker DTS 100i	294
Automated Cascade Impactor Preparation and Recovery	
Impactor Coater™ IC 200i	296
Gentle Rocker™ GR 200i	298
Impactor Genie™ IG 200i	300
Sample Preparation Unit SPU 200i	302
Impactor Cleaning System	306




Contents



Orally Inhaled & Nasal Drug Products (OINDPs)


The range of OINDPs available is broad, encompassing inhalers (metered-dose, dry powder and soft mist), nebulisers (jet, ultrasonic and vibrating mesh) and nasal sprays, aerosols and powders (aqueous-based, propellant-based and dry powder).



- A Metered-Dose
- D Soft Mist
- B Dry Powder
- E Nasal Spray
- C Nebuliser
- F Nasal Powder








Orally Inhaled Drug Products Metered-Dose Inhalers (MDIs)

MDIs use a propellant to deliver a fixed volume of liquid solution or suspension to the patient in the form of an aerosol.

They are small, inexpensive, convenient for the user and suitable for a wide range of drugs. However, the use of MDIs requires good coordination and technique to actuate the device. The actuation force needed means they are not always suitable for elderly or paediatric users. The use of breath-actuated MDIs or add-on devices such as spacers or valved holding chambers (VHCs) can help resolve these problems.



Conventional Pressurised

Comprises a pressurised canister containing the medication and propellant, together with a delivery device – normally a metering valve linked to an actuator. Pressing down on the canister releases the drug in the form of an aerosol cloud – this is then inhaled into the lungs.




Breath-Actuated

Senses the patient's inhalation through the actuator and synchronises dose delivery with it.



Spacers/VHCs

Add-on devices such as these reduce or eliminate
a) the need for coordination between actuation and inhalation and b) the cold
Freon® effect (see page 255), enhancing drug delivery.



Spacers/VHC: Coordinated v Uncoordinated use

Performance is optimal and directly comparable with a standard MDI if the patient inhales from the spacer/VHC as the device is actuated. This is called 'coordinated use'.


In contrast, the worst case scenario is if actuation coincides with exhalation, i.e. 'uncoordinated use'.

Dry Powder Inhalers (DPIs)

As the name suggests, with a DPI the medication comes in the form of a dry powder, rather than a liquid.

Typically, the active pharmaceutical ingredient(s) is mixed with a coarser excipient, such as lactose, to which it attaches. During aerosolisation the active is stripped from the carrier and inhaled whilst the carrier particles impact on the mouth and throat and are ingested.

However, their relatively high cost and reliance on inhalation strength and duration are potential drawbacks.



Passive


The majority of DPIs are passive devices, that is to say drug delivery is driven solely by the inspiration of the patient. There is no need to coordinate breathing with the actuation - the patient simply inhales deeply to access the drug.

Pre-Metered



The dose is pre-measured during manufacture (for example, blisters, capsules or similar cavities).

Unit Dose



The pre-measured dose in the form of a gelatine capsule or blister is loaded by the patient prior to use.

Device-Metered



The drug is contained in a reservoir within the device which measures each dose on actuation.



Some DPIs actively generate the aerosol, reducing dependence on patient inhalation, whilst simultaneously improving the accuracy and reproducibility of the delivered dose.

Such devices are normally termed 'active' DPIs and are particularly useful where the patient's own inspiration capability is compromised. Assistance normally comes in the form of pressurised/compressed air or through vibrations generated by a piezoelectric transducer.

Nebulisers

Nebulisers convert a liquid into aerosol droplets to produce a respirable cloud suitable for inhalation. They are widely used at home and in hospital and require little or no coordination for effective use. Nebulisers are normally loaded with the drug before each treatment and usually operate continuously once loaded.


The main advantage of nebulisers is that their use requires little or no coordination on the part of the patient. However, they tend to be cumbersome and require either compressed air or an electrical supply. Expense, inefficiency and inter-brand variability can also cause issues.

Ultrasonic



Use electricity to vibrate a piezoelectric crystal at high frequency. The resultant vibrations are transmitted to a reservoir containing the liquid drug formulation, creating a series of waves from which liquid droplets separate to form an aerosol.

Jet



Use a compressed air supply to atomise the liquid drug formulation to produce a fine mist using the Bernoulli principle.

Can be subdivided into three types depending on their output during exhalation.

Mesh



Uses the piezoelectric effect to vibrate a mesh at ultrasonic frequency which results in droplets being formed by fluid moving through holes in the mesh (holes either electroformed or laser drilled) to form a cloud prior to inhalation.

Standard

Constant output throughout the respiratory cycle.

Some jet and mesh nebulisers incorporate sensing devices to detect the patient's inspiration in order to provide breath-enhanced, breath-activated or breath-integrated systems, and there are smart versions that use Adaptive Aerosol Delivery technology to analyse the patient's breathing pattern to determine the timing of aerosol delivery during inhalation.

Breath-Enhanced

Continuous aerosolisation but provides higher output during inhalation.

Breath-Actuated

Aerosol produced only during inhalation.


Soft Mist Inhalers (SMIs)

Both MDIs and DPIs suffer from the same two inherent problems: low lung deposition (typically 5-20%) and dose variability (often due to patient difficulties in coordination or inspiration).

SMIs (often known as "Inhalation Metered Sprays" or "Aqueous Droplet Inhalers") actively aerosolise the liquid, forming a 'soft mist' to overcome these problems. These inhalers generally deliver a higher fine particle fraction than MDIs or DPIs. However, as with any multi-dose liquid system, microbial contamination can be a problem.

SMIs do not use a propellant to aerolise the liquid. Methods of aerosol generation include:

- (a) Forcing liquid through a nozzle
- (c) Thermal generation
- (b) Electrospraying
- (d) Vibration mesh



Nasal Drug Products

Like inhalers, nasal products can be liquid-, propellant- or powder-based. They are commonly multi-dose although unit dose devices are popular for delivering vaccines and pain relief.

Nasal Sprays



Mechanical metered-dose spray pumps are designed to deliver an accurate and consistent dose to the user.

Multi-dose spray pumps have dominated the nasal market and are widely available through a number of device manufacturers.

Unit-dose devices that deliver one or two shots (one per nostril), are usually based on the syringe principle.

Nasal Aerosols



Nasal aerosols are propellant-based and directly analogous to pressurised MDIs. An angled nosepiece or nozzle facilitates insertion into the nostril.

Nasal Powders



Available in both multi- and unit-dose formats, powder-based devices offer preservative-free delivery and can produce longer nasal retention times than liquids.

Powder-based nasal sprays are more ideal for peptides, hormones and antigens than liquid formulations and where high dose concentrations are required.

Applications of OINDPs

Pulmonary and nasal delivery offers a number of advantages compared to traditional oral and parenteral (subcutaneous injection) routes:

Directly targets the site of action

Rapid onset of drug action

Drugs effective in relatively low doses

Fewer side effects

Avoids first pass metabolism

Non-invasive administration


Such drugs include treatments for diverse applications such as diabetes, erectile dysfunction, migraine, osteoporosis and for vaccine delivery.

Orally Inhaled Drug Product Applications

Orally inhaled drugs are becoming increasingly popular as a means of delivering local or systemic therapy via the lungs.


Local Treatment

To treat lung diseases such as asthma and chronic obstructive pulmonary disease (COPD), and to deliver locally acting drugs such as antibiotics and antivirals directly to the lungs to curb infection.



Systemic Treatment

Considerable research and development has been devoted to delivering new drugs into the systemic circulation via the inhaled route - no doubt attracted by the large surface area and easy air/blood interface provided by the respiratory system.




Nasal Drug Product Applications

Traditionally, nasal preparations have been used for the local administration of antihistamines, decongestants and steroids in order to alleviate cold or allergy symptoms and nasal congestion.

More recently attention has focused on two other areas:


Systemic Circulation

There is potential for rapid drug absorption into the systemic circulation via the turbinates. This route is already in use in a number of areas, e.g. migraine and pain relief, osteoporosis, vaccines.



Central Nervous System

The potential of "Nose to Brain" entry to the central nervous system is presented by the olfactory region at the top of the nasal cavity for the treatment of, for example, diseases of aging such as Alzheimer's disease.



Organisations and their Roles

The ultimate responsibility for the safety, quality and efficacy of medicines and medical devices lies with the various national regulatory bodies designated to safeguard public health.

Regulatory Bodies in the UK, European Union, China, Japan and USA

At present, there are no worldwide standards that are specifically applicable to OINDPs.

In the European Union, the responsibility for the regulation of medicines and medical devices lies with the European Medicines Agency (EMA) in the form of the Committee for Medicinal Products for Human Use (CHMP).

The EMA was set up in 1995 to harmonise the work of existing national regulatory bodies in Europe.

The main guidance from the EMA relating to OINDPs is contained in two guidelines:

- CPMP (2006), "Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products"
- CPMP (2009), "Guideline on the requirements for clinical documentation for orally inhaled products (OIP) including the requirements for demonstration of therapeutic equivalence between two inhaled products for use in the treatment of asthma and chronic obstructive pulmonary disease (COPD) in adults and for use in the treatment of asthma in children and adolescents"

These guidelines give a comprehensive list of the parameters that are critical to the safety, quality and efficacy of the final product dependent on the specific type of inhaled or nasal preparation concerned.

From 1 January 2024, developers of new medicines can now submit applications via the UK's Medicines and Healthcare products Regulatory Agency (MHRA) new International Recognition Procedure (IRP). The IRP was developed by the MHRA following the UK's departure from the European Union. The EMA and Food and Drug Administration (FDA) are two of a number of Reference Regulators (RR) to the MHRA, the IRP being open to applicants that have already received an authorisation from an RR.

A similar regulatory function is provided by the National Medical Products Administration (NMPA) in China and the Ministry of Health, Labour and Welfare (MHLW) in Japan. The PMDA (Pharmaceuticals and Medical Devices Agency) is the main agency working alongside the MHLW.

In the USA, the regulatory function is performed by the Food and Drug Administration (FDA) through two centres, the Center for Drug Evaluation and Research (CDER) in respect of medicines and the Center for Devices and Radiologic Health (CDRH) in respect of medical devices.

The relevant current thinking from the FDA is reflected in the following regulatory Guidelines for Industry:

- CDER (1998), "Metered-Dose Inhaler (MDI) and Dry Powder Inhaler (DPI) Drug Products", Chemistry, Manufacturing and Controls Documentation - Draft
- CDER (2001), "Sterility Requirements for Aqueous-Based Drug Products for Oral Inhalation", Small Entity Compliance
- CDER (2002), "Nasal Spray and Inhalation Solution, Suspension and Spray Drug Products", Chemistry, Manufacturing and Controls Documentation
- CDER (2003), "Integration of dose-counting mechanisms into MDI products", Clinical Medical
- CDER (2003), "Bioavailability and bioequivalence studies for nasal sprays for local action",
 Biopharmaceutics - Draft

Since December 2013, the FDA has issued a series of product specific guidance relating to various active pharmaceutical ingredients (APIs) including Fluticasone Propionate (FP), Salmeterol, Tiotropium, and Albuterol, amongst others, intended to help generic manufacturers navigate the Abbreviated New Drug Application (ANDA) process (see Special Applications, page 260).

Additionally, the FDA has been focusing on further strategies to support the development of generics, notably complex generics like OINDPs. The document "Alternative In Vitro Bioequivalence (BE) Pathways Which Can Reliably Ensure In Vivo Bioequivalence of Product Performance with a Generic." (Generic Drug User Fee Amendments (GDUFA)) states, "Additional research is ongoing to explore physicochemical API properties and device characteristics to demonstrate structural similarities (Q3) between test and reference Dry Powder Inhaler (DPI), Metered Dose Inhaler (MDI), and nasal products. A series of projects are exploring these Q3 characteristics, using Morphologically Directed Raman Spectroscopy (MDRS) in conjunction with in vitro dissolution, more realistic Aerodynamic Particle Size Distribution (APSD) measurement under realistic in vitro testing conditions, and particle surface

characterisation. The goal of this initiative is to provide greater understanding of the complex interactions between device, formulation, and patient factors, and eventually be able to predict the therapeutic behaviour based on these *in vitro* characteristics".

In April 2018, the FDA published a new Draft Guidance for Industry for comment (Revision 1) entitled "Metered Dose Inhaler (MDI) and Dry Powder Inhaler (DPI) Products - Quality Considerations".

This guidance which covers both quality and performance issues as well as CMC information is a revision of the previous 1998 Guidance "updated to reflect current standards and requirements to enhance understanding of appropriate development approaches for these products consistent with the quality by design (QbD) paradigm".



ICH Quality Guidelines				
Q1A - Q1F Stability	Q7 - Good Manufacturing Practice			
Q2 - Analytical Validation	Q8 - Pharmaceutical Development			
Q3A - Q4B Impurities	Q9 - Quality Risk Management			
Q4 - Q4B Pharmacopoeias	Q10 - Pharmaceutical Quality System			
Q5A - Q5E Quality of Biotechnological Products	Q11 - Development and Manufacture of Drug Substances			
Q6A - Q6B Specifications	Q12 - Lifecycle Management			
Q13 - Continuous Manufacturing of Drug Substances and Drug Products	Q14 - Analytical Procedure Development			

International Regulation and Harmonisation

The International Conference on Harmonisation of Technical Requirements for Registration of Pharmaceuticals for Human Use (ICH) is a unique organisation consisting of representatives from the EMA, MHLW and the FDA, and experts from the pharmaceutical industry in the associated regions, in a single forum.

The purpose of the ICH is to promote greater harmonisation in the way in which the individual regulatory bodies regulate new drugs such that the medicine reaches the patient economically and with the minimum delay, whilst maintaining the standards of safety, quality and efficacy necessary to safeguard public health. (Note: A similar organisation, the Global Harmonisation Task Force (GHTF) exists for medical devices).

Whilst not OINDP-specific, the ICH has concentrated on the preparation of four quality related guidelines:

- · ICH Q8(R2) Pharmaceutical Development
- · ICH Q9 Quality Risk Management
- · ICH Q10 Pharmaceutical Quality System
- ICH Q11 Development and Manufacture of Drug Substances

All of which have now been implemented by the regulatory authorities concerned (EMA, FDA and MHLW).

Collectively, these provide the guidelines for a new Pharmaceutical Quality System (PQS) described in ICH Q10. Based on International Standards Organisation (ISO) quality concepts, the new system includes Good Manufacturing Practice (GMP) regulations where applicable and complements ICH Q8 and ICH Q9.

One of the key features of the new PQS is the decision to extend the system to include all parts of the product lifecycle, namely:

- · Pharmaceutical Development
- Technology Transfer, e.g. from development to manufacturing
- Manufacturing
- · Product Discontinuation

This decision to extend the PQS to include Pharmaceutical Development through the concept of Quality by Design (QbD) is described in more detail in ICH Q8(R2) Part II Pharmaceutical Development - Annex.

The ICH Q8(R2) Annex describes the principles and gives examples of many of the essential concepts employed in QbD including Critical Quality Attributes (CQAs), Design Space and Control Strategy and its implementation through Process Analytical Technology (PAT) Tools.

ICH Q9 describes the principles of quality risk management and their application in a pharmaceutical environment.

ICH Q10 provides a model PQS covering the different stages of a product life cycle and thus a link between pharmaceutical development and manufacturing. As a guideline, ICH Q10 is not enforceable – however, it is likely that the regulators will consider it as standard best practice.

The practical implementation of the guidelines with respect to OINDPs is not easy because of (a) the complexities involved in manufacturing inhalation products, (b) the difficulties in applying real-time test methods to them, and (c) the lack of clear *in vitro – in vivo* correlations (IVIVCs) for most formulations. This continues to be an area of considerable discussion in pharmaceutical development, quality and regulatory circles.

ICH Q11 provides a Guideline to the "Development and Manufacture of Drug Substances" including the type and extent of information to be submitted in regulatory dossiers.

Mention should also be made of ICH Q12 which works with ICH Q8-Q11 guidelines to provide a framework to facilitate the management of the entire "Pharmaceutical Product Lifecycle".

Finally, two further topics were endorsed by the Assembly (ICH Q13 and ICH Q14) in June 2018.

ICH Q13, adopted in November 2021, outlines Current Good Manufacturing Practices (CGMP) specific to the Continuous Manufacturing (CM). The guideline will also provides guidance to industry and regulatory agencies regarding regulatory expectations on the development, implementation, and assessment of CM technologies used in the manufacture of drug substances and drug products.

The ICH Q14, adopted in May 2022, comes with a revision to the ICH Q2(R1) Guideline on Validation of Analytical Procedures, with a view to potentially combine both documents into one, for simplification and clarity.



Drug Safety, Quality and Efficacy - The Pharmacopoeias

The main role of the Pharmacopoeias is to define the standards with which medicines shall comply and the methods by which compliance will be adjudged.

As with the regulatory groups, the leading Pharmacopoeias tend to be those of the European Union, USA, China and Japan.

a) European Pharmacopoeia (Ph. Eur.)

In the Ph.Eur., the initial information relating to the control of OINDPs is contained in the monograph associated with the dosage form concerned, e.g. "Preparations for Inhalation (0671)" with cross references to appropriate methods of testing, e.g. "2.9.18. Preparations for Inhalation: Aerodynamic Assessment of Fine Particles".

The Ph.Eur. is also responsible for "Pharmeuropa", a bi-monthly publication available free online, which contains "Draft Monographs and General Texts for Comment" and "International Harmonisation". This publication is a good indicator of new and/or amended monographs, e.g. - "Calibration and Mensuration Issues for the Standard and Modified ACI" Vol.12.4, p.584-588 (2000) - "2.9.44 Preparations for Nebulisation: Characterisation" Vol. 18.2, p.280-283 (2006).

b) United States Pharmacopeia (USP)

Historically, the USP has adopted a similar approach to the Ph.Eur. but placed more emphasis on the Physical Tests and Determinations, e.g. "Aerosols, Nasal Sprays, Metered-Dose Inhalers and Dry Powder Inhalers <601>" than the type of dosage form, e.g. "Pharmaceutical Dosage Forms <1151>".

However, in USP 38, the Pharmacopeia introduced a series of new chapters, <1> through to <5>, which provide general information about the Critical Quality Attributes (CQAs) applicable to various dosage forms based on their route of administration.

These chapters detail the test procedures relevant to each dosage form, divided between those relating to product quality and those to product performance.

Product quality tests assess physical, chemical and microbial attributes. Product performance tests assess drug release from the dosage form concerned.

In the case of "Inhalation and Nasal Drug Products", the quality tests are described in Chapter <5> whereas the performance tests are described in Chapter <601>.

Both Ph.Eur. 2.9.44 and USP <1601> include chapters on tests designed to characterise nebulisers.

In addition, USP Chapter <1602> covers testing of the "Spacers and Valved Holding Chambers used with Inhalation Aerosols - Characterization Tests" and Chapter <1603> covers "Cascade Impactor Practices". In December 2023, Chapter <1604> "Data Interpretation of Aerodynamic Particle Size Distribution Measurements for Orally Inhaled Products" was made official, covering APSD data handling and analysis.

The USP also includes a series of product-specific monographs intended to provide clarification of the testing of certain generics by methods not previously specified in the general chapters.

Like Ph.Eur., USP produce a bi-monthly publication which contains discussion documents relating to new and/ or amended chapters and monographs. "Pharmacopeial Forum" features items relating to "In-Process Revision", "Harmonisation" and "Stimuli to the Revision Process".

c) Chinese Pharmacopoeia (ChP)

The ChP has four chapters contained within its Volume IV applicable to OINDPs, <0111>, <0112>, <0113> and <0951>, plus five drug specific monographs.

Chapter <0111> relates to general requirements applicable to MDIs, DPIs and nebulisers (incl. DDU) whilst <0951> describes those methods relating to APSD measurement for OINDPs.

d) Japanese Pharmacopoeia (JP)

The JP has two chapters related to OIPs, "Chapter <6.14> on Delivered Dose Uniformity" and "Chapter <6.15> on Particle Size Distribution". In addition to these, a General Chapter "G6.4 General Information" is available and applicable to OINDPs.

Device Safety, Quality and Efficacy -International Standards Organisation (ISO)

Most OINDPs are unique dosage forms in so far as that they comprise two components:

- (a) The drug formulation(s)
- (b) The medical device delivering that formulation to the patient

The responsibility of defining the standards relating to the medical device resides with the ISO.

The relevant standards are "ISO 20072 Aerosol drug delivery device design verification - Requirements and test methods" for inhalers and "ISO 27427 Anaesthetic and respiratory equipment - Nebulising systems and components" for nebulisers.



Expert Groups

In addition to the above, there are a number of industry and quasi-industry expert groups whose role it is to assist the regulatory bodies in establishing best practice in their thinking and guidance.



European Pharmaceutical Aerosol Group (EPAG)

A group of 17 member companies active in the OINDP market within Europe, formed to establish scientifically-based best practice, provide consensus comment to industry and government agencies on safety and quality issues, and recommend harmonised standards and methodology. Copley is an invited member of the cascade impactor sub-team.



International Pharmaceutical Consortium on Regulation and Science (IPAC-RS)

A group of 16 members and 10 associate members committed to advancing consensus-based, scientifically driven standards and regulations for OINDPs worldwide. Copley is an associate member.



Product Quality Research Institute (PQRI)

PQRI is a collaborative, research organisation involving the FDA's CDER, industry and academia.

It was formed to provide consensus advice on the scientific information to be submitted in a regulatory filing to CDER and has been involved in a number of OINDP-related products.

Organisational Chart: Guidelines and Regulations

	Metered-Dose Inhaler (MDI)*	Dry Powder Inhaler (DPI)	Soft Mist Inhaler	Nasal Products	Nebuliser
Regulatory					
	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products (2006)				
EMA Guidelines	Guideline on the Requirements for Clinical Documentation for Orally Inhaled Products (OIP) including the Requirements for Demonstration of Therapeutic Equivalence between Two Inhaled Products for use in the Treatment of Asthma and Chronic Obstructive Pulmonary Disease (COPD) in Adults and for use in the Treatment of Asthma in Children and Adolescents (2009)				
FDA Draft		e Inhaler (MDI) & Dry Powde cts (2018) - Quality Conside	, ,		
Guidance for Industry				Nasal Aerosols and Nasal Sprays for Local Action (2003)	
FDA Guidance for Industry				Nasal Spray, Inhalation Solution, Suspension & Spray Drug Products (2002)	
Drug Efficacy					
European Pharmacopoeia 2021 (10.5)		ns for Inhalations (Dosage F ssessment of Fine Particles	•	Nasal Preparations (Dosage Forms 0676)	Preparations for Nebulisation (Chapter 2.9.44)
US Pharmacopeia 2020 (USP 43)	Inhalation & Nasal Drug Products - General Information & Product Quality Tests <5> Aerosols, Nasal Sprays, Metered-Dose Inhalers and Dry Powder Inhalers <601> Uniformity of Dosage Units <905> Cascade Impactor Practices <1603> Data Interpretation of Aerodynamic Particle Size Distribution Measurements for Orally Inhaled Products <1604> Pharmaceutical Dosage Forms (Aerosols - Inhalations) <1151>				
Chinese	Spacers & VHCs <1602> Inhalation Pr	roducts - Metered-Dose, Dr	y Powder Inhalers and Nebu	ulisers - Delivered Dose Unifo	ormity <0111>
Pharmacopoeia 2020			c Particle Size Distribution (
Japanese Pharmacopoeia (JP17)	Delivered Dose Uniformity <6.14> Particle Size Distribution <6.15> General Information <6.4>				
Device Efficacy					
International Standards Organisation	Aerosol Drug Delivery Devices - Requirements and test methods (ISO 2007), 2013				Nebulizing Systems (ISO 27427: 2013)
Expert Groups					
European Pharmaceutical Aerosol Group (EPAG)	EPAG European based industry expert group involved in orally inhaled and nasal drug products				
International Pharmaceutical Consortium on Regulation & Science (IPAC-RS)	IPAC-RS US based industry expert group involved in orally inhaled and nasal drug products				
Product Quality Research Institute (PQRI)		A collaborative research or	PQRI ganisation involving FDA's C	DER, industry and academia	

Delivered Dose Uniformity (DDU)

One of the four **Critical Quality Attributes (CQAs)** that determine the safety, quality and efficacy of orally inhaled and nasal drug products (OINDPs) as discussed in the previous chapter, delivered dose is the total amount of drug emitted from the drug device that is available to the user, when the device is actuated correctly.


The delivered dose is measured by firing the drug device into a sampling apparatus containing a filter. The dose is captured, dissolved in solvent and an aliquot is then analysed, normally using high pressure liquid chromatography (HPLC).

Each OINDP dose typically contains a mixture of one or more active pharmaceutical ingredients (API) together with excipients designed to help with dose delivery to the patient. It is critical to assess that the API dosage delivered is consistent, or 'uniform' with each administration to ensure the correct drug amount is delivered to the patient each time.

The uniformity of the delivered dose, or DDU, of an OINDP must be ensured within and between devices. A number of tests have been defined by the various regulatory authorities, which are designed to demonstrate:



Inter-batch dose consistency



Intra-dose consistency for multi-dose inhalers throughout device life



The number of deliveries are greater than or equal to the label claim



In the case of dry powder inhalers (DPIs), different flow rates specific to the device resistance are considered

DDU Over the Entire Contents

Both the European Pharmacopoeia (Ph. Eur.) and United States Pharmacopoeia (USP) state that DDU tests should be carried out on all orally inhaled products (OIPs) and that in the case of multiple-dose devices* tests should be carried out throughout the life of the inhaler i.e. dose uniformity over the entire contents.

In the case of Ph.Eur., for example, this involves the collection of 10 doses throughout the life of each individual inhaler: three doses at the beginning, four in the middle and three at the end (see below).

* In the case of Ph. Eur., for DPIs this only applies to reservoir type devices.

Example: Ph. Eur. DDU Over the Entire Contents Requirements				
Inhaler Life	Beginning	Middle	End	
No. required doses	3 shots	4 shots	3 shots	
Dose no.	2, 3, 4	49, 50, 51, 52	98, 99, 100	
100 labelled doses		90 shots to waste		
Dose no.	2, 3, 4	99, 100, 101, 102	198, 199, 200	
200 labelled doses		190 shots to waste		

Similar testing requirements exist for other pharmacopoeias and regulatory guidance (see page 12). To obtain the required doses for analysis, the remaining contents of the inhaled device must be wasted (and done so appropriately, i.e. reproducibly and safely).

Collection Devices for DDU Testing

Depending on the type of inhaler device under test, different apparatus set-ups are required. The key collection devices are highlighted below. For further information about device-specific testing, please proceed to the relevant sections within this chapter.

For MDIs, DPIs, BAIs and SMIs


Dose Collection

Dose Uniformity Sampling Apparatus (DUSA)

Two types of DUSA are available for DDU testing - a DUSA for MDIs and a DUSA for DPIs; each are additionally suitable for testing other device types.

Typically, the device is connected to the DUSA via a mouthpiece or nosepiece adapter (see page 211). The drug-laden plume released upon actuation of the

device is drawn into the DUSA using a vacuum pump (see page 188) connected to the outlet via a suitable length of tubing.




DUSA for MDIs

Suitable for DDU testing of: MDIs BAIs SMIs

Also suitable for DDU testing of Nasal Sprays & Nasal Aerosols (see page 26)

The DUSA for MDIs consists of a sample collection tube, a filter to capture the delivered dose and a connector to connect the DUSA with the wider test set-up. It has been designed to enhance productivity and ensure ease-of-use.

Schematic of a DUSA for MDIs





Automating Dose Collection

Compatible with most MDIs, the Vertus® III automated shake and fire range offers extensive paramter control and monitoring for precise and reproducible testing.

Easing the burden associated with routine manual dose uniformity testing, the Vertus III range eliminates firing errors, controls air flow speed and automates leak testing.

The Vertus III+ model offers the additional benefit of automated shot weight collection.

For further information, please see page 284.

Dose Uniformity Sampling Apparatus (DUSA) for MDIs

Cat. No. Description

8201 Dosage Unit Sampling Apparatus for MDIs (Silicone Rubber Seals)

8201A Dosage Unit Sampling Apparatus for MDIs (LDPE Seals)

Accessories

8211 Stand for 10 MDI Collection Tubes

Note: Aluminium or 316 Stainless Steel DUSAs are available, if required

Spare Parts


Cat. No.	Description
8202	Set of 3 Silicone Rubber Seals for MDI
8202A	Set of 3 LDPE Seals for MDI
8203	Collection Tube for MDI
8204	Filter Support Cap for MDI
8205	Rinsing Cap (Silicone Rubber Seal) for MDI
8205A	Rinsing Cap (LDPE Seal) for MDI
8206	Flow Meter Cap (Silicone Rubber Seal) for MDI
8206A	Flow Meter Cap (LDPE Seal) for MDI
8207	Stainless Steel Filter Support Disc for MDI
8210	Pack of 500 Glass Fibre Filters for MDI

DUSA for DPIs

Suitable for DDU testing of: ODPIs


Also suitable for DDU testing of Nasal Powders (see page 26)

The DUSA for DPIs is a larger version of the DUSA for MDIs and is designed specifically to sample at flow rates up to 100 L/min. It is also used to characterise the flow resistance of DPIs. The pressure port (P1) in its wall is used to connect a critical flow controller to measure the pressure drop across the device.



During testing, air is drawn through the sampling apparatus to broadly simulate inhalation. A critical flow controller is required to control air flow supply to the inhaler and ensure critical (sonic) flow conditions during testing.

Basic system set-up for DDU testing of DPIs (according to Ph. Eur. and USP)



Dose Uniformity Sampling Apparatus (DUSA) for DPIs

Cat. No. D	escription
------------	------------

8601 Dosage Unit Sampling Apparatus for DPIs (Silicone Rubber Seals)8601A Dosage Unit Sampling Apparatus for DPIs (LDPE Seals)

Accessories

8604 Stand for 10 DPI Collection Tubes

Note: Aluminium or 316 Stainless Steel DUSAs are available, if required

Spare Parts

Cat. No.	Description		
8602	Set of 3 Silicone Rubber Seals for DPI		
8602A	Set of 3 LDPE Seals for DPI		
8603	Pack of 100 Glass Fibre Filters for DPI		
8606	Filter Support Cap for DPI		
8607	Rinsing Cap (Silicone Rubber Seal) for DPI		
8607A	Rinsing Cap (LDPE Seal) for DPI		
8608	Collection Tube with P1 Port for DPI		
8608A	Collection Tube without P1 Port for DPI		
8609	Flow Meter Cap (Silicone Rubber Seal) for DPI		
8609A	Flow Meter Cap (LDPE Seal) for DPI		
8610	Stainless Steel Filter Support Disc for DPI		

Waste Shot Collection

Firing inhaled drug product shots to waste requires an evacuation system, which captures the aerosol emitted from repeated actuations of the device. The system must be capable of trapping large quantities of the drug for safe disposal. We offer both manual and automated fire-to-waste systems.

Waste Shot Collector: WSC2



The Waste Shot Collector WSC2 is a compact vacuum filtration system suitable for use with a range of devices. It can be used in either standalone mode or integrated into the Inhaler Testing Workstation™ ITW (see page 204), via a switching valve, whereby the vacuum pump used for the DUSA powers both sampling and waste collection units.

The external dimensions of the inlet of the WSC2 are identical to those of the DUSA. This means that:

- the same mouthpiece adapter (and therefore inhaler) can be used with both pieces of equipment
- the two pieces of equipment are interchangeable within a test set-up so all shots are collected or discharged to waste under identical test conditions

Waste doses are captured in a disposable cartridge which collects and traps the contents in an integral HEPA filter, retaining 99.97% of particles over 0.3 microns in diameter.

The WSC2 is also suitable for nasal drug product waste dose collection. See page 28 for further information.

Waste Shot Collector WSC2

Description		
Waste Shot Collector WSC2 (including 1 Cartridge)		
Spare Filter Cartridge for Waste Shot Collector		
FRS Flow Meter Adapter		
DFM Flow Meter Adapter		
Waste Shot Tally Counter		

Automating Waste Dose Collection

Automating firing-to-waste is highly advantageous from the perspective of conserving analyst time, eliminating the risk of repetitive strain injury, and maximising the repeatability of test data; firing-to-waste under well-defined, closely controlled conditions eliminates a potential source of variability in testing.

Priming & Waste Module: Vertus® III/Vertus III+ For MDIs

The Priming & Waste Module integrates firing-to-waste into automated dose uniformity test collection methods, enabling compendial entire contents testing with minimal manual input. Vertus III and Vertus III+ can switch automatically between priming and test levels, firing-to-waste or to dose collection as required, without operator intervention, enabling highly efficient testing procedures, most notably to meet through-life test requirements for DDU. Additionally, the Priming & Waste Module can be used as a standalone interface for waste shot collection. For further information, see page 284.



DecaVertus® III For MDIs

DecaVertus III is a high-throughput automated shake and fire to waste system for reproducible and controlled waste shot collection. Accommodating up to 10 inhalers per run, DecaVertus III ensures firing-to-waste occurs under closely controlled conditions every time. Since DecaVertus III is fully compatible with Vertus III/Vertus III+, methods can be easily transferred between systems, with DecaVertus III often used to alleviate the burden of through-life testing.

For further information, see page 290.



British Pharmacopoeia (BP) Content Uniformity Apparatus for MDIs

In addition to the Ph.Eur. and USP specified DUSA, the BP has its own unique apparatus for determining the "Content of Active Ingredient delivered by actuation of the valve", likely retained for historical reasons. This comprises a stainless steel base plate having three legs and a central hole to accept the actuator stem in a small vessel (to which solvent is added) suitable for shaking.

BP Content Uniformity Apparatus for MDIs

Cat. No. Description

8212 BP Content Uniformity Apparatus for MDIs

For MDIs with Spacers/VHCs and for Nebulisers

Filter Holder

The Filter Holder is designed for DDU testing for both MDIs with spacers VHCs and for nebulisers.

The Filter Holder is designed for use together with a breathing simulator, which is used to apply the specific breathing profile required for representative device operating conditions (see page 156). A filter is contained within the holder, to capture the delivered dose. The device under test is interfaced with the filter holder using a suitable mouthpiece adapter. For assessing the effects of a facemask for each device type, see page 244.



Filter Holder for MDIs with Spacers/VHCs and for Nebulisers

Cat. No.	Description
9102	Filter Holder and Adapter for Breath Simulator BRS 100i
9102A	Filter Holder and Adapter for Breath Simulator BRS 200i/300i
9103	Pack of 100 Filters for Filter Holder
9104	Angle Adapter for Breath Simulator BRS 100i

For Nasal Sprays, Nasal Aerosols and Nasal Powders

We offer multiple in vitro fire-to-sample and fire-to-waste options for new and generic nasal drug product testing, plus options for automation to not only ease the burden associated with routine analysis but reduce variability and improve data integrity.

Dose Collection Dose Uniformity Sampling Apparatus (DUSA)

As recommended in Ph. Eur. 0676 and USP <601>, we offer the Inhaler Testing Workstation™ ITW with a specially designed DUSA holder that can be oriented in a vertical or near-vertical position for more representative manual nasal spray DDU testing. Holders are available for both the DUSA for MDIs (for Nasal Spray and Nasal Aerosol testing) and DUSA for DPIs (for Nasal Powder testing).

Product orientation for fire-to-waste can also match dose collection to help ensure data capture is consistent and truly reflects performance with our innovative vertical Waste Shot Collector WSC2 attachment (see page 28).





Automating Dose Collection For Nasal Sprays and Nasal Aerosols

Simplifying the measurement of delivered dose uniformity testing for nasal drug products in accordance with the European Pharmacopoeia (Ph. Eur.) Chapter 0676 and United States Pharmacopoeia (USP) Chapter <601>, the DUSA Interface Plate enables a simple, leak-free connection between a standard DUSA and the nasal drug product to ensure complete dose capture in a vertical orientation.

Designed for use with the Vertus® III/III+ automated shake and fire system, routine test set-ups for nasal drug products are now easily automated using our complete integrated solution. Vertus III+ offers the additional benefit of automated shot weight collection.

For further information see page 284.



Nasal Spray Dose Collector NSDC For Nasal Sprays

Loss of nasal spray sample due to dripping and leakage is common, due to the need to fire nasal sprays upwards to simulate actual product use. The NSDC apparatus represents a significant advancement in the

field of nasal spray testing. Designed with precision and functionality in mind, the NSDC (patent pending) offers several key features aimed at enhancing nasal sprays testing accuracy and reliability.



With an opening large enough to prevent splashback, but small enough to greatly reduce the risk of drips and leakage, the NSDC not only ensures accurate collection of the full dose, but also enables dose collection to occur in a vertical orientation to better replicate

in vivo usage. Ensuring accurate collection of the full dose in the vertical orientation, the NSDC can be utilised as a complementary or alternative solution to dose collection with a DUSA.

How does the NSDC work?


The unique design enables vertical actuation of the nasal spray, while minimising any loss of sample caused by dripping out of the collection device. The internal geometry of the NSDC has been engineered to gently direct the nasal spray plume into a well where the sample can be collected for subsequent analysis.




The 'shark fin' design deflects spray away from the centre point of the nozzle in an aerodynamic fashion to minimise the risk of any rebound



All points on the fin itself slope away from the centre point thereby encouraging any drips that form to run away from the centre





Automating Dose Collection

For Nasal Sprays

The NSDC can be used as a standalone device for manual nasal spray dose collection, but automation minimises variability while simultaneously improving productivity and reducing the health and safety risks associated with repeat manual actuation, such as the risk of repetitive strain injury.

The NSDC directly interfaces with Vertus® III/ Vertus III+ shake and fire systems to fully automate actuation and test flow control, meeting the need for highly consistent actuation, as required by USP <601>.

Additionally, the NSDC can be switched easily for the Nasal Spray Waste Collector NSWC (see page 29) for controlled and representative "fire-to-waste" for entire contents testing of multi-dose devices. Vertus III+ offers the additional benefit of automated shot weight collection.

For further information about automation with the Vertus III range, see page 284.



Nasal Spray Dose Collector

Cat. No. Description

9735 Nasal Spray Dose Collector (NSDC)9737 Nasal Spray Holder for use with NSDC



Waste Shot Collection

For nasal devices containing multiple doses, delivered dose testing may need to be conducted throughout the life of the device, i.e. dose uniformity over the entire contents. Regulatory guidance also recommends that waste collection occurs under similar conditions to dose collection. We offer a range of waste shot collection devices designed specifically for the waste shot collection of nasal drug products in accordance with the regulatory guidance.

Waste Shot Collector WSC2

The WSC2 (see page 24) can be mounted vertically at 90° for vertical (or near-vertical) nasal product waste shot collection as recommended in Ph. Eur. 0676 and USP <601>. Suitable for nasal sprays, nasal aerosols and nasal powders.

Nasal Spray Waste Collector NSWC For Nasal Sprays

The NSWC is designed to collect high volumes of waste doses with no splashback onto the nozzle, for safe and convenient disposal of the waste drug.



Automating Waste Dose Collection

For Nasal Sprays

The NSWC may be used for manual waste shot collection, but automation ensures that waste shot collection occurs under closely controlled and repeatable conditions. Interfacing directly with the Vertus® III/Vertus III+, the NSWC streamlines shot disposal in a time-efficient way, while reducing the health and safety risks associated with repeat manual actuation, such as the risk of repetitive strain injury.

Additionally, the NSWC can be switched easily for the Nasal Spray Dose Collector NSDC (see page 27) or DUSA Interface Plate for controlled and representative dose collection. Vertus III+ offers the additional benefit of automated shot weight collection.

For further information about automation with the Vertus III range, see page 284.



Nasal Spray Waste Collector

Cat. No. Description

9736 Nasal Spray Waste Collector (NSWC)


USP Monographs

The USP has product-specific monographs for a number of APIs including Albuterol (Salbutamol), and Fluticasone Propionate (FP)/Salmeterol combinations, which are used globally to treat asthma and COPD. Due to their widespread use and application, these active ingredients are routine targets for generic development.

These monographs cover both DDU testing and Aerodynamic Particle Size Distribution (APSD) measurement since these metrics are required for all OIPs due to their defining influence on the success and consistency of drug delivery.

We offer a range of test equipment that closely replicates the original apparatus used in the development of these reference labelled drugs (RLD), enabling bioequivalence testing in accordance with these monographs.

For more information about the various apparatus used, see page 270.



Sample Collection Apparatus for FP/Salmeterol Aerosols

Choose your Delivered Dose Collection Device

	DUSA for MDIs	DUSA for DPIs	BP Content Uniformity Apparatus for MDIs	Filter Holder	Nasal Spray Dose Collector (NSDC)	USP Monographs
MDI	Y	N	Y	N	N	Y
MDI with Spacer/VHC	N	N	N	Y	N	N
DPI	N	Y	N	N	N	Y
Nebuliser	N	N	N	Υ	N	N
SMI	Y	N	N	N	N	N
Nasal Spray	Y	N	N	N	Y	N
Nasal Aerosol	Y	N	N	N	N	N
Nasal Powder	N	Υ	N	N	N	N



Delivered Dose Uniformity

Metered Dose Inhalers (MDIs)

MDI aerosol characteristics are relatively insensitive to changes in air flow rate because the aerosolisation and dispersion mechanisms are dependent on the force generated by the propellant, rather than the patient's inspiratory effort. Therefore, for MDIs, the test flow rate is fixed at an arbitrary value of 28.3 L/min.

A vacuum pump is used to draw air through the assembled test set-up at this flow rate.

However, these test conditions are not applied for DDU testing when the MDI is intended for use with

an add-on device such as a spacer or valved holding chamber (VHC).

Further information about the DDU testing of MDIs with a spacer or VHC can be found on page 40.

Regulations & Guidelines

The sampling procedure and acceptance criteria for the DDU of MDIs varies according to the regulatory authority concerned.

Organisation	Chapter(s)/Guidance	Key DDU Tests	
ЕМА	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products 2006	Pharmaceutical Development: • DDU Through Container Life • DDU Over Patient Flow Rate Range Product Manufacture: • Mean Delivered Dose • Delivered Dose Uniformity • Content Uniformity / Uniformity of Dosage Units	
Ph. Eur.	Chapter 0671	Uniformity of Delivered Dose Number of Deliveries per Inhaler	
FDA	MDI & DPI Products - Quality Considerations Draft Guidance 2018	Delivered Dose Uniformity	
USP	Chapter <601>	Delivered Dose Uniformity of Product Dose Uniformity Over the Entire Unit Life	
Ch.P.	Chapter 0111	Delivered Dose Uniformity	
JP	Chapter 6.14	Delivered Dose Uniformity	

DDU Over the Entire Contents

BBO OVER the Entire contents				
Organisation	1st Test Tier No. of Inhalers	1st Test Tier Criteria	2nd Test Tier No. of Inhalers	2nd Test Tier Criteria
Ph.Eur	10 Inhalers/ 1 prime 3 beginning of life 4 middle of life 3 end of life	9/10 doses to be 75-125% of Mean All doses to be 65-135% of Mean Mean to be 85-115% of LC*	20 Inhalers/ 1 dose	27/30 doses to be 75-125% of Mean Value All doses to be 65-135% of Mean Value Mean Value to be 85-115% of LC*
USP	10 Inhalers/ 1 prime 1 beginning of life 1 end of life	N/A	N/A	N/A
EMA	As per Ph.Eur.	As per Ph.Eur.	As per Ph.Eur.	As per Ph.Eur.
FDA	10 Inhalers/ 1 beginning of life 1 end of life	18/20 doses to be 80-120% 20/20 to be 75-125% of TDD** Mean to be 85-115% of TDD**	20 Inhalers/ 1 beginning of life 1 end of life	54/60 doses to be 80-120% of TDD 60/60 to be 75-125% of TDD** Mean to be 85-115% of TDD**
Ch.P.	1 Inhaler/10 doses (MDIs) and Multidose DPIs) 10 inhalers/1 dose of each	9/10 doses to be 75-125% and all to be 65-135% of Average Delivered Dose Mean to be 80-120% of LC*	2 inhalers/20 doses 20 inhalers/1 dose of each	27/30 doses to be 75-125% and all to be 65-135% of Average Delivered Dose Mean to be 80-120% of LC*
JÞ	1 Inhaler/10 doses	9/10 to be 75-125% of Mean Value All to be 65-135% of Mean Value Mean to be 85-115% of LC*	2 inhalers/20 doses	27/30 doses to be 75-125% of Mean Value All to be 65-135% of Mean Value Mean to be 85-115% of LC*

DDU of MDIs Manual Test System Set-Up


The minimum set-up for DDU testing as specified by the Ph. Eur. comprises a sample collection tube, fitted at one end with a suitable mouthpiece adapter to accept the inhaler under test and connected at the other end to a vacuum pump capable of continuously drawing 28.3 L/min through the inhaler.

In addition to the specifications laid down by the Ph. Eur., the FDA recommends and the USP specifies that the volume of air to be sampled should not exceed 2

litres; this being the volume of air adjudged to be typical of the average patient.

This additional criterion can be met by positioning an electronically operated timer controlled two-way solenoid valve, such as that incorporated in the Breath Actuation Controller BAC 100i.

DDU for MDIs: Test Specifications		
Flow Rate (Q)	28.3 L/min	
Air Volume (Ph. Eur./EMA)	Not defined	
Air Volume (USP/FDA)	2 litres	



Related Accessories



DUSA Collection Tube Stand

Designed for the convenient transfer of multiple DUSA for MDIs around the laboratory. See page 22.

Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 179.

MDI Actuation Sensor/Footswitch

Learn more on page 196.

Suitable for most commercially available MDI canisters, the MDI Actuation Sensor connects directly to the Breath Actuation Controller BAC 100i to ensure precise synchronisation of MDI actuation. Alternatively, a Footswitch can be attached to trigger actuation. See page 179.

DDU of MDIs: Manual Test System Component Parts



Dose Uniformity Sampling Apparatus DUSA for MDIs

See page 20.

In addition to the DUSA for MDIs, the following is needed to complete a fully-operational test set-up for the delivered dose testing of MDIs:

Vacuum Pump

Designed for optimal operation at the low flow rates required for MDI testing, the Low Capacity Pump LCP6 represents the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.

See page 188 for further information about our Vacuum Pump range.





Breath Actuation Controller

Ensuring that the volume of air sampled does not exceed pharmacopoeial specifications, the Breath Actuation Controller BAC 100i contains an electronically operated, timer-controlled two-way solenoid valve and is positioned between the DUSA and vacuum pump.

See page 176 for further information about our Flow Controller range.



The BAC 100i can also be used for the testing of Breath-Actuated (or Breath-Operated) MDIs. In this case, the BAC 100i is used to initiate the flow, simultaneously triggering the breath-actuated inhaler.

Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.

See page 184 for further information about flow rate measurement.



DDU of MDIs: Manual Test System Component Parts



Inhaler Testing Workstation™

Designed to keep the sampling apparatus organised during testing and improve workflow efficiency, the Inhaler Testing Workstation ITW holds the DUSA collection tube, vacuum connector, flow meter and Waste Shot Collector WSC2.

See page 204 for further information.

Waste Shot Collector and Switching Valve

A compact vacuum filtration system, the Waste Shot Collector WSC2 captures aerosols emitted from repeated actuations of the inhaler, trapping large quantities of the drug for safe disposal. The Switching Valve is used to re-direct air flow between the collection device and WSC2 for quick and easy dose wasting.

See page 24 for further information about the WSC2. Alternatively, automate labour-intensive MDI waste shot collection with the Vertus® III/Vertus III+ and DecaVertus® III (see page 282).





Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler and the test apparatus. For a list of available Mouthpiece Adapters see page 211.

Custom Mouthpiece Adapters are available upon request.

Oualification

these requirements.

Good Manufacturing Practices (GMP) regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- · Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



Copley provides a range of qualification documentation, services and tools to meet

See page 310 for further information.

DDU of MDIs

Automated Test System Set-Up

The Vertus® III automated shake, fire and shot waste range is made up of integrated turn-key solutions for precise, controlled and reproducible MDI testing.

Compatible with most MDIs, the Vertus range offers analysts complete control over:

- The speed, angle and duration of shaking, ahead of actuation
- Firing force and the speed of application and release of that force
- The time delay between the end of shaking and device actuation



Improve inhaler testing accuracy and reproducibility



Increase productivity and reduce hassle



Replicate test methods across different sites with ease



Reduce handling errors and costly out-ofspecification results

Vertus III & Vertus III+

Offering high productivity, walkaway MDI testing, the Vertus III and Vertus III+ can collect doses at the start, middle and end of product life (including shots to waste as required).

The Vertus III+ also offers optional shot weight collection.





DecaVertus® III

Accepting up to 10 inhalers per run, DecaVertus III is a high-throughput shake and fire-to-waste system, ideal for alleviating the burden of tedious through-life testing.

Replaces the need for:

Vacuum Pump





Flow Controller

See page 282 for further information about the Vertus III and DecaVertus III range.

Inhaler Testing Workstation $^{\mathsf{TM}}$



Waste Shot Collector



Related Applications

We also offer a range of equipment for additional MDI testing application support:



For better *in vitro-in vivo* correlation (IVIVC) testing

See page 222



For cold Freon® effect testing

See page 255



For USP product-specific monograph testing

See page 270

Automation Tools



Improve efficiency



Reduce variability



Eliminate handling errors



Increase testing capacity



DUSA Shaker™ DTS 100i

The DTS 100i automates the internal rinsing of the DUSA collection tubes for full, fast and repeatable drug dissolution and drug recovery.

See page 294 for further information.

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320





MDIs with a Spacer/VHC


Add-on devices such as spacers, VHCs and reverse VHCs reduce or eliminate the need for coordination between actuation and inhalation and are widely used together with MDIs to overcome coordination issues.

When a patient uses an MDI without an add-on device, the drug particles contained within the delivered dose are inhaled almost instantaneously as the formulation is aerosolised. In contrast, when an add-on device such as a spacer or VHC is used, the patient inhales the drug from a reservoir of aerosolised particles.

The additional dead volume provided by this reservoir allows aerosol expansion, but also an opportunity for particle impaction, settling and/or electrostatic deposition within the chamber itself, all of which can change the delivered dose.

As the use of add-on devices has grown, the regulatory authorities have become increasingly aware of the need to test with add-on devices as distinct from MDIs alone.

The amount of drug received by the patient using an add-on device with an MDI will be directly influenced by the inhalation profile of the user concerned. For that reason, tests call for the application of specific breathing profiles to reflect the physiology of the intended user, see Table 1.



Regulation & Guidelines

The sampling procedure for the DDU testing of MDIs with a spacer/VHC varies according to the regulatory authority concerned.

Organisation	Chapter(s)/Guidance	Key DDU Tests
FDA	MDI & DPI Products - Quality Considerations Draft Guidance 2018	Effect of Flow Rate and Inhalation Delay on MDIs with Spacers
USP	Chapter <1602>	Mass of drug delivered - fully coordinated and fully uncoordinated


Table 1: Representative Tidal Breathing Patterns					
	Paediatric Ad			ult	
Parameter	Neonate	Infant	Child	Normal 1	Normal 2
Tidal Volume (mL)	25	50	155	770	500
Frequency (cycles/min)	40	30	25	12	13
I/E Ratio	1:3	1:3	1:2	1:2	1:2
Minute Volume (mL)	1000	1500	3875	9240	6500

For DDU over the entire contents testing of MDIs with a spacer/VHC and a facemask, see page 246.

DDU of MDIs with a Spacer/VHC Test System Set-Up

The standard sampling apparatus for MDIs with an add-on device consists of a breathing simulator to generate the specified breath profile, a filter holder containing the filter to capture the delivered dose and a suitable mouthpiece adapter to connect the filter holder to the mouthpiece of the spacer/VHC concerned.

In the case of VHCs, tests are also carried out to compare the dose received when use is coordinated or uncoordinated with device actuation, to assess the impact of valve operation.



Related Accessories



MDI Actuation Sensor/Footswitch

Suitable for most commercially available MDI canisters, the MDI
Actuation Sensor connects directly to the Breath Actuation Controller
BAC 100i to ensure precise synchronisation of MDI actuation.
Alternatively, a Footswitch can be attached to trigger actuation.
See page 179.

Considering the effects of environmental variability?

Our environmental control solutions are designed to help improve the accuracy, sensitivity and reproducibility of test data.

Learn more on page 196.



TOP TIP

The constant 28.3 L/min air flow rate applied during the testing of MDIs is replaced by a specific patient relevant tidal breath profile more representative of the conditions applied by the patient when using an add-on device.

DDU of MDIs with a Spacer/VHC: Test System Component Parts



Filter Holder (with Adapter for Breath Simulator Model BRS 100i)

See page 25.

In addition to the Filter Holder, the following is needed to complete a fully-operational test set-up for the delivered dose testing of MDIs with a spacer/VHC.

Breathing Simulator

Providing breathing profiles more representative of *in vivo* behaviour than conventional systems offering a constant flow rate, the Breathing Simulator BRS 200i is ideal for assessing the effects of a spacer or VHC on the DDU of MDIs.

Alternatively, a basic entry-level model, the Breathing Simulator BRS 100i, is also available.

Find out more about our range of Breathing Simulators on page 156.





Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the spacer/VHC and the test apparatus. For a list of available Mouthpiece Adapters see page 211.

Custom Mouthpiece Adapters are available upon request.

Qualification

GMP regulations require that





 Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.

Related Applications

We also offer a range of equipment for additional MDIs with a spacer/VHC testing application support:



For facemask testing See page 244

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321




Servicing See page 310



Support See page 320



Design See page 320






Dry Powder Inhalers (DPIs)

For DPIs, the test regime is more complex than for MDIs, since aerosolisation depends on the strength and duration of a single inhalation by the user.

During a single, deep inhalation, a typical adult produces a pressure drop over the device of approximately 4 kPa. Depending on the device flow resistance this will yield a flow rate, typical of the mean patient inhalation flow rate, that is then used for all the required testing of that device.

DDU for DPIs: Test Specifications		
Flow Rate (Q) Device dependent (4 kPa)		
Air Volume (Ph. Eur./EMA)	4 litres	
Air Volume (USP/FDA)	2 litres	

Pressure difference between lungs and atmosphere when inhaling through a DPI



Regulations & Guidelines

The sampling procedure and acceptance criteria for the DDU testing of DPIs varies according to the regulatory authority concerned.

Organisation	Chapter(s)/Guidance	Key DDU Tests
ЕМА	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products 2006	Pharmaceutical Development: • DDU Through Container Life • DDU Over Patient Flow Rate Range Product Manufacture: • Mean Delivered Dose • Delivered Dose Uniformity • Content Uniformity / Uniformity of Dosage Units
Ph. Eur.	Chapter 0671	Uniformity of Delivered Dose Number of Deliveries per Inhaler
FDA	MDI & DPI Products - Quality Considerations Draft Guidance 2018	Delivered Dose Uniformity
USP	Chapter <601>	Delivered Dose Uniformity of Product Dose Uniformity Over the Entire Unit Life
Ch.P.	Chapter 0111	Delivered Dose Uniformity
JÞ	Chapter 6.14	Delivered Dose Uniformity


DDU Over the Entire Contents

Organisation	1st Test Tier No. of Inhalers	1st Test Tier Criteria	2nd Test Tier No. of Inhalers	2nd Test Tier Criteria
Ph.Eur	10 Inhalers/ 1 prime 3 beginning of life 4 middle of life 3 end of life	9/10 doses to be 75-125% of Mean All doses to be 65-135% of Mean Mean to be 85-115% of LC*	20 Inhalers/1 dose	27/30 doses to be 75-125% of Mean Value All doses to be 65-135% of Mean Value Mean Value to be 85-115% of LC*
USP	10 Inhalers/ 1 prime 1 beginning of life 1 end of life	N/A	N/A	N/A
EMA	As per Ph.Eur.	As per Ph.Eur.	As per Ph.Eur.	As per Ph.Eur.
FDA	10 Inhalers/ 1 beginning of life 1 end of life	18/20 doses to be 80-120% 20/20 to be 75-125% of TDD** Mean to be 85-115% of TDD**	20 Inhalers/ 1 beginning of life 1 end of life	54/60 doses to be 80-120% of TDD 60/60 to be 75-125% of TDD** Mean to be 85-115% of TDD**
Ch.P.	1 Inhaler/10 doses (MDIs) and Multidose DPIs) 10 inhalers/1 dose of each	9/10 doses to be 75-125% and all to be 65-135% of Average Delivered Dose Mean to be 80-120% of LC*	2 inhalers/20 doses 20 inhalers/1 dose of each	27/30 doses to be 75-125% and all to be 65-135% of Average Delivered Dose Mean to be 80-120% of LC*
JP	1 Inhaler/10 doses	9/10 to be 75-125% of Mean Value All to be 65-135% of Mean Value Mean to be 85-115% of LC*	2 inhalers/20 doses	27/30 doses to be 75-125% of Mean Value All to be 65-135% of Mean Value Mean to be 85-115% of LC*

DDU of DPIs Test System Set-Up

The basic requirements for DPI DDU testing are the same as for MDI testing, namely DUSA, mouthpiece adapter, vacuum pump and flow meter. However, a critical flow controller (e.g. Critical Flow Controller TPK 100i) to measure the pressure drop across the device and control the flow conditions during testing is also required.

This is mandatory because most DPIs are passive breath-actuated devices which rely on the patient's inspiration rather than a propellant for dose aerosolisation and delivery. The testing of DPIs is further complicated by the fact that different inhalers provide varying degrees of flow resistance, i.e. some require more effort to inhale through than others. Find out more about critical flow control on page 172.



Related Accessories



DUSA Collection Tube Stand

Designed for the convenient transfer of multiple DUSA for DPIs around the laboratory. See page 23.

Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 183.

Footswitch

Connecting directly to the Critical Flow Controller TPK 100i, the Footswitch enables precise synchronisation of DPI device actuation with the onset of flow. See page 183.

DDU of DPIs: Test System Component Parts



Dose Uniformity Sampling Apparatus DUSA for DPIs

See page 22.

In addition to the DUSA for DPIs, the following is needed to complete a fully-operational test set-up for the delivered dose testing of DPIs:

Vacuum Pump

Ideal for the higher, sonic flow rate testing requirements of DPIs, the High Capacity Pump HCP6 and Super Capacity Pump SCP6 represent the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.

See page 188 for further information about our Vacuum Pump range.





Critical Flow Controller

Simplify DPI test system set-up in accordance with pharmacopoeial recommendations with the Critical Flow Controller series. Positioned between the DUSA and vacuum pump, the Critical Flow Controller TPK 100i ensures critical (sonic) flow conditions during testing. It measures and records all parameters required for testing and controlling flow conditions.

See page 172 for further information about our Flow Controller Range.

Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.

See page 184 for further information about flow rate measurement.



DDU of DPIs: Test System Component Parts



Inhaler Testing Workstation™

Designed to keep the sampling apparatus organised during testing and improve workflow efficiency, the Inhaler Testing Workstation ITW holds the DUSA collection tube, vacuum connector, flow meter and Waste Shot Collector WSC2

COPLEY

See page 204 for further information.

Waste Shot Collector and Switching Valve

A compact vacuum filtration system, the Waste Shot Collector WSC2 captures aerosols emitted from repeated actuations of the inhaler, trapping large quantities of the drug for safe disposal. The Switching Valve is used to re-direct air flow between the collection device and WSC2 for quick and easy dose wasting. Please note: only required for multi-dose devices.

See page 24 for further information about the WSC2.





Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus. For a list of available Mouthpiece Adapters see page 211.

Custom Mouthpiece Adapters are available upon request.

DDU Over the Entire Contents

In the case of DPI reservoir type devices, tests should be carried out throughout the life of the inhaler, i.e. dose uniformity over the entire contents. For further information, see page 19.

Oualification

GMP regulations require that

- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.



Related Applications

We also offer a range of equipment for additional DPI testing application support:



For better *in vitro-in vivo* correlation (IVIVC) testing

See page 222



For USP product-specific monograph testing

See page 270

Automation Tools



Improve efficiency



Reduce variability



Eliminate handling errors



Increase testing capacity



DUSA Shaker™ DTS 100i

The DTS 100i automates the internal rinsing of the DUSA collection tubes for full, fast and repeatable drug dissolution and drug recovery.

See page 294 for further information.

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320



Nebulisers


The delivered dose testing of nebulisers is carried out to determine the total amount of drug a patient might be expected to receive during a treatment period, rather than through one inhalation.

Given the mode of operation of nebulisers, well-defined tidal breathing profiles for specific patient types are specified for testing (see Table 2). These profiles can be reliably achieved using breathing simulators (see page 156).

Delivered Dose Testing Requirements for Nebulisers

The delivered dose of a nebuliser is quantified via two discrete metrics: the active substance delivery rate and the total active substance delivered.

To measure active substance delivery rate the output from the nebuliser is captured on a filter, under appropriate test conditions, over a specified time (typically 60 seconds). Replacing the filter and continuing the test until nebulisation stops, because the reservoir is empty, enables calculation of the second metric – total active substance delivered. This is the total mass collected during steps 1 and 2 of the test.



Regulations and Guidelines

The Filter Holder apparatus is used to perform those tests specified in the Pharmacopoeias relating to:

- Preparations for Nebulisation: Characterisation (Ph. Eur. 2.9.44)
- General Information: Products for Nebulization Characterization Tests (USP <1601>)

Organisation	Chapter(s)/Guidance	Key DDU Tests listed
ЕМА	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products	Drug Delivery Rate Total Drug Delivered
Ph. Eur.	Chapter 2.9.44. Preparations for Nebulisation: Characterisation	Ph. Eur. : Active Substance Delivery Rate Total Active Substance Delivered
FDA	Guidance for Industry: Nasal Spray and Inhalation Solution, Suspension and Spray Drug Products - Chemistry, Manufacturing and Controls Documentation	Content Uniformity
USP	Chapter <1601> Products for Nebulization - Characterization Tests	Drug Substance Delivery Rate Total Drug Substance Delivered
Ch.P.	Chapter 0111	Delivered Dose Uniformity
JP	-	-

Table 2 : Breathing Simulator Specifications for Nebuliser Characterisation Tests				
	Adult	Neonatal	Infant	Child
Total Volume	500 ml	25 ml	50 ml	155 ml
Frequency	15 cycles/min	40cycles/min	30 cycles/min	25 cycles/min
Waveform	Sinusoidal	Sinusoidal	Sinusoidal	Sinusoidal
I/E Ratio	1:1	1:3	1:3	1:2

DDU of Nebulisers Test System Set-Up

The sampling apparatus for nebulisers (mouthpiece-based products) consists of a breathing simulator to generate the specified breathing profile, a filter holder containing the filter to capture the delivered dose and a suitable mouthpiece adapter to connect the filter holder to the nebuliser under test.



DDU of Nebulisers: Test System Component Parts



Filter Holder (with Angle Adapter and Adapter for Breathing Simulator Model BRS 100i)

See page 25.

In addition to the Filter Holder, the following is needed to complete a fully-operational test set-up for the delivered dose testing of nebulisers:

Breathing Simulator

Providing breathing profiles more representative of *in vivo* behaviour than conventional systems offering a constant flow rate, the Breathing Simulator BRS 200i is ideal for assessing the DDU of nebulisers.

A basic entry-level model, the Breathing Simulator BRS 100i, is also available.

Find out more about our range of Breathing Simulators on page 156.





Mouthpiece Adapter


Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the nebuliser and the test apparatus. For a list of available Mouthpiece Adapters See page 211.

Custom Mouthpiece Adapters are available upon request.

Qualification

GMP regulations require that

- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.



Related Applications

We also offer a range of equipment for additional nebuliser testing application support:



For better *in vitro-in vivo* correlation (IVIVC) testing

See page 222



For facemask testing

See page 244

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321




Servicing See page 310



Support See page 320



Design See page 320





Soft Mist Inhalers (SMIs)

Since they are active, aqueous-based devices, the DDU testing of SMIs is similar to that of MDIs, with testing carried out at a constant flow rate of 28.3 L/min.


Regulations and Guidelines

The sampling procedure and acceptance criteria for the DDU testing of SMIs varies according to the regulatory authority concerned.

Organisation	Chapter(s)/Guidance	Key DDU Tests
ЕМА	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products 2006	Delivered Dose Uniformaty
Ph. Eur.	-	-
FDA	Guidance for Industry: Nasal Spray and Inhalation Solution, Suspension and Spray Drug Products - Chemistry, Manufacturing and Controls Documentation	Content Uniformity
USP	-	-
Ch.P.	Chapter 0111	Delivered Dose Uniformity
JP	-	-

DDU of SMIs

Test System Set-Up



- A Vacuum Pump
- B Breath Actuation Controller
- C Inhaler Testing
 Workstation™
- D Flow Rate Sensor

- Mouthpiece Adapter
- F DUSA for
- G Waste Shot Collector
- H Switching Valve

Considering the effects of environmental variability?

Our environmental control solutions are designed to help improve the accuracy, sensitivity and reproducibility of test data.

Learn more on page 196.

COPLEY

Related Accessories



DUSA Collection Tube Stand

Designed for the convenient transfer of multiple DUSA for MDIs around the laboratory. See page 22.

Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 179.

Footswitch

Connecting directly to the Breath Actuation Controller BAC 100i, the Footswitch enables precise synchronisation of SMI device actuation with the onset of flow. See page 179.

DDU of SMIs: Test System Component Parts



Dose Uniformity Sampling Apparatus DUSA for MDIs

See page 20.

In addition to the DUSA for MDIs, the following is needed to complete a fully-operational test set-up for the delivered dose testing of SMIs:

Vacuum Pump

Designed for optimal operation at the low flow rates required for SMI testing, the Low Capacity Pump LCP6 represents the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.







Breath Actuation Controller

Ensuring that the volume of air sampled does not exceed pharmacopoeial specifications, the Breath Actuation Controller BAC 100i contains an electronically operated, timer-controlled two-way solenoid valve and is positioned between the DUSA and vacuum pump.

See page 172 for further information about our Flow Controller range.

Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.

See page 184 for further information about flow rate measurement.





Inhaler Testing Workstation™

Designed to keep the sampling apparatus organised during testing and improve workflow efficiency, the ITW holds the DUSA collection tube, vacuum connector, flow meter and Waste Shot Collector WSC2.

See page 204 for further information.

Waste Shot Collector and Switching Valve

A compact vacuum filtration system, the Waste Shot Collector WSC2 captures aerosols emitted from repeated actuations of the inhaler, trapping large quantities of the drug for safe disposal. The Switching Valve is used to re-direct air flow between the collection device and WSC2 for quick and easy dose wasting. Please note: only required for multi-dose devices.

See page 24 for further information about the WSC2.





Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus. For a list of available Mouthpiece Adapters see page 211.

Custom Mouthpiece Adapters are available upon request.

DDU Over the Entire Contents

In the case of multiple dose devices, tests might need to be carried out throughout the life of the inhaler, i.e. dose uniformity over the entire contents. For further information, see page 19.

Qualification

GMP regulations require that

- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.



Related Applications

We also offer a range of equipment for additional SMI testing application support:



For better in vitro-in vivo correlation (IVIVC) testing See page 222



For USP product-specific monograph testing See page 270

Automation Tools



Improve efficiency



Reduce variability



Eliminate handling errors



testing capacity



DUSA Shaker™ DTS 100i

The DTS 100i automates the internal rinsing of the DUSA collection tubes for full, fast and repeatable drug dissolution and drug recovery.

See page 294 for further information.

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



Training See page 321



Servicing See page 310



Support See page 320



Design See page 320





Nasal Sprays

According to regulatory guidance, for the DDU testing of nasal sprays, the test unit should be actuated in a vertical or near-vertical, valve-up position with adequate controls over the critical mechanical actuation parameters, such as actuation force, speed and rest periods. USP <601> also requires the need for highly consistent actuation.

Regulations and Guidelines

The sampling procedure and acceptance criteria for the DDU testing of nasal sprays varies according to the regulatory authority concerned.

Organisation	Chapter(s)/Guidance	Key DDU Tests
ЕМА	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products	Delivered Dose Uniformity Through Container Life
Ph. Eur.	Chapter 0676	Uniformity of Delivered Dose
FDA	Guidance for Industry: Nasal Spray and Inhalation Solution, Suspension and Spray Drug Products - Chemistry, Manufacturing and Controls Documentation	Content Uniformity
USP	Chapter <601> Inhalation and Nasal Drug Products: Aerosols, Sprays, and Powders - Performance Quality Tests	Delivered Dose Uniformity of Product
Ch.P.	Chapter 0111	Delivered Dose Uniformity
JP	-	-

DDU of Nasal Sprays Automated Test System Set-Up

Routine DDU test set-ups for nasal sprays are easily automated with Vertus® III. Compatible with most nasal sprays, the Vertus III range simplifies the delivered dose uniformity testing of nasal sprays in accordance with Ph. Eur. Chapter 0676 and USP Chapter <601>.

The Vertus III range offers analysts complete control over:

- The speed, angle and duration of shaking, ahead of actuation
- · Firing force and the speed of application and release of that force
- The time delay between the end of shaking and device actuation



Improve nasal spray testing accuracy and reproducibility



Replicate test methods across different sites with ease



Reduce handling errors and costly out-ofspecification results



Increase productivity and reduce hassle

Vertus III+ with DUSA Interface Plate





Automated shot weight measurement via an integrated balance

- A Vertus® III+ with Balance
- - DUSA Interface Plate



To find out more about our range of Automated Shake & Fire systems, see page 282.

Manual Test System Set-Up

As recommended in Ph. Eur. 0676 and USP <601>, we offer the Inhaler Testing Workstation™ ITW with a DUSA holder that can be oriented in a vertical or near-vertical position for more representative manual nasal spray DDU sampling. Product orientation for fire-to-waste should also match dose collection to help ensure data capture is more consistent and truly reflects performance with our innovative vertical Waste Shot Collector WSC2 attachment.



NSDC: Manual Test System Set-Up

Used together with its manual holder, the NSDC is a compact dose collection system designed for manual DDU sampling of nasal sprays. This convenient system is ideal for quick, hassle-free DDU testing. See page 27 for further information.



Considering the effects of environmental variability?

Our environmental control solutions are designed to help improve the accuracy, sensitivity and reproducibility of test data.

Learn more on page 196.



DDU Over the Entire Contents

For more information about testing throughout the life of the nasal spray, in the case of multiple dose devices, i.e. dose uniformity over the entire contents, see page 19.

Qualification

GMP regulations require that





 Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.

Automation Tools



Improve efficienc



Reduce variability



Eliminate handling errors



Increase testing capacity



DUSA Shaker™ DTS 100i

The DTS 100i automates the internal rinsing of the DUSA collection tubes for full, fast and repeatable drug dissolution and drug recovery.

See page 294 for further information.

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



ServicingSee page 310



Support See page 320



Design See page 320



Nasal Aerosols

DDU testing of nasal aerosols follows a similar process to that of MDIs (page 32), since both use a propellant to deliver a specified volume of active ingredient(s) upon actuation of a metered valve system. Testing is typically conducted at a fixed flow rate of 28.3 L/min using a DUSA for MDIs for sample collection.

Regulations and Guidelines

The sampling procedure and acceptance criteria for the DDU of nasal aerosols varies according to the regulatory authority concerned.

Organisation	Chapter(s)/Guidance	Key DDU Tests
EMA	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products	Delivered Dose Uniformity Through Container Life
Ph. Eur.	Chapter 0676	Uniformity of Delivered Dose
FDA	Guidance for Industry: Metered Dose Inhaler (MDI) and Dry Powder Inhaler (DPI) Drug Products	Content Uniformity
USP	Chapter <601> Inhalation and Nasal Drug Products: Aerosols, Sprays, and Powders - Performance Quality Tests	Delivered Dose Uniformity of Product
Ch.P.	Chapter 0111	Delivered Dose Uniformity
JP	-	-

DDU of Nasal Aerosols Test System Set-Up

Routine DDU test set-ups for nasal aerosols are easily automated with Vertus® III. Compatible with most nasal aerosols, the Vertus III range simplifies the delivered dose uniformity testing of nasal aerosols in accordance with Ph. Eur. Chapter 0676 and USP Chapter <601>.

The Vertus III range offers analysts complete control over:

- The speed, angle and duration of shaking ahead of actuation
- · Firing force and the speed of application and release of that force
- · The time delay between the end of shaking and device actuation



Improve nasal aerosol testing accuracy and reproducibility



Replicate test methods across different sites with ease



Reduce handling errors and costly out-ofspecification results



Increase productivity and reduce hassle

Vertus III+ with DUSA Interface Plate



For more information about automation with the Vertus® range, see page 284.




Automated shot weight measurement via an integrated balance





Manual Test System Set-Up

As recommended in Ph. Eur. 0676 and USP <601>, we also offer the Inhaler Testing Workstation™ ITW with a DUSA holder that can be oriented in a vertical or near-vertical position for more representative manual nasal aerosol DDU testing. With our innovative vertical Waste Shot Collector WSC2 attachment, product orientation for fire-to-waste can also match dose collection to help ensure dose capture is more consistent and truly reflects performance.



DDU Over the Entire Contents


For more information about testing throughout the life of nasal sprays, in the case of multiple dose devices, i.e. dose uniformity over the entire contents, see page 28.

Automation Tools











DUSA Shaker™ DTS 100i

The DTS 100i automates the internal rinsing of the DUSA collection tubes for full, fast and repeatable drug dissolution and drug recovery.

See page 294 for further information.

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320



Nasal Powders


The minimum requirements for nasal powder delivered dose testing are the same as for DPI testing (see page 46), namely DUSA, nosepiece adapter, vacuum pump and flow meter, plus a critical flow controller to measure the pressure drop across the device and control flow conditions during testing.

Regulations and Guidelines

The sampling procedure and acceptance criteria for the DDU testing of nasal powders varies according to the regulatory authority concerned.

Organisation	Chapter(s)/Guidance	Key DDU Tests
ЕМА	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products	Delivered Dose Uniformity Through Container Life
Ph. Eur.	Chapter 0676	Uniformity of Delivered Dose
FDA	Guidance for Industry: Metered Dose Inhaler (MDI) and Dry Powder Inhaler (DPI) Drug Products	Content Uniformity
USP	Chapter <601> Inhalation and Nasal Drug Products: Aerosols, Sprays, and Powders - Performance Quality Tests	Delivered Dose Uniformity of Product
Ch.P.	Chapter 0111	Delivered Dose Uniformity
JP	-	-

DDU of Nasal Powders Test System Set-Up



- Vacuum Pump
- Controller
- Sensor FRS
- Collector

- DUSA for DPIs
- Nosepiece . Adapter
- Inhaler Testing Workstation™
- Switching Valve

Considering the effects of environmental variability?

Our environmental control solutions are designed to help improve the accuracy, sensitivity and reproducibility of test data.

Learn more on page 196.



Related Accessories



DUSA Collection Tube Stand

Designed for the convenient transfer of multiple DUSA for DPIs around the laboratory. See page 23.

Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 183.

Footswitch

Connecting directly to the Critical Flow Controller TPK 100i, the Footswitch enables precise synchronisation of nasal powder device actuation with the onset of flow. See page 183.

DDU of Nasal Powders: Test System Component Parts



Dose Uniformity Sampling Apparatus (DUSA) for DPIs

See page 22.

In addition to the DUSA for DPIs, the following is needed to complete a fully-operational test set-up for the delivered dose testing nasal powders:

Vacuum Pump

Ideal for the higher, sonic flow rate testing requirements of nasal powders, the High Capacity Pump HCP6 represents the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.







Critical Flow Controller

Simplify nasal powder test system set-up in accordance with pharmacopoeial recommendations with the Critical Flow Controller series. Positioned between the DUSA and vacuum pump, the Critical Flow Controller TPK 100i ensures critical (sonic) flow conditions during testing. It measures and records all required parameters required for testing and for controlling flow conditions.

See page 172 for further information about our Flow Controller range.

Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.

See page 184 for further information about flow rate measurement.





Inhaler Testing Workstation™

Designed to keep the sampling apparatus organised during testing and improve workflow efficiency, the Inhaler Testing Workstation ITW holds the DUSA collection tube, vacuum connector, flow meter Waste Shot Collector WSC2.

See page 204 for further information.

Waste Shot Collector and Switching Valve

A compact vacuum filtration system, the Waste Shot Collector WSC2 captures aerosols emitted from repeated actuations of the inhaler, trapping large quantities of the drug for safe disposal. The Switching Valve is used to re-direct air flow between the collection device and WSC2 for quick and easy dose wasting. Please note: only required for multi-dose devices.

See page 28 for further information about the WSC2.





Nosepiece Adapter

Special nosepiece adapters are available to accommodate the nasal powder device and interface it with the test set-up.

See page 211 for further information.

DDU Over the Entire Contents

In the case of multiple dose devices, tests might need to be carried out throughout the life of the nasal powder, i.e. dose uniformity over the entire contents. For further information, see page 28.

Qualification

GMP regulations require that


- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.




Automation Tools











DUSA Shaker™ DTS 100i

The DTS 100i automates the internal rinsing of the DUSA collection tubes for full, fast and repeatable drug dissolution and drug recovery.

See page 294 for further information.

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320




Aerodynamic Particle Size Distribution

Together with delivered dose, aerodynamic particle size distribution (APSD) is typically identified as a Critical Quality Attribute (CQA) for orally inhaled and nasal drug products (OINDPs) making it a primary focus for in vitro characterisation. The APSD of an OINDP defines how particles behave in a moving air stream. It is intuitively relevant to the understanding of likely lung deposition and hence potential drug efficacy.

To be therapeutically effective, inhaled drug particles should ideally be in the range of 1 to 5 microns to deposit in the lungs. Particles more than 5 microns will generally impact in the oropharynx and be swallowed, whereas below 1 micron particles will likely remain

entrained in the air stream and be exhaled. The mass of dose delivered at a particle size below 5 microns is normally described as the fine particle mass (FPM) or dose (FPD) and is an important metric for OIPs.



Particle Dae > 10 microns (Mouth/Throat)

Particle $D_{ae} = 1 - 5$ microns (Deep Lungs)

Particle D_{ae} = 5 - 10 microns (Upper Respiratory Tract)

Particles D_{ae} = 1 micron (Exhaled)

ГОР

 $D_{ae} = D_p \frac{1}{2} f(S)$

D = Geometric diameter

p = Particle density

S = Shape factor




 $D = 3 \mu m$ $D_{ae} = 3 \mu m$



 $D_{ae} = 3 \mu m$

An Introduction to Cascade Impaction

The cascade impactor is the instrument of choice for both regulators and pharmacopoeias when measuring the APSD of inhaled drug products due to some unique features. Cascade impactors separate a sample on the basis of particle inertia (which is a function of velocity and aerodynamic particle size) without the need to know either particle density or shape.





The term "impactor" is generally used for an instrument where the particles "impact' on a dry impaction plate or cup. The term "impinger" is used to describe instruments where the particles impinge into a liquid or onto a moist collection surface. Cascade impactors have three unique features which make them the ideal tool for particle size assessment of inhaled products.

1. Cascade impactors measure aerodynamic particle size data

Cascade impactors measure aerodynamic particle size which is a function of particle density, as well as the physical dimensions and shape of the particles concerned. This is a more relevant parameter when studying how particles behave in a moving air stream (as exemplified by the respiratory tract) rather than simple "geometric" size.


2. Cascade impactors deliver active pharmaceutical ingredient (API) specific measurements


Cascade impactors provide a direct means of recovering and quantifying API contained in the aerosol cloud. The aerosol clouds generated by pharmaceutical inhalers typically comprise a combination of API(s) and other excipients or components, but it is the size distribution of the API that influences efficacy. Cascade impaction generates an APSD specifically for the API to meet this informational need.


3. Cascade impactors capture the entire dose

Cascade impactors, unlike other sizing techniques, which just provide a snapshot of part of the dose, capture the entire dose allowing complete characterisation of the aerosol under test.

The pharmacopoeias recommend a number of commercially available impactors for the routine testing of OINDPs including the Next Generation Impactor (NGI) and the Andersen Cascade Impactor (ACI), both of which are used globally for the testing of metered-dose inhalers (MDIs), dry powder inhalers (DPIs) and soft mist inhalers (SMIs).








How Does a Cascade Impactor Work?

Airborne particles are directed towards the surface of the collection plate for a particular stage.



APSD Data Analysis Software: Inhalytix®

Following sampling, the particle mass collected at each stage is recovered using a suitable solvent and analysed, usually via High Pressure Liquid Chromatography (HPLC) to determine the amount of drug present. Following chemical analysis, important metrics such as Fine Particle Dose, Fine Particle Fraction, Mass Median Aerodynamic Diameter and Geometric Standard Deviation must be calculated from the raw assay data to characterise APSD in accordance with pharmacopoeial and/or regulatory guidance. Where do you start?

Inhalytix is a fully validated, end-to-end APSD data management solution designed to automate the complex transformation of raw inhaler testing data into performance-defining metrics. Accepting data from standard and customised impactors

and impingers, Inhalytix harnesses the power of advanced algorithms to seamlessly transform raw data into actionable insights, supporting analysts in making informed decisions.









Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.

Other Considerations



Impactor Mensuration

Stage mensuration replaces the need for repetitive calibration using standardised aerosols and ensures that only impactors conforming to specification are used in testing. It involves individually inspecting every jet on every stage of the impactor to ensure compliance. Mensuration should be conducted at least annually to ensure the impactor remains in conformance and to assess if any stages might require intervention or repair.

All cascade impactors (including induction ports and preseparators), supplied by Copley, are checked at every stage of manufacture using the very latest in metrology equipment and are provided with a mensuration certificate prior to release.

To find out more about our Servicing options, please see page 312.



Impactor Leak Testing

The ability of a cascade impactor to accurately size separate particles relies on maintaining a fixed volumetric flow rate of air through it. Leaks between impactor stages that allow air to become entrained into the impactor from the outside can modify this flow rate and cause incorrect particle sizing. Performing a leak test prior to each test is recommended to ensure data integrity.

To find out more about our Impactor Leak Testing Kit, please see page 315.




Impactor Cleaning

Cascade impactors are precision instruments and should be treated with care. Regular cleaning and drying is an essential element of good impactor practice and ensures that the instrument is free of product residue and debris prior to testing and that the unit remains in optimum condition throughout its life.

To find out more about our Impactor Cleaning System, please see page 306.



Types of Cascade Impactor



Next Generation Impactor (NGI)

The NGI is a high performance, precision cascade impactor suitable for the APSD characterisation of all types of OINDPs. Ideal for testing at all flow rates specified in the relevant pharmacopoeias, the highly flexible NGI is the cascade impactor of choice for many laboratories throughout the world.



Meets and exceeds all Ph.Eur. and USP specifications



Low inter-stage wall losses for good drug recovery (mass balance)



Seven stages; five with cutoffs between 0.54 and 6.12 microns at flow rates from 30 to 100 L/min



Electrically conductive; unaffected by static



Excellent stage efficiency (GSD <1.2), accuracy and reproducibility




User friendly design for maximum throughput and easy automation


NGI: Key Features






The sample-laden air flow passes through the NGI in a saw-tooth pattern across stages arranged in a horizontal plane.






NGI Principle of Operation

Schematic of Seal Body Showing Orientation of the Various Stages



NGI Cut-Off Diameters

	15	30	60	100	L/min
Stage 1	14.10	11.72	8.06	6.12	microns
Stage 2	8.61	6.40	4.46	3.42	microns
Stage 3	5.39	3.99	2.82	2.18	microns
Stage 4	3.30	2.30	1.66	1.31	microns
Stage 5	2.08	1.36	0.94	0.72	microns
Stage 6	1.36	0.83	0.55	0.40	microns
Stage 7	0.98	0.54	0.34	0.24	microns



NGI: Component Parts

A number of supporting component parts are required in addition to the NGI itself:



NGI Induction Port

Manufactured from 316 stainless steel, the tapered and hardened outlet of the NGI Induction Port provides an airtight seal with the inlet to Stage 1 and the mouthpiece adapter.

NGI Preseparator

The NGI requires the use of a preseparator when used with DPIs in order to catch any powder boluses and large non-inhalable particles. Offering high capacity, high efficiency, two-stage separation, the NGI Preseparator provides a sharp and reproducible cut-point of between 10 and 15 microns depending on flow rate.





Filter Holder

In most cases, the MOC eliminates the need for a final paper filter, having an 80% collection efficiency of 0.3 micron particles at 30 L/min. If ultra-fine particles are present and at flow rates below 30 L/min, then an internal or external filter holder can be used.

Sample Collection Cups

Four special types of sample collection cups are available in addition to those supplied as standard with the NGI:

Gravimetric Cup - for APSD determinations based on sample weight

Deep Cup - to bypass a stage, obviating impaction

Exhaust Cup - to bypass a downstream portion of the impactor

Glass Disc Cup - for Malvern Panalytical Morphologi system



NGI: Accessories



NGI Cup Rack

For the convenient storage of a full set of NGI Cups, protecting the critical surfaces from inadvertent damage and dust collection when not in use.

NGI Carrying/Wash Rack

For transporting the NGI system components around the laboratory and storing them, protecting the critical surfaces from damage and scratches. The rack is also designed to hold the components in place when using our Impactor Cleaning System.






Rinsing Caps

Silicone Rubber and 316 Stainless Steel Rinsing Caps are available for capping off the open ends of the NGI Induction Port and the NGI Preseparator during manual and automated drug recovery.

Storage Cabinet for Impactor Collection Trays

Accommodates up to six NGI Collection Cup Trays (NGI Collection Cup Trays not included).





All NGIs supplied by Copley are machined to the same precision tolerances to guarantee reproducibility between impactors. Each NGI is supplied with a full stage mensuration report (system suitability).

Recommended annually, NGI stage mensuration replaces the need for repetitive, difficult and typically unreliable calibration and ensures that only impactors conforming to specification are used in testing.

For more information on our Servicing options, see page 310.

Further details regarding the design and archival calibration of the NGI can be found in the Journal of Aerosol Medicine Volume 16(3), 2003 and Volume 17(4), 2004.

NGI: Technical Specifications

Flow Rate Range	15 - 100 L/min
Particle Size Range	0.24 - 14.1 microns (dependent on flow rate)
Number of Stages	7
Operation Method	Impaction
Inter-Stage Losses	Low (<5%)
Method of Drug Assay	Chemical analysis - HPLC - Ultra Performance Liquid Chromatography (UPLC) - Infrared Spectroscopy (IR)
Material(s) of Construction	Nickel Plated Aluminium or 316 Stainless Steel

Next Generation Impactor (NGI)

Impactors

Cat. No.	Description
5201	Next Generation Impactor (NGI)
5201A	NGI+ Next Generation Impactor
5202	NGI+ Next Generation Impactor Upgrade

Component Parts

Induction Ports

5203	NGI Induction Port
5239	FRS Flow Meter Adapter
5238	DFM Flow Meter Adapter

Preseparators for testing DPIs

5204	NGI Preseparator (Nickel Plated Aluminium)
5204A	NGI Preseparator with Stainless Steel Insert

Filter Holders

5206	Internal Filter Holder
5210	External Filter Holder
5240	Box of 100 Filters (for Internal/External Filter Holder)

Sample Collection Cups

5243A	Deep Cup, Small (to bypass a stage,
	obviating impaction)
5242A	Malvern Glass Disc Cup, Small (for Malvern
	Panalytical Morphologi system)
5243	Exhaust Cup, Small (to bypass downstream stages
	of impactor)
5241	Gravimetric Cup Small (for APSD determinations
	based on weight)
5241A	Pack of 100 Filters for Small and Large Gravimetric Cup
5244	Gravimetric Cup Large (for APSD determinations
	based on weight)

Accessories

Cat. No.	Description
5222	NGI Collection Cup Rack
5205	NGI Carrying/Wash Rack
5265	Set of 2 Silicone Rubber Rinsing Caps
	for NGI Induction Port
5266	Set of 2 Silicone Rubber Rinsing Caps
	for NGI Preseparator
5227	Set of 2 Stainless Steel Rinsing Caps
	for NGI Induction Port
5228	Set of 2 Stainless Steel Rinsing Caps
	for NGI Preseparator
5232	Set of 2 Silicone Rubber Stoppers
	for NGI I.P./Preseparator
5224	Storage Cabinet for Impactor Collection Trays

NGI Cooler™

5009	NGI Cooler
5011	NGI Cooler Qualification Documentation
5012	NGI Cooler Qualification Tools
5013	Re-calibration of NGI Cooler Qualification Tools

Spare Parts

5208	Collection Cup Tray
5209	Set of 8 Collection Cups (2 Large, 6 Small)
5245	Welded Cup Tray Manifold
5211	Set of 18 Seals for the Next Generation Impactor
5246	Set of 10 Seals for the NGI Preseparator
5247	Set of 10 Seals for the NGI Internal Filter Holder
5248	Set of 10 Seals for the NGI External Filter Holder
5249	NGI Outlet Diameter Reducing Adapter





Andersen Cascade Impactor (ACI)

Well-established and readily accepted by the regulatory authorities, the ACI has been used for the APSD characterisation of OINDPs for over 30 years.



Meets and exceeds all Ph.Eur. and USP specifications



Low flow resistance at high flow rates when Stages 6 & 7 are removed



60 and 90 L/min Conversion Kits available for high flow rate testing, whilst retaining the 28.3 L/min cut-off diameters



Electrically conductive; unaffected by static



Reduced stack option for work with nasal aerosols and sprays





ACI: Materials of Construction

316 Stainless Steel

Superior corrosion resistance and durability to extend impactor life.

Titanium


Lightweight handling, superior corrosion resistance.

Aluminium

Lightweight, lower cost, where corrosion resistance is not an issue.



When used at calibration flow rates, cascade impactors can be operated at different angles which may be useful when testing device performance at different positions.



Unlike the NGI, the stages of the ACI are arranged vertically. The aerosol flow passes first through the stage at the top of the impactor, through to the last stage and a final filter at the bottom of the impactor arrangement.

ACI: Modified Configurations

The standard ACI is designed for use at 28.3 L/min. In some cases (particularly with low resistance DPIs), it is necessary to operate at flow rates greater than 28.3 L/min, if a pressure drop over the inhaler of 4 kPa is to be achieved. However, it is important to consider the

change in cut-points that would occur for each stage with any change to the flow rate. We offer two modified configurations of the ACI for operation at calibrated flow rates of 60 and 90 L/min to help address this.



ACI: Component Parts

A number of supporting component parts are required in addition to the ACI itself:



USP Induction Port

Provides an airtight seal achieved between the ACI inlet and the mouthpiece adapter.

ACI Preseparator

Designed to collect the large mass of non-inhalable powder boluses emitted from powder-based inhalers prior to their entry into the impactor, the ACI Preseparator is ideal for DPI testing applications. Preseparators are available for testing at 28.3, 60 and 90 L/min.



ACI: Accessories



ACI Quick Clamp

Constructed from stainless steel, the ACI Quick Clamp enables quick and efficient adjustment of the ACI plate stack.

ACI Collection Plate Rack

For the convenient storage of the ACI collection plates, protecting the critical collection surfaces from inadvertent scratches and dents when not in use.



ACI: Accessories



ACI Carrying/Wash Rack

Constructed from heavy duty polypropylene and fitted with neoprene cushions, the ACI Carrying/Wash Rack is ideal for transporting the ACI system components around the laboratory and storing them, protecting the critical surfaces from damage and scratches. The rack is also designed to hold the components in place when used with our Impactor Cleaning System.

Rinsing Caps

Silicone Rubber Rinsing Caps are available for capping off the open ends of the ACI Induction Port during manual and automated drug recovery.



Storage Cabinet for Impactor Collection Trays

Accommodates up to six NGI Collection Cup Trays (NGI Collection Cup Trays not included).





between impactors. Each ACI is supplied with a full stage mensuration report (system suitability).

ACI: Technical Specifications

Flow Rate Range	28.3 L/Min Modified configurations: Conversion kits for 60 L/Min and 90 L/Min available
Particle Size Range	0.4 - 9.0 microns (28.3 L/Min) 0.3 - 8.6 microns (60 L/Min) 0.2 - 8.0 microns (90 L/Min)
Number of Stages	8
Operation Method	Impaction
Inter-Stage Losses	Low to High (depending on product)
Method of Drug Assay	Chemical analysis - HPLC - UPLC - IR
Material(s) of Construction	Aluminium, 316 Stainless Steel or Titanium

Andersen Cascade Impactor (ACI)

Impactors

Cat. No.	Description
8301	28.3 L/Min Andersen Cascade Impactor*
8301-60	60 L/Min Andersen Cascade Impactor*
8301-90	90 L/Min Andersen Cascade Impactor*

Conversion Kits for the standard 28.3 L/min ACI

8318	Conversion Kit for 60 L/min operation*
8319	Conversion Kit for 90 L/min operation*

Component Parts

Induction Ports

8501	USP Induction Port*	
8510	USP Induction Port (One-piece 316 Stainless Steel)	
5239	FRS Flow Meter Adapter	
5238	DFM Flow Meter Adapter	

Preseparators for testing DPIs

8401	28.3 L/min Preseparator*
8420	60 L/min Preseparator*
8420-90	90 L/min Preseparator*

Accessories

Cat. No.	Description
5212	'Quick Clamp' for Andersen Cascade Impactor
5441	ACI Collection Plate Rack
5401	ACI Carrying/Wash Rack
5224	Storage Cabinet for Impactor Collection Trays

Accessories

Cat. No.	Description
Rinsing Caps	
8504	Set of 2 Silicone Rubber Rinsing Caps
	for ACI Induction Port

Spare Parts

8307	Complete Set of 13 ACI Silicone Rubber O-Rings
8314	Set of 8 Stainless Steel Collection Plates (28.3 L/min)
8314-60	Set of 8 Stainless Steel Collection Plates (60 L/min)
8314-90	Set of 8 Stainless Steel Collection Plates (90 L/min)
8316	Box of 100 Glass Fibre Filters
8306	Set of 6 O-Rings for Spring Clamp
8308	Set of 3 Spring Clamps
8309	Set of 3 PVC End Caps for Spring Clamps
8403	Set of 4 O-Rings for Preseparator
8395	ACI Carrying Case
8351	Inlet Cone*
8352	Stage -2A*
8353	Stage -1A (for 90 L/min operation)*
8354	Stage -1 (for 60 L/min operation)*
8355	Stage -0*
8356	Stage 0*
8357	Stage 1*
8358	Stage 2*
8359	Stage 3*
8360	Stage 4*
8361	Stage 5*
8362	Stage 6*
8363	Stage 7*
8364	Stage F (Filter)*
8365	Base (including Hose Fitting)*
*Please spe	ecify Aluminium (A), 316 Stainless Steel (S) or Titanium (T

 $^{^*}$ Please specify Aluminium (A), 316 Stainless Steel (S) or Titanium (T) when placing your order.





Multi-Stage Liquid Impinger (MSLI)

A traditional apparatus for routine testing and research applications in industry and academia, the MSLI comprises four impaction stages and a final filter stage. Whilst it does not offer the number of stages of the ACI or NGI, it has virtually no inter-stage losses.

Also, unlike the ACI and NGI, the collection stages of the MSLI are kept moist, which eliminates the problem of particle bounce associated with conventional impactors.



Ph.Eur. Chapter 2.9.18 compliant for MDIs and DPIs



Eliminates particle bounce and re-entrainment problems



Choice of construction materials to suit all budgets and needs



Quick and easy to mensurate



Virtually no inter-stage losses





MSLI: Materials of Construction

316 Stainless Steel

durability to extend impactor life.

Titanium


Lightweight handling, superior corrosion resistance.

Aluminium

Lightweight, lower cost, where



A stage mensuration certificate and leak test certificate are included with each MSLI as standard. During the mensuration, the sintered glas impingement stages are positioned using calibrated gauge blocks to ensure that the correct iet-to-plate distance is maintained.



The aerosol stream is drawn into the top of the MSLI, passing first through Stage 1 which acts as a preseparator. Particles with sufficient inertia will impact on the moist surface of the sintered glass disc. Those with insufficient inertia will pass through to Stage 2. The same process of impaction and particle selection takes place until the final filter stage (Stage 5), which captures any remaining fine particles.

The cut-off diameters for the relevant stages at a volumetric flow rate of 60 L/min are shown in the table below.

MSLI Cut-Off Diameters

	60	L/Min
Stage 1	13.0	microns
Stage 2	6.8	microns
Stage 3	3.1	microns
Stage 4	1.7	microns
Stage 5 (Filter)	< 1.7	microns

MSLI: Technical Specifications

Flow Rate Range	Between 30 and 100 L/min
Particle Size Range	1.7 - 13.0 microns (dependent on flow rate)
No. of Stages	4
Operation Method	Impingement
Inter-Stage Losses	Zero
Method of Drug Assay	Chemical Analysis - HPLC - UPLC - IR
Material(s) of Construction	Aluminium, 316 Stainless Steel or Titanium

Multi-Stage Liquid Impinger (MSLI)

Cat. No.	Description	
8801	Multi-Stage Liquid Impinger (MSLI)*	
8501	USP Induction Port*	
8510	USP Induction Port (One-piece 316 Stainless Steel)	
5239	FRS Flow Meter Adapter	
5238	DFM Flow Meter Adapter	
Options		
8851	Torque Adjuster for MSLI	

Spare	Parts	

8805	Set of 3 O-Rings
8807	Set of 8 Inter-Stage PTFE Gaskets (Code M)
8814	Filter Support Plate (Code S)
8834	Pack of 10 Silicone Rubber Stoppers
8839	Pack of 100 Glass Fibre Filters
8840	Ground Glass Cylinder (Code E)
8844	Set of 4 Sintered Glass Discs (Code D)

^{*} Please specify Aluminium (A), 316 Stainless Steel (S) or Titanium (T) when placing your order.







Glass Twin Impinger (GTI)

Retained as Apparatus A in Ph.Eur. 2.9.18 due to its value as a simple and inexpensive routine quality control tool, the two-stage GTI is ideal for use where batch-to-batch variability in FPD is required and a coarser test may be acceptable.

Its usage is typically restricted to the assessment of nebulisers, MDIs, nasal sprays and DPIs where it can be demonstrated that a flow rate of 60 (+/-5) L/min is suitable.



Ph.Eur. 2.9.18 compliant (Apparatus A)



Regular mensuration is not required





No inter-stage losses



Ideal for routine quality control applications

GTI Key Features:





The GTI operates on the principle of liquid impingement to divide the dose emitted from the inhaler into respirable and non-respirable portions.

Prior to testing, 7 mL of solvent is typically dispensed into the upper impingement chamber and 30 mL to the lower impingement chamber.

The upper impingement chamber (stage 1) is designed such that at a flow rate of 60 L/min through the impinger, the particle cut-off is 6.4 microns. Particles smaller than 6.4 microns pass into the lower impingement chamber (stage 2).

After the test is complete, the active drug collected in the lower impingement chamber is assayed and expressed as a respirable fraction (or percentage) of the delivered dose.

GTI: Technical Specifications

Flow Rate Range	60 L/Min
Particle Size Range	6.4 microns only
Number of Stages	1
Operation Method	Impingement
Inter-Stage Losses	Zero
Method of Drug Assay	Chemical Analysis - HPLC - UPLC - IR
Material(s) of Construction	Glass

Glass Twin Impinger (GTI)

Cat. No.	Description	Spare	Parte
8901	Glass Twin Impinger	Spare	raits
8999	Modification for Nasal Sprays (acc. to Aiache & Beyssac)	8906	Coupling Tube (Ph.Eur. Code E)
8920	FRS Flow Meter Adapter for GTI/FP Ind Port	8907	Screwthread Side-Arm Adapter (Ph.Eur. Code F)
	. ,	8912	Lower Jet Assembly (Ph.Eur. Code G)
		8908	Lower Impingement Chamber (Ph.Eur. Code H)
Spare P	arts	8909	Throat Flow Meter Adapter (Ph.Eur. Code I)
•		8910	Vacuum Pump Adapter (Ph.Eur. Code J)
8903	Throat (Ph.Eur. Code B)	8913	Set of 2 Conical Joint Clips (Yellow)
8904	Neck (Ph.Eur. Code C)	8914	Set of 4 Conical Joint Clips (Green)
8905	Upper Impingement Chamber (Ph.Eur. Code D)	8916	Spare Set of Glassware (incl. clips and Lower
			Jet Assembly)

Technical Specifications: Comparison Summary



Choose your Impactor

Device Type		NGI	ACI	MSLI	GTI	Pharmacopoeia
		Υ	Υ	Υ	Υ	Ph. Eur./EMA
MDI		Υ	Υ	N	N	USP/FDA
		Υ	Υ	N	Υ	ChP
		Y	Y	Y	N	JP
		Υ	Υ	Υ	Y	Ph. Eur./EMA
MDI with a Spacer/		Υ	Υ	N	N	USP/FDA
Valved Holding Chamber (VHC)		Υ	Υ	N	Υ	ChP
		Υ	Υ	N	N	JP
		Υ	Υ	Υ	Υ	Ph. Eur./EMA
DPI		Υ	Υ	Υ	N	USP/FDA
DPI		Υ	Υ	N	Υ	ChP
		Υ	Υ	Υ	N	JP
		Υ	N	N	N	Ph. Eur./EMA
Nebuliser		Υ	N	N	N	USP/FDA
Nebuliser		Υ	N	N	N	ChP
		Υ	N	N	N	JP
		Υ	Υ	N	N	Ph. Eur./EMA
SMI		Υ	Υ	N	N	USP/FDA
SMI		Υ	Υ	N	N	ChP
		Υ	Υ	N	N	JP
		Υ	Υ	N	N	Ph. Eur./EMA
Nasal Products	A	Υ	Υ	N	N	USP/FDA
FIOUUCUS	U	Υ	Υ	N	N	ChP
		Υ	Υ	N	N	JP



Aerodynamic Particle Size Distribution

Metered Dose Inhalers (MDIs)

The APSD testing of MDIs is typically performed at a flow rate of 28.3 L/min when using an ACI or 30 L/min when using an NGI. For Breath Actuated MDIs (BAIs) a Breath Actuation Controller may also be used to generate a time delay.

There is no requirement for a preseparator in MDI measurement. Plate and/or cup coating may be used to prevent particle bounce and re-entrainment, but is generally not required if the formulation includes a surfactant. Multiple doses are typically required to


achieve analytical sensitivity.

For further information on the APSD testing of MDIs with a Spacer or Valved Holding Chamber (VHC), see page 111.

Regulations and Guidelines


	Organisation	Chapter/Guidance		
	Ph. Eur. / EMA	2.9.18 App E		
	USP / FDA	<601> App 6		
NGI	ChP	<0951> App 3		
	JP	6.15.5 App 3		
oÆ0	Organisation	Chapter/Guidance		
\=-	Ph. Eur. / EMA	2.9.18 App D		
	USP / FDA	<601> App 1		
 #	ChP	<0951> App 2		
ACI	JP	6.15.5 App 2		
files	Organisation	Chapter/Guidance		
	Organisation Ph. Eur. / EMA	Chapter/Guidance 2.9.18 App C		
	-			
	Ph. Eur. / EMA	2.9.18 App C		
₩SLI	Ph. Eur. / EMA USP / FDA	2.9.18 App C		
MSLI	Ph. Eur. / EMA USP / FDA ChP	2.9.18 App C <601> App 1 -		
MSLI	Ph. Eur. / EMA USP / FDA ChP JP	2.9.18 App C <601> App 1 - 6.15.5 App 1		
MSLI	Ph. Eur. / EMA USP / FDA ChP JP Organisation	2.9.18 App C <601> App 1 - 6.15.5 App 1 Chapter/Guidance		
MSLI	Ph. Eur. / EMA USP / FDA ChP JP Organisation Ph. Eur. / EMA	2.9.18 App C <601> App 1 - 6.15.5 App 1 Chapter/Guidance		

APSD of MDIs Manual Test System Set-Up




- Mouthpiece Adapter
- B Induction
- Alternative Impactors/Impingers

- Next Generation
 Impactor NGI
- D Flow Rate
- Breath Actuation Controller
- F Vacuum Pump





Related Accessories



Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 179.

MDI Actuation Sensor/Footswitch

Suitable for most commercially available MDI canisters, the MDI Actuation Sensor connects directly to the Breath Actuation Controller BAC 200i to ensure precise synchronisation of MDI actuation. Alternatively, a Footswitch can be attached to trigger actuation.

See page 179.

APSD of MDIs: Manual Test System Component Parts



Next Generation Impactor NGI

The recommended test set-up is with an NGI, but an ACI may also be used. Impactors with 7 or 8 stages are preferred by regulators, as they provide good APSD resolution. However, for some established methods the MSLI or GTI may be acceptable.









In addition to the above, the following is needed to complete a fully-operational test set-up for APSD measurement of MDIs:

Vacuum Pump

Designed for optimal operation at the flow rates required for MDI testing, the Low Capacity Pump LCP6 represents the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.

See page 188 for further information about our Vacuum Pump range.













Breath Actuation Controller



Ensuring that the volume of air sampled does not exceed the pharmacopoeial specifications, the Breath Actuation Controller BAC 100i contains an electronically operated, timer-controlled two-way solenoid valve and is positioned between the impactor and Vacuum Pump to control air flow through the inhaler.

See page 172 for further information about our Flow Controller range.













Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias. See page 184 for further information about flow rate measurement.









Inhaler Testing Workstation™

Designed to keep the apparatus organised during testing and improve workflow efficiency, the Inhaler Testing Workstation ITW holds the casacde impactor and flow meter in position throughout the testing process.

See page 204 for further information.

Recommended for:









Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus. For a list of available Mouthpiece Adapters see page 211.

Custom Mouthpiece Adapters are available upon request.









Oualification

GMP regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- · Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.



APSD of MDIs

Automated Test System Set-Up

The Vertus® III automated shake, fire and shot waste range is made up of integrated turn-key solutions for precise, controlled and reproducible MDI testing.

Compatible with most MDIs, the Vertus III systems offer analysts complete control over:

- The speed, angle and duration of shaking, ahead of actuation
- Firing force and the speed of application and release of that force
- The time delay between the end of shaking and device actuation



Improve inhaler testing accuracy and reproducibility



Increase productivity and reduce hassle



Replicate test methods across different sites with ease



Reduce handling errors and costly out-ofspecification results



Replaces the need for:





Breath Actuation Controller



Inhaler Testing Workstation™

See page 284 for further information about the Vertus III range.


APSD Data Analysis Software: Inhalytix®



Inhalytix is a fully validated, end-to-end APSD data management solution for the entry, analysis and reporting of APSD data for all inhaled products.

Automating the complex transformation of raw inhaler testing data into performance-defining metrics, Inhalytix accepts data from both standard and customised impactors and impingers.







Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.


Automation Tools





Reduce variability








Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis. See page 298.







Sample Preparation Unit SPU 200i

Simplifies and standardises drug recovery from induction ports and preseparators. See page 302.









Impactor Cleaning System

Standardises cleaning and drying procedures to help ensure the NGI and ACI remain in optimum condition throughout their life. See page 306.





Related Applications

We also offer a range of equipment for additional MDI testing application support:



For better *in vitro-in vivo* correlation (IVIVC) testing

See page 222



For cold Freon® effect testing

See page 255



For USP product-specific monographs

See page 270

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320




Aerodynamic Particle Size Distribution

MDIs with a Spacers/VHC

Due to the potential opportunity for particle expansion, impaction and deposition within the chamber of add-on devices such as spacers or VHCs, the APSD characteristics may be substantially altered from what is emitted when the MDI is used alone. This potential for change must be appropriately assessed.

Regulations and Guidelines



In Section 3 of USP Chapter <1602> Spacers and Valved Holding Chambers used with Inhalation Aerosols, two tests are specified relating to the APSD characterisation of add-on devices used with the MDIs:

Test 3.1


Designed to measure the APSD from the spacer/VHC when used under optimal conditions, that is, with no delay following actuation of the inhaler. Direct comparisons can then be made between the APSD produced by the MDI both with and without the add-on device.

Test 3.2

For testing VHCs only and designed to measure the APSD from the VHC when used under "worst case" conditions, i.e. with a delay of 2 or more seconds between inhaler actuation and patient inspiration.

The delay can be simulated by placing a timer controlled two-way solenoid valve such as the Breath Actuation Controller BAC 100i between the impactor and the pump.

APSD of MDIs with a Spacer/VHC Test System Set-Up





Considering

the effects of environmental variability?



Our environmental control solutions are designed to help improve the accuracy, sensitivity and reproducibility of test data.

Learn more on page 196.

Alternative Impactors



Related Accessories



Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 179.

MDI Actuation Sensor/Footswitch

Suitable for most commercially available MDI canisters, the MDI
Actuation Sensor connects directly to the Breath Actuation Controller
BAC 100i to ensure precise synchronisation of MDI actuation.
Alternatively, a Footswitch can be attached to trigger actuation.
See page 179.

APSD of MDIs with a Spacer/VHC: Test System Component Parts



Andersen Cascade Impactor ACI

If the spacer/VHC is intended for adults, then the standard ACI or NGI should be used with a suitable vacuum pump capable of producing 28.3 or 30 L/min respectively. If the add-on device is intended for neonates, infants or small children, then only the NGI should be used as this can be used at the lower flow rate of 15 L/min.



In addition to the above, the following is needed to complete a fully-operational test set-up for the APSD measurement of MDIs with a spacer or VHC:

Vacuum Pump

Designed for optimal operation at the low flow rates required for MDI testing, the Low Capacity Pump LCP6 represents the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.

See page 188 for further information about our Vacuum Pump range.









Breath Actuation Controller



Ensuring that the volume of air sampled does not exceed the pharmacopoeial specifications, the Breath Actuation Controller BAC 100i contains an electronically operated, timer-controlled two-way solenoid valve and is positioned between the impactor and Vacuum Pump to control air flow supply to the inhaler.

See page 172 for further information about our Flow Controller range.







Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.

See page 184 for further information about flow rate measurement.





Inhaler Testing Workstation™

Designed to keep the apparatus organised during testing and improve workflow efficiency, the Inhaler Testing Workstation ITW keeps the cascade impactor and flow meter in position throughout the testing process.

See page 204 for further information.

Recommended for:







Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler/add-on device combination under test and the test apparatus. For a list of available Mouthpiece Adapters see page 211.

Custom Mouthpiece Adapters are available upon request.

Required for:






Oualification

GMP regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- · Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.




APSD Data Analysis Software: Inhalytix®



Inhalytix is a fully validated, end-to-end APSD data management solution for the entry, analysis and reporting of APSD data for all inhaled products.

Automating the complex transformation of raw inhaler testing data into performance-defining metrics, Inhalytix accepts data from both standard and customised impactors and impingers.







Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.


Automation Tools






Reduce variability







Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis. See page 298.







Sample Preparation Unit SPU 200i

Simplifies and standardises drug recovery from induction ports and preseparators. See page 302.









Impactor Cleaning System

Standardises cleaning and drying procedures to help ensure the NGI and ACI remain in optimum condition throughout their life. See page 306.





Related Applications

We also offer a range of equipment for additional application testing support:



For facemask testing
See page 244

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320





Aerodynamic Particle Size Distribution

Dry Powder Inhalers (DPIs)

The APSD measurement of DPIs is typically performed under the same conditions as DDU testing. However there are some differences.

A preseparator is typically interposed between the induction port and stage 0 of cascade impactor to capture the large, non-inhalable carrier particles, to prevent impactor over-loading.

As for delivered dose testing of DPIs, test flow rate is set on the basis of a 4 kPa pressure drop across the

device, to approximate the mean patient inhalation flow rate achieved during clinical use.


Cup-coating should be considered and validated as part of method development to reduce particle bounce and re-entrainment.

Regulations and Guidelines

	Organisation	Chapter/Guidance
	Ph. Eur. / EMA	2.9.18 App. E
	USP / FDA	601 App. 5
NGI	ChP	<0951> App. 3
	JP	6.15.5 App 3
	Organisation	Chapter/Guidance
	Ph. Eur. / EMA	2.9.18 App. D
4	USP / FDA	601 App. 2
	ChP	<0951> App. 2
ACI	JP	6.15.5 App 2
	Organisation	Chapter/Guidance
	Ph. Eur. / EMA	2.9.18 App. C
I I Ø	USP / FDA	601 App. C
	ChP	-
MSLI	JP	6.15.5 App 1
	Organisation	Chapter/Guidance
	Ph. Eur. / EMA	2.9.18 App. A
1 1/2		
	USP / FDA	-
	USP / FDA ChP	- <0951> App. 1

APSD of DPIs

Test System Set-Up



- Mouthpiece Adapter
- B Induction
- C Preseparator
- D Next Generation Impactor NGI
- Flow Rate Sensor
- F Breath Actuation Controller
- G Vacuum Pump







Our environmental control solutions are designed to help improve the accuracy, sensitivity and reproducibility of test data.

Learn more on page 196.

Alternative Impactors/Impingers



Related Accessories



Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 183.

Footswitch

Connecting directly to the Critical Flow Controller TPK 100i, the Footswitch enables precise synchronisation of DPI actuation with the onset of flow. See page 183.

APSD of DPIs: Test System Component Parts



Next Generation Impactor NGI

The recommended test set-up is with an NGI, but an ACI may also be used. Impactors with 7 or 8 stages are preferred by the regulators, as they provide good APSD resolution. However, for some established methods the MSLI or GTI may be acceptable.









FORGET



Preseparator

For the collection of large mass, non-inhalable powder boluses typically emitted from a DPI, prior to entry into the impactor. Different preseparators are available for the NGI and ACI.

See pages 87 and 93 respectively.

Note: Preseparators are not required for APSD testing of DPIs using an MSLI or GTI.





In addition to the above, the following is needed to complete a fully-operational test set-up for the APSD measurement of DPIs:

Vacuum Pump

Ideal for the higher, sonic flow rate testing requirements of DPIs, the High Capacity Pump HCP6 and Super Capacity Pump SCP6 represent the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.

See page 188 for further information about our Vacuum Pump range.



















Induction Port P1 Measurement Adapter



Critical Flow Controller

Simplify DPI test system set-up in accordance with pharmacopoeial recommendations with the Critical Flow Controller series. Positioned between the impactor and vacuum pump, the Critical Flow Controller TPK 100i ensures critical (sonic) flow conditions during testing. It measures and records all parameters required for testing and for controlling flow conditions.

See page 172 for further information about our Flow Controller range.







Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.

See page 184 for further information about flow rate measurement.















Inhaler Testing Workstation™

Designed to keep the apparatus organised during testing and improve workflow efficiency, the Inhaler Testing Workstation ITW keeps the cascade impactor and flow meter in position throughout the testing process.

See page 204 for further information.

Recommended for:





Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus. For a list of available Mouthpiece Adapters see page 211.

Custom Mouthpiece Adapters are available upon request.












Oualification

GMP regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- · Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.



APSD Data Analysis Software: Inhalytix®



Inhalytix is a fully validated, end-to-end APSD data management solution for the entry, analysis and reporting of APSD data for all inhaled products.

Automating the complex transformation of raw inhaler testing data into performance-defining metrics, Inhalytix accepts data from both standard and customised impactors and impingers.





Meaningful metrics,



Secure and

Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.

Automation Tools



Improve efficiency



Reduce variability



Eliminate handling errors



Increase testing capacity



Impactor Coater™ IC 200i

Standardises impaction surface coating for both NGI Collection Cups and ACI Collection Plates. See page 296.

Recommended for:







Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis. See page 298.

Recommended for:







Impactor Genie™ IG 200i

An innovative 2-in-1 solution combining the coating capabilities of the Impactor Coater IC 200i with the drug recovery features of the Gentle Rocker GR 200i. See page 300.

Recommended for:





Sample Preparation Unit SPU 200i

Simplifies and standardises drug recovery from the Induction Ports and Preseparators. See page 302.

Recommended for:









Impactor Cleaning System

Standardises cleaning and drying procedures to help ensure the NGI and ACI remain in optimum condition throughout their life. See page 306.





Related Applications

We also offer a range of equipment for additional DPI testing application support:



For better in vitro-in vivo correlation (IVIVC) testing

See page 222



For USP product-specific monographs

See page 270

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



Training See page 321



Servicing See page 310



Support See page 320



Design See page 320



Aerodynamic Particle Size Distribution


Nebulisers

For devices such as nebulisers, the evaporation of droplets exacerbated by the thermal mass of the impactor can be a problem, especially for drugs in solution.

Loss of solvent reduces droplet size, producing artificially low APSD measurements, compromising the integrity of the resulting data. Cooling the impactor to approximately 5°C is the recommended method for overcoming this problem.


The recommended flow rate of 15 L/min employed in the APSD testing of nebulisers is lower than that of other OINDPs in order to better represent the tidal breathing conditions employed in their use.

Regulations and Guidelines



Organisation	Chapter/Guidance
Ph. Eur. / EMA	2.9.44 App. E
USP / FDA	<1601> App. 6
ChP	0951 App. 3
JP	-

APSD of Nebulisers Test System Set-Up




- A Mouthpiece Adapter
- (B) Induction Port
- Next Generation Impactor NGI
- D NGI Cooler™
- E Flow Rate Sensor
- Breath Actuation
 Controller
- G Vacuum Pump



Determine sampling time (T_o) by balancing the risk of impacto overload with the requirement for analytical sensitivity. Time chosen should be sufficient to ensure an adequate sample is collected for analysis without overloading the collection cups, which causes liquid streaking.

Related Accessories



Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 179.

Footswitch

Connecting directly to the Breath Actuation Controller BAC 100i, the Footswitch enables precise synchronisation of nebuliser device actuation with the onset of flow. See page 179.



Our environmental control solutions are designed to help improve the accuracy, sensitivity and reproducibility of test data.

Learn more on page 196.

APSD of Nebulisers: Test System Component Parts



Next Generation Impactor NGI

The APSD characterisation of a nebuliser should be conducted using an NGI. This is because the NGI is calibrated for use at 15 L/min and has collection cups well suited to retaining liquid droplets.

In addition to the above, the following is needed to complete a fully-operational test set-up for the APSD measurement of nebulisers:

Vacuum Pump

Designed for optimal operation at low flow rates required for nebuliser testing, the Low Capacity Pump LCP6 represents the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.







Breath Actuation Controller

Ensuring that the volume of air sampled does not exceed the pharmacopoeial specifications, the Breath Actuation Controller model BAC 100i contains an electronically operated, timer-controlled two-way solenoid valve and is positioned between the impactor and Vacuum Pump to control air flow supply to the nebuliser.

See page 172 for further information about our Flow Controller range.

Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.


See page 184 for further information about flow rate measurement.



NGI Cooler™

Accommodating the NGI both open and closed, the NGI Cooler allows the NGI to be operated in a temperature controlled environment. Additional space allows for cooling of extra sets of Collection Cups, so multiple tests can be undertaken in quick succession.







Filter Holder

In most cases, the MOC eliminates the need for a final paper filter, having an 80% collection efficiency of 0.3 micron particles at 30 L/min. If ultra-fine particles are present and at flow rates below 30 L/min, then an internal or external filter holder can be used.

Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus. For a list of available Mouthpiece Adapters see page 211.


Custom Mouthpiece Adapters are available upon request.



Oualification

GMP regulations require that

- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.




APSD Data Analysis Software: Inhalytix®



Inhalytix is a fully validated, end-to-end APSD data management solution for the entry, analysis and reporting of APSD data for all inhaled products.

Automating the complex transformation of raw inhaler testing data into performance-defining metrics, Inhalytix accepts data from both standard and customised impactors and impingers.







Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.

Automation Tools



Improve efficiency



Reduce variability



handling errors





Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis. See page 298.







Sample Preparation Unit SPU 200i

Simplifies and standardises drug recovery from induction ports and preseparators. See page 302.









Impactor Cleaning System

Standardises cleaning and drying procedures to help ensure the NGI and ACI remain in optimum condition throughout their life. See page 306.





Related Applications

We also offer a range of equipment for additional nebuliser testing application support:



For facemask testing See page 244



For better *in vitro-in vivo* correlation (IVIVC) testing

See page 222

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320





Aerodynamic Particle Size Distribution

Soft Mist Inhalers (SMIs)

For SMIs as for nebulisers, the evaporation of droplets exacerbated by the thermal mass of the impactor can be a problem.


Loss of solvent reduces droplet size, producing artificially low APSD measurements, compromising the integrity of the resulting data. Cooling the impactor to approximately 5°C is the recommended method for overcoming this problem.

Classified as active devices, the recommended flow rate for SMI testing is 28.3 L/min for the ACI or 30 L/min for the NGI.

Regulations and Guidelines

Whilst there is no current pharmacopoeial or regulatory guidance for SMIs, they are considered to combine the metered-dose technology of MDIs with the aqueous

aerosol droplet generation of nebulisers. Testing, and the equipment that features in this section, reflects this combined technology.



APSD of SMIs Test System Set-Up



- Mouthpiece Adapter
- Induction
- Port
- Next Generation Impactor NGI

Flow Rate

- - **Breath Actuation** Controller

NGI Cooler™


Pump



Alternative Impactors/Impingers



Related Accessories



Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 179.

Footswitch

Connecting directly to the Breath Actuation Controller BAC 100i, the Footswitch enables precise synchronisation of SMI device actuation with the onset of flow. See page 179.

APSD of SMIs: Test System Component Parts



Next Generation Impactor NGI

The recommended test set-up is with an NGI. An ACI can also be used for the assessment of SMIs.





In addition to the above, the following is needed to complete a fully-operational test set-up for the APSD measurement of SMIs:

Vacuum Pump

Designed for optimal operation at low flow rates required for SMI testing, the Low Capacity Pump LCP6 represents the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.

See page 188 for further information about our Vacuum Pump range.









Breath Actuation Controller BAC



Ensuring that the volume of air sampled does not exceed the pharmacopoeial specifications, the Breath Actuation Controller BAC 100i contains an electronically operated, timer-controlled two-way solenoid valve and is positioned between the impactor and Vacuum Pump to control air flow supply to the inhaler.

See page 172 for further information about our Flow Controller range.







Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.

See page 184 for further information about flow rate measurement.





NGI Cooler™

Accommodating the NGI both open and closed, the NGI Cooler allows the NGI to be operated in a temperature controlled environment. The NGI Cooler can also be used with the ACI. The additional space inside facilitates the cooling of extra sets of collection cups/multiple ACIs, allowing multiple tests to be undertaken efficiently.

See page 202 for further information about the NGI Cooler.

Required for:









Mouthpiece Adapter

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus. For a list of available Mouthpiece Adapters see page 211.

Custom Mouthpiece Adapters are available upon request.





Qualification

GMP regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- · Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.




APSD Data Analysis Software: Inhalytix®



Inhalytix is a fully validated, end-to-end APSD data management solution for the entry, analysis and reporting of APSD data for all inhaled products.

Automating the complex transformation of raw inhaler testing data into performance-defining metrics, Inhalytix accepts data from both standard and customised impactors and impingers.







Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.

Automation Tools




Improve efficiency



Reduce variability







Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis. See page 298.

Recommended for:







Sample Preparation Unit SPU 200i

Simplifies and standardises drug recovery from induction ports and preseparators. See page 302.









Impactor Cleaning System

Standardises cleaning and drying procedures to help ensure the NGI and ACI remain in optimum condition throughout their life. See page 306.





Related Applications

We also offer a range of equipment for additional SMI testing application support:



For better *in vitro-in vivo* correlation (IVIVC) testing
See page 222

F



For cold Freon® effect testing
See page 255



For USP product-specific monographs See page 270

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320





Aerodynamic Particle Size Distribution

Nasal Sprays

Nasal sprays typically produce droplets in the range 20-200 microns, which is outside the effective range of cascade impactors. However, most sprays deliver a proportion (typically <5%) of fine droplets in the <10 micron range.

It is important to quantify the amount of droplets in this range since it is the amount of dose that can penetrate beyond the nasal tract and into the lower respiratory tract or lungs, which may be undesirable. Regulators recommend the use of a cascade impactor in conjunction with a high volume expansion chamber to quantify the amount of drug in the <10 micron range, to assess the potential risk of deposition in the lungs.

Regulations and Guidelines

	Organisation	Chapter/Guidance
	Ph. Eur. / EMA	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products
	USP / FDA	Nasal Spray and Inhalation Solution, Suspension, and Spray Drug Products — Chemistry, Manufacturing, and Controls Documentation
NGI	ChP	-
	JP	-
0月0	Organisation	Chapter/Guidance
\=/	Ph. Eur / EMA	-
	USP / FDA	Nasal Spray and Inhalation Solution, Suspension, and Spray Drug Products — Chemistry, Manufacturing, and Controls Documentation
	ChP	-
ACI	JP	-
ÇP.	Organisation	Chapter/Guidance
	Ph. Eur. / EMA	-
#	USP / FDA	-
<i>(;</i>)	ChP	-
GTI	JP	-

APSD of Nasal Sprays Automated Test System Set-Up

Routine APSD test set-ups for nasal sprays are easily automated with Vertus® III. Compatible with most nasal sprays, the Vertus III systems simplify the aerodynamic particle size distribution measurement of nasal sprays in accordance with Ph. Eur. Chapter 0676 and USP Chapter <601>.

The Vertus III range offer analysts complete control over:

- The speed, angle and duration of shaking, ahead of actuation
- · Firing force and the speed of application and release of that force
- The time delay between the end of shaking and device actuation

Vertus III+ with Next Generation Impactor NGI and Glass Expansion Chamber





Automated shot weight measurement via an integrated balance

- A Vertus III+ with Balance
- B Glass Expansion Chamber
- B Next Generation Impactor NGI

Alternative Vertus® III Interface Plates



Andersen Cascade Impactor ACI with Glass Expansion Chamber



Fast Screening Andersen FSA with Glass Expansion Chamber



Glass Twin Impinger


For more information about automation with the Vertus® range, see page 284.

Manual Test System Set-Up

We offer **Glass Expansion Chambers** suitable for the quantification of nasal spray drug product present in the form of particles or droplets that are less than 10 microns.

We offer two sizes for the characterisation of nasal sprays:

- 2 L chamber: to maximise aerosolisation and impactor deposition for regular nasal sprays.
- 5 L chamber: for powerful nasal sprays where increased volume is required to allow a full aerosol plume to generate.



- A Vacuum Pump Page 188
- B Flow Rate Sensor Page 186
- C Adapter and Clamp Page 208
- D Glass Expansion Chamber Page 208
- Nosepiece Adapter
- F Next Generation Impactor NGI Page 84

environmental variability? Our environmental control solutions are designed to help improve the accuracy, sensitivity and reproducibility of test data.

Considering the effects of

Learn more on page 196.

Alternative Impactors/Impingers



Andersen Cascade Impactor ACI with Glass Expansion Chamber



Fast Screening Andersen FSA with Glass Expansion Chamber




Glass Twin Impinger GTI

Qualification

GMP regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- · Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.




APSD Data Analysis Software: Inhalytix®



Inhalytix is a fully validated, end-to-end APSD data management solution for the entry, analysis and reporting of APSD data for all inhaled products.

Automating the complex transformation of raw inhaler testing data into performance-defining metrics, Inhalytix accepts data from both standard and customised impactors and impingers.







Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.

Automation Tools



Improve efficiency



Reduce variability



Eliminate



Increase testing capacity



Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis. See page 298.









Impactor Cleaning System

Standardises cleaning and drying procedures to help ensure the NGI and ACI remain in optimum condition throughout their life. See page 306.







Related Applications

We also offer a range of equipment for additional nasal spray testing application sup



For better *in vitro-in vivo* correlation (IVIVC) testing

See page 222

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320



Design See page 320





Aerodynamic Particle Size Distribution

Nasal Aerosols

Like nasal sprays, nasal aerosols typically produce droplets in the range 20-200 microns, which is outside the effective range of cascade impactors. However, nasal aerosols deliver a proportion (typically <5%) of fine droplets in the <10 micron range. Unlike nasal sprays, nasal aerosols are propellant-driven.

It is important to quantify this FPD since it can penetrate beyond the nasal tract and into the lower respiratory tract or lungs, which may be undesirable. Regulators recommend the use of a cascade impactor in conjunction with a high volume expansion chamber to quantify the amount of drug in the <10 micron range, to assess the potential risk of deposition in the lungs.

Regulations and Guidelines

NGI	Organisation	Chapter/Guidance
	Ph. Eur. / EMA	Guideline on the Pharmaceutical Quality of Inhalation and Nasal Products
	USP / FDA	Nasal Spray and Inhalation Solution, Suspension, and Spray Drug Products – Chemistry, Manufacturing, and Controls Documentation
	ChP	-
	JP	-
0月0		
0月0	Organisation	Chapter/Guidance
	Organisation Ph. Eur. / EMA	Chapter/Guidance -
	_	Chapter/Guidance - Nasal Spray and Inhalation Solution, Suspension, and Spray Drug Products — Chemistry, Manufacturing, and Controls Documentation
	Ph. Eur. / EMA	- Nasal Spray and Inhalation Solution, Suspension, and Spray Drug Products —

APSD of Nasal Aerosols Automated Test System Set-Up

Routine APSD test set-ups for nasal aerosols are easily automated with Vertus® III. Compatible with most nasal aerosols, the Vertus III systems simplify the aerodynamic particle size distribution measurement of nasal aerosols in accordance with Ph. Eur. Chapter 0676 and USP Chapter <601>.

The Vertus III range offer analysts complete control over:

- The speed, angle and duration of shaking, ahead of actuation
- · Firing force and the speed of application and release of that force
- The time delay between the end of shaking and device actuation







Automated shot weight measurement via an integrated balance

- A Vertus III+ with Balance
- B Glass Expansion Chamber
- B Next Generation Impactor NGI

Alternative Vertus® III Interface Plates



Andersen Cascade Impactor ACI with Glass Expansion Chamber



Fast Screening Andersen FSA with Glass Expansion Chamber


For more information about automation with the Vertus® range, see page 284.

Manual Test System Set-Up

We offer Glass Expansion Chambers suitable for the quantification of nasal aerosol drug product present in the form of particles or droplets that are less than 10 microns.

We offer one size ideal for the characterisation of nasal aerosols:

1L chamber: to maximise drug deposition below the top stage of the impactor.



- Vacuum Pump Page 188
- Adapter and Clamp Page 208
- Nosepiece Adapter Page 211
- Flow Sensor Page 184

Considering the effects of environmental variability?

Glass Expansion Chamber Page 208 Our environmental control solutions are designed to help Next Generation Impactor NGI improve the accuracy, sensitivity Page 84 and reproducibility of test data. Learn more on page 196.

Alternative Impactors/Impingers



Andersen Cascade Impactor ACI with Glass **Expansion Chamber**



Fast Screening Andersen FSA with Glass Expansion Chamber

Qualification

GMP regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- · Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.




APSD Data Analysis Software: Inhalytix®



Inhalytix is a fully validated, end-to-end APSD data management solution for the entry, analysis and reporting of APSD data for all inhaled products.

Automating the complex transformation of raw inhaler testing data into performance-defining metrics, Inhalytix accepts data from both standard and customised impactors and impingers.







Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.

Automation Tools





Reduce variability



Fliminate handling errors



Increase testing capacity



Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis. See page 298.









Impactor Cleaning System

Standardises cleaning and drying procedures to help ensure the NGI and ACI remain in optimum condition throughout their life. See page 306.

Recommended for:







Related Applications

We also offer a range of equipment for additional nasal aerosol testing application support:



For better *in vitro-in vivo* correlation (IVIVC) testing

See page 222

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310




Support See page 320



Design See page 320





Aerodynamic Particle Size Distribution


Nasal Powders

Like nasal sprays and aerosols, nasal powders typically produce droplets in the range 20-200 microns, which is outside the effective range of cascade impactors. However, nasal powders deliver a proportion (typically <5%) of fine droplets in the <10 micron range.

It is important to quantify this FPD since it can penetrate beyond the nasal tract and into the lower respiratory tract or lungs, which may be undesirable. Regulators recommend the use of a cascade impactor to quantify the amount of drug in the <10 micron range, to assess the potential risk of deposition in the lungs.


The APSD measurement of nasal powders is typically performed under similar conditions as the APSD measurement of DPIs. However a preseparator is not required.

Regulations and Guidelines



APSD of Nasal Powders

Test System Set-Up



- A Nosepiece Adapter
- B Induction Port
- C Preseparator
- Next Generation Impactor NGI
- Flow Rate Sensor
- F Critical Flow Controller


G Vacuum Pump



Alternative Impactors/Impingers



Related Accessories



Temperature and Relative Humidity Sensor

Ideal for measuring environmental test conditions. See page 183.

Footswitch

Connecting directly to the Critical Flow Controller TPK 100i, the Footswitch enables the precise synchronisation of nasal powder device actuation with the onset of flow. See page 183.

TOP TIP

The angle of the impactor can be adjusted to replicate the angle that the nasal powder device may be used at, to investigate device performance under representative conditions. The APSD measurement is unaffected by gravimetric forces.

APSD of Nasal Powders: Test System Component Parts



Next Generation Impactor NGI

The test set-up is shown with an NGI but an ACI is equally suitable for the assessment of nasal powders. The Fast Screening Andersen (FSA) impactor is a reduced stack plus filter version of the ACI. As little deposition is expected in the lower stages, the FSA may be used to assess the APSD characteristics of nasal powders.

A Preseparator may be required for the collection of large mass, noninhalable powder boluses that may be emitted, prior to entry into the impactor. Different preseparators are available for the NGI, ACI (and FSA).

See page 263 for further information about the FSA.







In addition to the above, the following is needed to complete a fully-operational test set-up for the APSD measurement of nasal powders:

Vacuum Pump

Ideal for the higher, sonic flow rate testing requirements of nasal powders, the High Capacity Pump HCP6 and Super Capacity Pump SCP6 represent the latest in high performance, low maintenance, vacuum pump technology. Our Vacuum Pump range is specifically designed for use in the testing of OINDPs in accordance with pharmacopoeial requirements.

See page 188 for further information about our Vacuum Pump range.













Critical Flow Controller

Simplify nasal powder test system set-up in accordance with pharmacopoeial recommendations with the Critical Flow Controller series. Positioned between the impactor and vacuum pump, the Critical Flow Controller TPK 100i ensures critical (sonic) flow conditions during testing. It measures and records all parameters required for testing and for controlling flow conditions.

See page 172 for further information about our Flow Controller range.







Flow Rate Sensor

Used for establishing accurate and consistent inlet flow rate during testing, the Flow Rate Sensor FRS measures flow rates to the accuracy specified by the pharmacopoeias.

See page 184 for further information about flow rate measurement.











Designed to keep the apparatus organised during testing and improve workflow efficiency, the Inhaler Testing Workstation ITW keeps the cascade impactor and flow meter in position throughout the testing process.

See page 204 for further information.

Recommended for:









Nosepiece Adapter

Special Nosepiece Adapters are available to accommodate the nasal powder device and interface it with the test set-up.

See page 211 for further information.







Oualification

GMP regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- · Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



See page 310 for further information.



APSD Data Analysis Software: Inhalytix®



Inhalytix is a fully validated, end-to-end APSD data management solution for the entry, analysis and reporting of APSD data for all inhaled products.

Automating the complex transformation of raw inhaler testing data into performance-defining metrics, Inhalytix accepts data from both standard and customised impactors and impingers.



Automated data



Meaningful metrics,



Secure and

Unlock key APSD insights in 3 easy steps with Inhalytix. Learn more on page 214.

Automation Tools



Improve efficiency



Reduce variability








Impactor Coater™ IC 200i

Standardises impaction surface coating for both NGI Collection Cups and ACI Collection Plates. See page 296.









Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis. See page 298.









Impactor Genie™ IG 200i

An innovative 2-in-1 solution combining the coating capabilities of the Impactor Coater IC 200i with the drug recovery features of the Gentle Rocker GR 200i. See page 300.

Recommended for:









Sample Preparation Unit SPU 200i

Simplifies and standardises drug recovery from the Induction Ports and Preseparators. See page 302.









Impactor Cleaning System

Standardises cleaning and drying procedures to help ensure the NGI and ACI remain in optimum condition throughout their life. See page 306.

Recommended for:







Related Applications

We also offer a range of equipment for additional nasal powder testing application support:



For better *in vitro-in vivo* correlation (IVIVC) testing

See page 222

Training, Servicing & Support

We offer a comprehensive range of services from bespoke product design to installation, expert training and technical support, optimising all aspects of pharmaceutical testing from start to finish.



TrainingSee page 321



Servicing See page 310



Support See page 320




Design See page 320



Ancillaries

This chapter describes the ancillaries required in addition to the Dosage Unit Sampling Apparatus (DUSA) and cascade impactor to make up a fully-operational test set-up for determining the Delivered Dose Uniformity (DDU) and Aerodynamic Particle Size Distribution (APSD) of orally inhaled and nasal drug products (OINDPs).



Breathing Simulators

Used to apply a more clinically representative breathing profile (relative to a constant flow rate) during testing, our range of Breathing Simulators cover the variety of breathing patterns found in neonatal, infant, child and adult physiologies.

See page 156.

Flow Controllers

Our Breath Actuation Controller BAC is an electricallyoperated, timer controlled, two-way valve specifically designed for testing MDIs, BAIs, MDIs with add-on devices (spacers and valved holding chambers (VHCs), nebulisers, SMIs, nasal sprays and aerosols.

Designed to generate a standardised square-wave breath profile, our Critical Flow Controller TPK is ideal for the routine testing of 'passive' devices such as DPIs, where the drug aerosolisation is dependent on the strength and duration of the patient's inspiration.

See page 172.







Flow Rate Measurement

Flow rate is a critical parameter in the *in vitro* testing of OINDPs. We offer two devices for measuring flow rate with the required range and accuracy to ensure accurate and consistent inlet flow rate during testing. Both units will give similar readings provided they are calibrated and operated correctly.

See page 184.

Vacuum Pumps

Driving most inhaler testing systems is the vacuum pump. We offer a choice of three Vacuum Pumps dependent on the system set-up and the capacity required.







Environmental Control

Our range of environmental control tools are designed to help mitigate the impact of localised environmental conditions on OINDP test data integrity.

See page 194.

NGI Cooler™

Designed to maintain the integrity of the APSD data of aerolised droplets by eliminating evaporation induced by the thermal mass of the impactor, the NGI Cooler provides a temperature-controlled environment for testing.

See page 204.





Inhaler Testing Workstation™ ITW

Providing an 'extra pair of hands', the ITW holds key test equipment in place during testing. Available with attachments to support both DDU and APSD testing, the versatile ITW is the ideal benchtop companion for busy analysts.

See page 196.

Glass Expansion Chambers

Ideal for maximising the aerolisation of nasal drug products in the assessment of fine particles by cascade impaction, Glass Expansion Chambers are available for a wide range of nasal drug product applications.

See page 208.





Mouthpiece & Nosepiece Adapters

Our high quality silicone Mouthpiece and Nosepiece adapters are available for the most common devices on the market. A custom design service is also available for other devices.

See page 211.



Ancillaries

Breathing Simulators

Our range of Breathing Simulators are designed to generate an inhalation and/ or exhalation profile that mimics that of a human subject for more clinically representative testing.

Replacing the fixed flow rate normally used for regulatory testing with a breathing profile has become routine in orally inhaled product (OIP) assessment, with more and more laboratories turning to the use of

breathing simulators to measure the effects of different breathing profiles, flow rates and breathing techniques during product development.

Their use has two major applications:



Pharmacopoeial

To assess the DDU of:

- 1. Nebulisers as per Ph. Eur. 2.9.44 and USP chapter <1601>.
- 2. MDIs when used together with spacers and valved holding chambers, as per USP <1602>.



Improving in vitro-in vivo correlations (IVIVCs)

To apply more clinically representative conditions during *in vitro* testing so as to generate data that are more relevant to *in vivo* behaviour.



The use of breathing simulators is supported by the **Quality by Design (QbD)** strategy outlined in ICH Q8, which relies on scoping the potential impact of any variability that may arise from for example, difference in patient physiology or technique

Choose your Breathing Simulator

From the generation of simple sinusoidal patterns stated in USP and Ph.Eur. for testing of nebulisers and MDIs with a spacer/VHC to complex user-generated profiles for improving in-vitro in-vivo correlations (IVIVCs), our range of versatile Breathing Simulators can be used for a variety of testing applications.






Relevant Applications

Testing MDIs with a spacer/VHC

Improving IVIVCs for MDIs with a

Testing nebulisers







Volume 0 - 155 mL 0 - 900 mL	Volume 0 to 5000 mL
Patient Profile Suitability	Patient Profile Suitability
Neonate/Infant ✓	Neonate/Infant 🗶
Child ✓	Child X
Adult ✓	Adult ✓
21 CFR Part 11 Compliant	21 CFR Part 11 Compliant

Patient Profile S	uitability
Neonate/Infant	/
Child	✓
Adult	✓
21 CFR Part 11 C	ompliant
Frequence 12 - 40 bp	
I:E Ratio)

Frequency	
6 - 60 bpm	
I:E Ratio	
Variable	

Waveforms

Sinusoidal 🗸

Triangular 🗸

Square 🗸

User-defined ✓ (flow vs time)

Patient Profile S	uitability
Neonate/Infant Child	Х
Adult	✓
21 CFR Part 11 C	ompliant
Frequenc	ey
6 - 60 bpr	m
I:E Ratio Variable	

Waveforr	ns
Sinusoidal	/
Square	Х
Triangular	Х
User-defined	Х
Profiles	•

Inhalation 🗸

Exhalation 🗸

Control: Start


On inhalation 🗸

On exhalation 🗸

User-defined X

1:1, 1:2 or 1:3

Profiles	
Inhalation	/
Exhalation	/
Control: St	art
On inhalation	/
On exhalation	1
User-defined	





Breathing Simulator BRS 100i



Ph. Eur. 2.9.44 compliant



USP <1601> and <1602> compliant



ISO 27427:2013 compliant



Touchscreen user interface



MDI Actuation Sensor/ Footswitch remote start capability




Selectable start position (inhalation or exhalation) for spacers/VHCs



Extensive data output options

Key Features:



BRS 100i: User Interface



Setting a test parameter



Set v Actual test parameters (before test run)



Set v Actual test parameters (during test run, with test progress bar and enhanced visualisation of breath status)





Settings menu

BRS 100i connectivity options

BRS 100i: Technical Specifications

Volume	0 to 800 mL (manually adjust)
Frequency	12 - 40 bpm
I:E Ratio	1:1, 1:2 or 1:3
Cycle Number	1 - 9,999 breaths
Waveforms	Sinusoidal
Start	Select start on inhalation or exhalation stroke
User Interface	5 inch, resistive colour touchscreen
Dimensions	460 x 385 x 290 mm (w x d x h)
Connectivity	RS-232 RUN - IN - for MDI Actuation Sensor or Footswitch USB A (for connection with a USB printer) USB B (for connection with a PC)

BRS 100i Accessories



Angle Adapter

Used to angle the device to a position representative of in vivo usage.

Reporting

Extensive data output options are available as standard, including direct reporting to a printer or PC.

Reported parameters

- Start with: Inhalation/Exhalation
- I:E Ratio
- Breath frequency (bpm)
- · Number of breaths
 - Set
 - Actual









BRS Qualification Kit

Oualification & Maintenance

- · Calibration certificate of compliance to Ph. Eur./USP provided as standard
- Comprehensive IQ/OQ documentation packages and toolkits available
- · Qualification Kit available
- Extended Warranty available

Breathing Simulator BRS 100i

Cat. No.	Description
9231	Breathing Simulator Model BRS 100i
1014	BRS 100i Extended Warranty - 1 year
1015	BRS 100i Extended Warranty - 2 years

Accessories

8797	MDI Actuation Sensor
8791	Footswitch
9765	Label Printer
9117	IQ/OQ Documentation for BRS 100i/200i/300i
9115	Qualification Kit for BRSi Series
9118	Re-calibration of Qualification Kit for BRSi Series
9108	BRS 100i Re-calibration Certificate
9104	Angle Adapter for Breath Simulator BRS 100i



Breathing Simulator BRS 200i



Ph. Eur. 2.9.44 compliant



USP <1601> and <1602> compliant



ISO 27427:2013 compliant



21 CFR Part 11 compliant



Stores and recalls methods



Touchscreen user interface




Extensive data output options



Improved accuracy for infant and neonate profile volume requirements

Key Features:





For VHCs, simulate uncoordinated product use by starting the breathing profile on the exhalation portion of the profile.

BRS 200i: User Management

The user management feature of the BRS 200i helps ensure data remains compliant with 21 CFR Part 11. Take control of your data and grant appropriate levels of access to users:

Access Level	Permissions
1	Run approved methods
2	Run methods pending approval, and approved methods
3	Configure methods, run approved and pending methods
4	Approve methods
5	Assign user roles, modify system administration settings
6	Unrestricted access to all functions







Assigning user access level

With password-protected user logins, each test run is date and time stamped and attributable to that user, providing a clear audit trail.

BRS 200i: Method Management


The BRS 200i offers users a number of different ways to define their chosen breathing patterns:








BRS 200i: User Interface



Main run test screen (ready to test)



Main run test screen (test in progress)



Volume/piston selection



Settings menu

BRS 200i: Technical Specifications

Volume	2 cylinders, 2 volumes: 0 - 155 mL, 0 - 900 mL
Frequency	6 - 60 bpm
I:E Ratio	Variable
Waveforms	Sinusoidal, square, triangular, user-defined (flow vs time)
Profiles	Inhalation and/or exhalation
Start	Start on inhalation or exhalation stroke
User Interface	10.1 inch, capacitive colour touchscreen
Connectivity	RS-232 3 x USB A (for import/export of methods and connection with a USB keyboard or mouse) Ethernet - for computer networking Temperature/Humidity Sensor port RUN IN - for MDI Actuation Sensor or Footswitch RUN OUT - to trigger activation of other connected electronic devices

BRS 200i Accessories



EnviroMate™ Environmental Control Chamber

Changes in ambient temperature and/or relative humidity have been identified as variables that could impact nebuliser performance. The BRS 200i easily interfaces with EnviroMate for comprehensive environmental condition control.

See page 196 for further information.

NGI Cooler™ Stand for BRS 200i

Supports interfacing of the NGI Cooler with the BRS 200i, whilst saving precious benchtop space.

See page 202 for further information.





Real-Time Breath Verification Chamber BVC

Enables measurement and recording of the breathing profile generated during the test with the inhaler in place, using the Breath Profile Analyser BPA available in the Qualification Kit for BRSi Series (Cat. No. 9115). For use with the USP Induction Port only.

Reporting

Extensive data output options are available as standard, including direct reporting to a PC and export to a USB memory stick.

3 standard reports are available; Method Report, Run Report and Audit Report.

1) Method Report and 2) Run Report both report the following parameters:

- Waveform
- Volume (mL)
- Frequency (bpm)
- I:E Ratio
- Start Delay (s)
- Inhalation Duration (s)
- Inhalation Delay (s)
- Exhalation Duration (s)
- Exhalation Delay (s)
- Start with: Inhalation/ Exhalation

- Cycles
- Cycle Duration (s)
- Test Duration (s)
- Max. Flow (L/Min)
- Max. Acceleration (L/Min/Min)
- · Cylinder Size (mL)
- Method creation information (e.g. Status, Last Modified By) -Method Report only
- Last Run by (e.g. User, Last Run Date) - Run Report only

3) Audit report

All data changes reported with a date and time stamp attributable per user.





Oualification & Maintenance

- Calibration certificate of compliance to Ph. Eur./USP provided as standard
- Comprehensive IQ/OQ documentation packages and toolkits available
- · Qualification Kit available
- · Extended Warranty available

Breathing Simulator: BRS 200i

Cat. No.	Description
9176	Breathing Simulator Model BRS 200i
1016	BRS 200i/300i Extended Warranty - 1 year
1017	BRS 200i/300i Extended Warranty - 2 years

Accessories

8976	Temperature and Relative Humidity Sensor
8797	MDI Actuation Sensor
8791	Footswitch
9117	IQ/OQ Documentation for BRS 100i/200i/300i
9115	Qualification Kit for BRS 100i/200i/300i
9118	Re-calibration of Qualification Kit for BRSi Series
9119	BVC - Real-Time Breath Profile Verification Chamber



Breathing Simulator BRS 300i



Ph. Eur. 2.9.44 compliant



USP <1601> and <1602> compliant



ISO 27427:2013 compliant



21 CFR Part 11 compliant



Extensive data output options



Touchscreen user interface




Stores and recalls methods



Powerful drive system for generating challenging profiles

Key Features:





For VHCs, simulate uncoordinated product use by starting the breathing profile on the exhalation portion of the profile.

BRS 300i: User Management

The user management feature of the BRS 300i helps ensure data remains compliant with 21 CFR Part 11. Take control of your data and grant appropriate levels of access to users:

Access Level	Permissions
1	Run approved methods
2	Run methods pending approval, and approved methods
3	Configure methods, run approved and pending methods
4	Approve methods
5	Assign user roles, modify system administration settings
6	Unrestricted access to all functions



User login screen



Assigning user access level

With password-protected user logins, each test run is date and time stamped and attributable to that user, providing a clear audit trail.

BRS 300i: Method Management

The BRS 300i offers users a number of different ways to define their chosen breathing patterns:









BRS 300i: User Interface



Main run test screen (ready to test)



Main run test screen (test in progress)



Settings menu

BRS 300i: Technical Specifications

Volume	0 - 5000 mL (500 - 5000 mL certified)
Frequency	6 - 60 bpm
I:E Ratio	Variable
Waveforms	Sinusoidal, square, triangular, user-defined (flow vs time)
Profiles	Inhalation and/or exhalation
Start	Start on inhalation or exhalation stroke
User Interface	10.1 inch, capacitive colour touchscreen
Connectivity	RS-232 3 x USB A (for import/export of methods and connection with a USB keyboard or mouse) Ethernet - for computer networking Temperature/Humidity Sensor port RUN IN - for MDI Actuation Sensor or Footswitch RUN OUT - to trigger activation of other connected electronic devices

BRS 300i Accessories

Real-Time Breath Verification Chamber BVC

Enables measurement and recording of the breathing profile generated during the test with the inhaler in place, using the Breath Profile Analyser BPA available in the Qualification Kit for BRSi Series (Cat. No. 9115). For use with the USP Induction Port only.



Reporting

Extensive data output options are available as standard, including direct reporting to a PC and export to a USB memory stick.

3 standard reports are available; Method Report, Run Report and Audit Report.

1) Method Report and 2) Run Report both report the following parameters:

- Waveform
- · Volume (mL)
- Frequency (bpm)
- I:E Ratio
- Start Delay (s)
- Inhalation Duration (s)
- · Inhalation Delay (s)
- Exhalation Duration (s)
- Exhalation Delay (s)
- Start with: Inhalation/ Exhalation

- · Max. Flow (L/Min)
- Max. Acceleration (L/Min/Min)
- · Cylinder Size (mL)
- Method creation information Method Report only
- · Last Run by (e.g. User, Last Run Date) - Run Report only

Cycles

- Cycle Duration (s)
- Test Duration (s)
- (e.g. Status, Last Modified By) -

3) Audit report

All data changes reported with a date and time stamp attributable per user.





Oualification & Maintenance

- · Calibration certificate provided as standard
- · Comprehensive IQ/OQ documentation packages and toolkits available
- · Oualification Kit available
- · Extended Warranty available

Breathing Simulator BRS 300i

Cat. No.	Description
9186	Breathing Simulator Model BRS 300i
1016	BRS 200i/300i Extended Warranty - 1 year
1017	BRS 200i/300i Extended Warranty - 2 years

Accessories

8976	Temperature and Relative Humidity Sensor
8797	MDI Actuation Sensor
8791	Footswitch
9117	IQ/OQ Documentation for BRS 100i/200i/300i
9115	Qualification Kit for BRSi Series
9118	Re-calibration of Qualification Kit for BRSi Series
9119	BVC - Real-Time Breath Profile Verification Chamber



Ancillaries

Flow Controllers


Flow rate and volume of air control are crucial when it comes to the DDU testing and APSD measurement of OINDPs. The use of an appropriate flow controller is vital to comply with the regulatory requirements and streamline the testing, and when creating specific methods which are easy to follow and transfer as required.

The Ph. Eur. and USP require that test flow rate is controlled to within +/-5% of the specified value. This requirement can be met by selecting an appropriate flow control ancillary.

MDIs, MDIs with a Spacer/VHC, BAIs, Nebulisers, SMIs, Nasal Sprays & Nasal Aerosols

Regulatory requirements for these OINDPs call for the control of:

- Air flow rate to a defined constant flow rate or to apply defined breathing profiles. See page 156.
- Total air volume.
- Delay/synchronisation to begin sampling at a defined time.



DPIs

In the case of DPIs, flow control is particularly important. Since most DPIs are classified as "passive" devices (i.e. they rely solely on the patient's inspiration to operate), variations in flow rate can significantly affect device performance. It is therefore a regulatory requirement that critical flow conditions are applied during testing.

The testing of DPIs is further complicated by the fact that devices vary in terms of their resistance to flow i.e. some require more effort to inhale through than others.

Setting the flow rate for the testing of DPIs is more complex than for other types of OINDP. There are three variables which need to be established to determine the breath profile for DPI testing:

Flow Rate (Q) Inspiration Volume Critical Flow Control


1. Flow Rate (Q)

The *in vivo* strength and duration of the user's inspiration is broadly replicated by the flow rate used and the duration of testing.

To establish the correct flow rate, the flow must first be adjusted to produce a pressure drop comparable with that found at the mouth of the user *in vivo* when using the particular inhaler being studied.

Both the Ph.Eur. and USP suggest a pressure drop over the inhaler of 4 kPa as broadly representative of the pressure drop generated during inhalation by patients using DPIs.

The pressure drop created by drawing air through an inhaler can be determined by measuring the absolute pressure downstream of the inhaler mouthpiece and comparing this directly with atmospheric pressure. Using a flow control valve, it is then a simple matter to adjust the flow rate from the vacuum pump to produce the required pressure drop of 4 kPa and then, by replacing the inhaler with a suitable flow meter, to measure the flow rate, Q, required to produce this pressure drop.



It is this Flow Rate Q, that the pharmacopoeias state should be used for DDU testing and APSD measurement.

The only exception to this criterion is that if the flow required to produce a 4 kPa pressure drop is >100 L/min, as for example in the case of particularly low resistance inhalers, then 100 L/min should be used.

2. Inspiration Volume

>

Once the flow rate (Q) has been established, it is now necessary to control the volume of air drawn through the inhaler during testing to the 2 or 4 litres per simulated inhalation required by the pharmacopoeias/regulators.

This is to simulate the *in vivo* inspiration volume of the patient and is achieved by introducing a timer-controlled, fast-acting solenoid valve between the test device and the vacuum pump.

TOP TIP By using a timer to

By using a timer to control the time that the solenoid valve is open, it is possible to control the volume of air drawn through the inhaler to achieve the volume specified.



4 litres is considered to be the normal forced inhalation capacity of an average sized male weighing approx. 70 kg. In practice, it is not uncommon to widen the scope of the test parameters to cover a broader target patient population, such as geriatrics and paediatrics, as well as those already suffering from pulmonary problems, including typical use and unintentional misuse conditions.

Example Calculation

Volume: 4 litres (Ph. Eur) Flow Rate (Q): 100 L/min


Time = Volume * 60/Flow rate

= 2.4 seconds

Volume: 4 litres (Ph. Eur) Flow Rate (Q): 60 L/min

Time = Volume * 60/Flow rate

= 4 seconds



3. Critical Flow Control



Once the parameters to control the strength and duration of the simulated breathing cycle have been established, there is one final issue to be considered – flow rate stability.

Ensuring stable flow throughout the test is critical to the testing of DPIs, since, as passive devices, they can be sensitive to small changes in flow rate.

An easy way to validate flow rate stability is to ensure that critical (sonic) flow occurs in the flow control valve. This can be confirmed by simply measuring the absolute pressure at a point on either side of the valve.

Providing that the pressure downstream of the valve is less than half of the upstream pressure i.e. that the ratio $P3/P2 \le 0.5$ then critical (sonic) flow is assured and the flow rate can be assumed to be stable.

Schematic of APSD Measurement System for DPIs COPLEY Vacuum Pump Critical Flow Controller Controller

Conforming to the Ph. Eur. and USP specifications for a system that controls the key variables impacting the test conditions for DPIs (as described in the previous section), our Flow Controllers have become the industry-standard for both DDU and APSD applications.

Choose your Flow Controller





Device Type	BAC 100i/-R	TPK 100i/-R
MDI	Y	Υ
MDI with Spacer/VHC	Y	Υ
Breath-Actuated MDI	Υ	Υ
DPI	N	Υ
Nebuliser	Υ	Υ
SMI	Υ	Υ
Nasal Spray	Υ	Υ
Nasal Aerosol	Υ	Υ
Nasal Powder	N	Υ



Breath Actuation Controller **BAC 100i**



Ph. Eur. and USP compliant



Simplified workflow with user-guided test set-up



Integrated timer for control of solenoid valve



Fully automated In situ impactor leak testing



Extensive data output options



Intuitive touchscreen control



Spacer/VHC testing delay function




Inlet/In-line flow meter modes



Atmospheric pressure measurement

Key Features:





BAC 100i-R and BAC 100i

TOP TIP BAC 100i v BAC 100i-R Two versions of the unit are avails The BAC 100i-R (Reversed) is fund

The BAC 100i-R (Reversed) is functionally identical to the BAC 100i but the position of the pneumatic connections are reversed to improve connectivity with other inhaler testing equipment.


BAC 100i: User Interface



Guided test set up process



Target v Actual test parameters (before test run)



O ACTUAL

Settings menu



Leak test screen



Flow method screen



Easy setting of delay time

BAC 100i: Technical Specifications

User Interface	Resistive touchscreen
Flow Setting	Manual
Temperature/Relative Humidity Measurement Capabilities	Yes (see page 179)
Auto-Trigger	MDI Actuation Sensor Footswitch
Critical Flow Control	No
Solenoid Valve Opening/Closing Time	25/25 ms
Timer Range	0-600.0s resolution 0.1s
Dimensions	415 x 315 x 250 mm (w x d x h)

BAC 100i: Accessories

MDI Actuation Sensor

Enabling precise synchronisation of the MDI actuation with the onset of flow, the MDI Actuation Sensor simply clips on to most commercially available MDI canisters and connects directly to the BAC 100i.

Alternatively, a Footswitch can be used to synchronise the actuation of MDIs, nebulisers, SMIs and nasal aerosols with the onset of flow.

The MDI Actuation Sensor can also be used for the testing of MDIs with a spacer/ VHC in accordance with the specifications laid down in USP Chapter <1602>.





Temperature and Relative Humidity Sensor

The Temperature and Relative Humidity Sensor is designed to provide analysts with accurate data about environmental conditions.

Label Printer

The Label Printer offers space-efficient printing for laboratories, with easy setup, intuitive operation, and high-quality report output. Integrating seamlessly with compatible Copley products, the Label Printer prints critical test parameters on to adhesive labels; ideal for affixing directly into laboratory notebooks.



Reporting

Extensive data output options are available as standard, including direct reporting to a printer or PC.

Available reports:

- Run test
- Test setup
- Leak test
- Calibration



Oualification & Maintenance

- · Certificate of compliance to Ph. Eur./USP provided as standard.
- · Comprehensive IQ/OQ documentation packages and toolkits available.
- · Extended warranty available

Breath Actuation Controller BAC 100i

Cat. No.	Description
8975	Breath Actuation Controller Model BAC 100i
8975-R	Breath Actuation Controller Model BAC 100i-R (Inlet
	Outlet Reversed)
1020	BAC 100i/R Extended Warranty - 1 year
1021	BAC 100i/R Extended Warranty - 2 years

Accessories

8976	Temperature and Relative Humidity Sensor
8797	MDI Actuation Sensor
8791	Footswitch
9765	Label Printer
8983	BAC 100i Re-calibration Certificate
8752	Flow Time Verification Kit
8753	Re-calibration of Flow Time Verification Kit



Critical Flow Controller **TPK 100i**



Ph. Eur. and USP compliant



Simplified workflow with user-guided test set-up



User warned if sonic flow conditions are not met



Fully automated In situ impactor leak testing



Extensive data output options



'Fly-by-wire' flow control valve - operation can be automated for more efficient and reproducible data



In-line flow measurement accommodated




Intuitive touchscreen control

Key Features:



The TPK can also be used as a Breath Actuation Controller (BAC) for testing MDIs with a spacer/VHC and BAIs in accordance with Ph.Eur. 0671 and USP Chapter <1602>.





TPK 100i-R and TPK 100i

TPK 100i v TPK 100i-R

TOP TIP

Two versions of the unit are available. The TPK 100i-R (Reversed) is functionally identical to the TPK 100i but the position of the pneumatic connections are reversed to improve connectivity between the TPK and other inhaler testing equipment.

TPK 100i: User Interface



Guided test set up process



Test set-up report



Target v Actual test parameters (before test run)



Leak test screen




Device resistance measurement



Settings menu



Guided calibration process



Flow method screen



Test settings

TPK 100i: Technical Specifications

User Interface	Resistive touchscreen
Flow Setting	Manual and Automated
Temperature/Relative Humidity Measurement Capabilities	Yes (see page 183)
Auto-Trigger	Footswitch MDI Actuation Sensor
Critical Flow Control	Yes
Solenoid Valve Opening/Closing Time	25 ms / 25 ms
Timer Range	0-600.0s resolution 0.1s
Dimensions	415 x 315 x 250 mm (w x d x h)

TPK 100i: Accessories

Temperature and Relative Humidity Sensor

The Temperature and Relative Humidity Sensor is designed to provide analysts with accurate data about environmental conditions.





Footswitch

Enabling precise synchronisation of device actuation with the onset of flow, the Footswitch connects directly to the TPK 100i.

Alternatively, an MDI Actuation Sensor can be used for synchronisation of MDI actuation and the onset of flow.

Label Printer

The Label Printer offers space-efficient printing for laboratories, with easy setup, intuitive operation, and high-quality report output. Integrating seamlessly with compatible Copley products, the Label Printer prints critical test parameters on to adhesive labels; ideal for affixing directly into laboratory notebooks.



Reporting

Extensive data output options are available as standard, including direct reporting to a printer or PC.

Available reports:

- · Run test
- Test setup
- Leak test
- Flow resistance
- Calibration



Qualification & Maintenance

- Certificate of compliance to Ph. Eur./USP provided as standard.
- Comprehensive IQ/OQ documentation packages and toolkits available.
- · Extended warranty available

Critical Flow Controller TPK 100i

Cat. No.	Description
8970	Critical Flow Controller Model TPK 100i
8970-R	Critical Flow Controller Model TPK 100i-R
	(Inlet/Outlet Reversed)
1018	TPK 100i/R Extended Warranty - 1 year
1019	TPK 100i/R Extended Warranty - 2 years

Accessories

8976

8791	Footswitch
8797	MDI Actuation Sensor
9765	Label Printer
8973	TPK 100i Re-calibration Certificate
8752	Flow Time Verification Kit
8753	Re-calibration of Flow Time Verification Kit

Temperature and Relative Humidity Sensor



Flow Rate Measurement

Air flow control is critical in the DDU and APSD testing of OINDPs. For many inhaled products, air flow triggers or drives aerosolisation of the formulation and it can therefore have a significant effect on both delivered dose and APSD. Equally importantly, air flow impacts the performance of the test apparatus, notably cascade impactors which are designed to function at a constant air flow rate.

In addition, for some devices, especially DPIs, the air flow through the device provides the motive force for dose delivery; indeed, some breath-actuated/operated devices trigger only when the flow rate through them exceeds a certain value.

DDU Testing

A constant, repeatable flow rate is required throughout testing to ensure conformance with the regulatory requirements and pharmacopoeial specifications.

APSD Measurement


Air flow rate has a direct influence on the aerodynamic performance of cascade impactors. The jet-to-plate distances on most commonly used impactors are fixed. Therefore, as long as the nozzle diameters remain within defined tolerances and there are no leaks in the system, the cutoff diameter of any given stage is directly related to the volumetric flow rate of air passing through it. A change in flow rate results in a change in the aerodynamic particle size characteristics of the stage or stages concerned altering the measured APSD.

Determining Test Flow Rate


Although patient inspiration subjects inhalers to varying flow rates, DDU testing and APSD measurement require a constant volumetric air flow. Within this constraint, flow rates are specified, as far as possible, to reflect the conditions of use. Because of the link between air flow rate and cascade impactor performance, flow meters for OINDP testing must:

- 1. Be capable of measuring volumetric flow (L/min)
- 2. Be calibrated for exit flow as opposed to inlet flow


We offer two flow meters that meet these criteria.



Choose your Flow Rate Measurement Device



Flow Rate Sensor FRS



FRS in Holder of Inhaler Testing Workstation™ ITW





Volumetric Mode:

Calculation of flow rate based on live temperature/pressure conditions



Standard Mode:

Calculation of flow rate based on pre-set temperature/pressure conditions



Fast data output, ideal for understanding changes in flow rate, such as rise-times



Low flow resistance

Technical Specifications

Operation principle	Thermal (MEMS)
Flow Rate Range	- 200 to + 200 Std L/min
Resolution	0.1 StdL/min
Accuracy	Typically +/-1.75% of reading Maximum +/-2.5% of reading or ± 0.2 Std L/min, whichever is greater
Flow Resistance	< 4 kPa at 200 Std L/min
Volumetric Flow Calculation	Accurate calculation from in-built temperature and pressure sensors
Inlet Filter	Required (one supplied, replacements available)
Connectivity	Interface to external devices, such as: - Breath Actuation Controller BAC 100i/-R - Critical Flow Controller TPK 100i/-R - PC
Response Time	1 ms
Reporting	Flow rate and calibration data via RS-232 and USB
Calibration	Factory calibrations only
Power	5V DC, mains power supply provided with the FRS

i

The FRS is compatible with the following versions of Breath Actuation Controller and Critical Flow Controller: BAC 100i/BAC 100i-R, firmware v1.2.0 and above. TPK 100i/TPK 100i-R, firmware v1.2.0 and above.

The FRS connects directly to a Flow Controller via an RS-232 connection. The inlet flow rate is displayed clearly on the flow controller screen.



Critical Flow Controller TPK 100i

Qualification & Maintenance

 Calibration certificate of compliance to Ph. Eur./USP provided as standard

Flow Rate Sensor FRS

Cat. No.	Description
8100	Flow Rate Sensor Model FRS
8105	Inline Adapter Kit for FRS
8139	ITW Holder for FRS
8106	Pack of 12 Inlet Filters for FRS
8110	Re-calibration Certificate for FRS
5239	FRS Flow Meter Adapter for Induction
	Port, DUSA, WSC2, Filter Holder and Child
	Alberta Idealised Throat
8517	FRS Flow Meter Adapter for Adult Alberta
	Idealised Throat
8547	FRS to Alberta Idealised Nasal Inlet
8920	FRS Flow Meter Adapter for Glass Twin
	Impinger and FP Induction Port

Flow Meter DFM 2000

Key Features:



Technical Specifications

Operation Principle	Hot Wire Mass Flow
Flow Rate Range	0 - 200 Std L/min
Resolution	0.1 L/min between 90 and 200 L/min
Accuracy	+/-2% of reading
Flow Resistance	~ 8 kPa at 200 Std L/min
Volumetric Flow Calculation	Accurate calculation from in-built temperature and pressure sensors
Inlet Filter	Inlet filter required
Connectivity	Interface to external devices, such as - Breath Actuation Controller BAC 100i/-R - Critical Flow Controller TPK 100i/-R
Response Time	< 4ms. 63% of final value at full scale flow
Reporting	Flow rate & calibrate date via RS-232
Calibration	Factory calibrations only

Qualification & Maintenance

 Calibration certificate of compliance to Ph. Eur./USP provided as standard

Flow Meter DFM 2000

Cat. No. Description

8764 Flow Meter Model DFM 2000

Accessories

5238 DFM Flow Meter Adapter

8765 Re-calibration Certificate for DFM 2000



Vacuum Pumps

We offer vacuum pumps specifically designed for use in the testing of MDIs, DPIs, nebulisers and nasal products in accordance with the specifications laid down in the Ph. Eur. and USP.



Choose your Vacuum Pump

					9
	Application	Low Capacity Pump LCP6	High Capacity Pump HCP6	2 x HCP6	Super Capacity Pump SCP6
	MDI	Υ	Υ	Υ	Y
	MDI with Spacer/VHC	Υ	Υ	Υ	Y
	DPI sonic flow with NGI @ > 80 L/Min	N	N	Υ	Y
	DPI sonic flow with NGI @ < 80 L/Min	N	Υ	Υ	Y
	Nebuliser	Y	Υ	Υ	Y
	SMI	Υ	Υ	Υ	Y
	Nasal Spray	Υ	Y	Υ	Y
	Nasal Aerosol	Υ	Υ	Υ	Y
A	Nasal Powder	N	Υ	Υ	Y

Low Capacity Pump LCP6

Key Features:





Low maintenance



Advanced sound insulation



Oil-free



Self-sealing compound carbon vanes continually adjust so that the pump effectively performs with "as new" efficiency throughout its service life.

Technical Specifications

Туре	Rotary Vane
Lubrication Type	Dry
Max. Flow in L/min (unrestricted)	133
Max. Sonic Flow through NGI	N/A
Max. Vacuum Level	<15 kPa
Applications: Nasal	Yes
Nebulisers	Yes
MDIs	Yes
DPIs	No
Routine Maintenance	None
Dimensions (w x d x h)	270 x 335 x 280 mm
Weight (kg)	18.4 kg

Oualification & Maintenance

- Included in IQ/OQ Documentation for Inhaler Testing Systems - see page 317
- · Extended Warranty available


Low Capacity Pump LCP6

Cat. No.	Description
7923	Low Capacity Pump Model LCP6
1022	LCP6 Pump Extended Warranty - 1 year
1023	LCP6 Pump Extended Warranty - 2 years
Λ	

Accessories

7904 Overhaul Kit for LCP6

High Capacity Pump HCP6



Technical Specifications

	1 x HCP6	2 x HCP6
Туре	Rotary Vane	Rotary Vane
Lubrication Type	Dry	Dry
Max. Flow in L/min (unrestricted)	416	833
Max. Sonic Flow through NGI (L/min)	80	100
Max. Vacuum Level	<15 kPa	<15 kPa
Applications: Nasal	Yes	Yes
Nebulisers	Yes	No
MDIs	Yes	No
DPIs	Yes	Yes
Routine Maintenance	None	None
Dimensions (w x d x h)	322 x 580 x 390 mm	750 x 580 x 390 mm
Weight (kg)	45	90

Oualification & Maintenance

- Included in IQ/OQ Documentation for Inhaler Testing Systems - see page 317
- · Extended Warranty available

High Capacity Pump HCP6


Cat. No.	Description
7921	High Capacity Pump Model HCP6
1024	HCP6 Pump Extended Warranty - 1 year
1025	HCP6 Pump Extended Warranty - 2 years

Accessories

7905 Overhaul Kit for HCP6

Super Capacity Pump SCP6

Key Features:



Technical Specifications

Туре	Rotary Vane
Lubrication Type	Oil
Max. Flow in L/min (unrestricted)	683
Max. Sonic Flow through NGI	100
Max. Vacuum Level	<0.1 kPa
Applications: Nasal	Yes
Nebulisers	Yes
MDIs	Yes
DPIs	Yes
Routine Maintenance	Oil/Filter Change
Dimensions (w x d x h)	423 x 653 x 455 mm
Weight (kg)	71

Oualification & Maintenance

- Included in IQ/OQ Documentation for Inhaler Testing Systems - see page 317
- Extended Warranty available


Super Capacity Pump SCP6

Cat. No.	Description
7928	Super Capacity Pump Model SCP6
1026	SCP6 Pump Extended Warranty - 1 year
1027	SCP6 Pump Extended Warranty - 2 years
	, , ,

Accessories


7909 Maintenance Kit for SCP67913 Replacement Lubricant (5 litres) and Funnel for SCP5





Environmental Control

Many factors have been identified that could give rise to variability in the testing of orally inhaled and nasal drug products (OINDPs). Each of these factors reduces the discriminatory power of tests to accurately determine product variability. Effective strategies to mitigate sources of variability in delivered dose uniformity (DDU) testing and aerodynamic particle size distribution (APSD) measurement have been the focus of rigorous investigation for many years. Environmental conditions is one key area.



Localised changes in laboratory temperature and humidity, and also the presence of electrostatic charge, are known to have a direct influence on the dose emission and aerosol generation performance of OINDPs, thereby compromising DDU and APSD test data integrity.



TEMPERATURE

- Particle/droplet evaporation effects for MDIs, nebulisers, ADIs, nasal sprays and aerosols
- Accuracy of volumetric air flow rate measurements during testing



HUMIDITY

- Particle/droplet evaporation effects for MDIs, nebulisers, ADIs, nasal sprays and aerosols
- Water absorption effects for hygroscopic powder-based DPI formulations
- Electrostatic charge-related issues exacerbated by low humidity



- Device generated triboelectrification of particles (via drug, propellant, metering valve system, inhaler materials, packaging etc.)
- · Analyst-induced electrostatic effects (via clothing, handling)

Use of proper control mechanisms

Ph. Eur. and USP specifically reference the control of environmental conditions in cases where temperature and/or humidity limits are stated on the product label and/or it is specified in the relevant monograph.

However, it is also good practice to implement environmental controls across all DDU and APSD testing applications to reduce variability and improve the accuracy, sensitivity and reproducibility of data.

COPLEY

EnviroMate™

Inadequate control of environmental conditions can affect the dose emission and aerosol generation performance of OINDPs, leading to erroneous data and costly testing delays. EnviroMate is a cost-effective, compact, benchtop solution that addresses these issues with considerable value for scientists faced with:

- · Variable laboratory conditions or inadequate climate control
- · OINDPs with high sensitivity to temperature, humidity and/or electrostatic charge
- · Poor reproducibility and unexplained out-of-specification (OOS) results
- Achieving better environmental control, in a cost-effective manner, without investing in a dedicated climatically-controlled laboratory for testing

Accommodating all types of dose uniformity sampling apparatus and cascade impactor, EnviroMate establishes and maintains uniform temperature and humidity throughout the chamber, whilst the built-in anti-static system helps minimise the unwanted effects of electrostatic charge.

Ideal for those struggling to achieve stable conditions for delivered dose uniformity (DDU) and aerodynamic particle size distribution (APSD) testing, EnviroMate provides users with consistent environmental control, in the immediate test area, enhancing data accuracy and repeatability.



Designed specifically for OINDP testing



Maintains uniform temperature and humidity throughout





Minimises electrostatic charge



As recommended in Ph. Eur. and USP



Compact, benchtop solution




No routine maintenance required



Energy efficient solution compared to dedicated environmental control rooms

Key Features:




EnviroMate™ Performance Data

EnviroMate has been specifically designed to accommodate a wide range of inhaler testing apparatus, including:

- Next Generation Impactor (NGI), Anderson Cascade Impactor (ACI), Multi-Stage Liquid Impinger (MSLI), Glass Twin Impinger (GTI)
- Inhaler Testing Workstation™ ITW with DUSA for MDIs and DPIs, SMIs and Nasal, Waste Shot Collector WSC2, Andersen Cascade Impactor (ACI) and Multi-Stage Liquid Impinger (MSLI)

EnviroMate can also interface with the BRS 200i and NGI Cooler[™] for nebuliser testing, and flow controllers and vacuum pumps for testing of other OINDPs.

The graph opposite shows the performance envelope of the EnviroMate, with the shaded area defining the optimal performance range.



EnviroMate™ Technical Specifications

User Interface	Digital display with set-point control buttons
Ambient Temperature	16 - 28°C
Temperature Control Range	17 - 35°C
Temperature Control Accuracy	Typically ±2°C
Ambient Humidity	35 - 85% RH
Humidity Control Range	15 - 85% RH
Humidity Control Accuracy	Typically ±5% RH
Electrostatic Minimisation System	Included as standard
Ancillary Connector Ports	2 x ports for quick-release connectors (left and right) 2 x ports for cables (left and right) 2 x ports for interface with BRS 200i and NGI Cooler™ (right)
Sound Level	63 dBA at 1m
Power Supply	Mains supply: 230V, 50Hz or 115V, 60Hz
Compressed Air Supply Connector	SMC Series KQ2 6 mm O.D. One-touch Fitting
	5 to 8 bar (G)
Compressed Air Supply Pressure	3 to 6 but (d)
Compressed Air Supply Pressure Capacity	280 litres



EnviroMate with Next Generation Impactor (NGI)



EnviroMate with Inhaler Testing Workstation™ ITW and DUSA for DPIs



EnviroMate with Andersen Cascade Impactor (ACI)



EnviroMate with Breathing Simulator BRS 200i for nebuliser testing




EnviroMate with NGI Cooler for nebuliser testing

EnviroMate™

Cat. No.	Description
5040	EnviroMate Environmental Chamber
5042	IQ/OQ Documentation for EnviroMate
5043	Qualification Tools for EnviroMate

5044 Recalibration of EnviroMate Qualification Tools

We also offer additional tools to help mitigate the impact of localised environmental conditions on OINDP test data integrity:



Anti-Static Grounding Kit

Dissipating electrostatic charge build-up introduced during handling of the test apparatus, inhaler and other non-conductive items coming into contact with the laboratory bench during test preparation, the Anti-Static Grounding Kit safely grounds the analyst for effective static elimination. The Anti-Static Grounding

Kit contains:

- · A comfortable and adjustable user wristband
- 1x bench mat
- 1 x earth plug for grounding (UK, EU and US versions available)

Cat. No. Description

9300 Antistatic Grounding Kit

Electrostatic Eliminator

Designed to effectively eliminate electrostatic charge over the lab bench area, the Electrostatic Eliminator is an efficient ioniser that is capable of neutralising static across a broad area, while still providing a comfortable working environment for analysts.

- Excellent coverage over a wide area (2 m x 0.6 m)
- · Rapid electrostatic discharge
- · Compact, benchtop unit

Cat. No. Description

9301 Electrostatic Eliminator





Digital Static Meter

Ideal for measuring the intensity and polarity of electrostatic charge in the test area, the Digital Static Meter is an easy-to-use, handheld device ideal for quick checks of the electrostatic charge level around the equipment prior to testing, to facilitate optimal control.

- · Compact and lightweight unit
- Measures electrostatic charge in the range 0 to ± 20 kV

Cat. No. Description

9302 Digital Static Meter





NGI Cooler[™]

Exacerbated evaporation caused by the thermal mass of the NGI may be an issue for devices such as nebulisers that deliver the drug as an aerolised solution. Loss of solvent reduces droplet size, producing artificially low particle size measurements and compromises the integrity of APSD data.

The NGI Cooler is designed to support testing in a temperature-controlled environment, cooling the impactor to 5°C to overcome the issue of droplet size change due to evaporation.




Ph. Eur. and USP compliant



Quiet operation



Precise temperature control



NGI Cooler Accessories

NGI Cooler Stand for BRS 200i

Supports interfacing of the NGI Cooler with the BRS 200i, whilst saving precious benchtop space.



NGI Cooler: Technical Specifications

Pharmacopoeial Compliance	Ph. Eur. 2.9.44 USP <1601> EPAG recommended
Temperature Range	0°C and ambient (typically 5°C to 10°C)
Temperature Accuracy	± 1.5°C
Dimensions (w x d x h)	1000 x 500 x 575 mm

Qualification & Maintenance

- Comprehensive IQ/OQ documentation packages and toolkits available
- · Extended Warranty available

NGI Cooler™

Cat. No.	Description
5009	NGI Cooler
1046	NGI Cooler Extended Warranty - 1 year
1047	NGI Cooler Extended Warranty - 2 years

Accessories

9114	NGI Cooler Stand for BRS 200i
5011	NGI Cooler Qualification Documentation
5012	NGI Cooler Qualification Tools
5013	Re-calibration of NGI Cooler
	Qualification Tools



Inhaler Testing Workstation™ ITW

The hub of an inhaler testing system, the ITW is a modular workstation designed to aid handling and manipulation of the various pieces of test apparatus and accessories, improving workflow.

The ITW offers analysts the flexibility to pick and choose the attachments necessary for their test set-up needs. Simply connect the required attachments and start testing with greater ease.



Suitable for DDU testing and APSD measurement applications



Quick-slide attachments for rapid method change



Suitable for both right- and left-handed configurations



Flexible configurations to suit different testing requirements



Stable and secure platform for test components



Supplied with quickrelease connectors for easy interfacing

ITW: DDU Testing

The ITW keeps the DUSA collection tube, vacuum connection, flow rate sensor and waste shot collector (WSC2) in place during the testing process.





Holder for DUSA Filter Holder



Tubing attachments ensure the workstation remains organised



Dose sampling and waste collection for nasal drug products

ITW: APSD Measurement

The ITW provides a stable support for the impactor during testing, together with the flow rate measurement device.



Also compatible with:



Multi-Stage Liquid Impinger (MSLI)

Inhaler Testing Workstation™ ITW

Cat. No.	Description
8120	Inhaler Testing Workstation - Baseplate and Upright
8125	Inhaler Testing Workstation for WSC2 with Switching Valve
8136	ITW Holder for ACI Base
8139	ITW Holder for FRS
8132	ITW Holder for DPI DUSA
8131	ITW Holder for MDI DUSA
8133	ITW Holder for MDI/DPI Filter Support Cap
8137	ITW Holder for USP Induction Port
8130	ITW QR Tube Holder

Spare/Additional Tubing




A variety of tubing is available to provide connections between the various components making up the inhaler testing system. The 3 mm tubing is designed to provide the connection between the DUSA for DPIs and Critical Flow Controller.

Tubing

Description
10 mm i.d. PVC Tubing (per metre)
16 mm i.d. Wire Reinforced PVC Tubing (per metre)
3 mm i.d. PVC Tubing (per metre)

Quick-Release Connectors



Quick-Release Connectors are provided as standard with various pieces of equipment. Additional connectors can be purchased if required in two sizes, 13 mm and 16 mm designed for use with 10 mm i.d. and 16 mm i.d. tubing respectively.

Quick-Release Connectors

Cat. No.	Description
5026	13mm Quick-Release Connector - 3/8" threaded QR Male
5027	13mm Quick-Release Connector - 1/2" threaded QR Male
5028	16mm Quick-Release Connector - 3/8" threaded QR Male
5029	16mm Quick-Release Connector - 1/2" threaded QR Male






Glass Expansion Chambers

The majority of nasal products are designed to generate droplets/particles with a mass median aerodynamic diameter (MMAD) of greater than 10 to 20 microns. This is to increase nasal deposition and minimise deposition in the lungs.

However, most sprays deliver a proportion (typically <5%) of fine droplets in the <10 micron range. It is important to quantify this dose since it can penetrate beyond the nasal tract and into the lower respiratory tract or lungs, which may be undesirable.

Cascade impactors are designed to capture particles in the range 0 to 10 microns and are widely used for this application.



The use of a cascade impactor in conjunction with a high volume expansion chamber is used to measure the amount of drug in small particles or droplets in respect of nasal sprays and aerosols.

In accordance with the draft guidance, we offer a range of glass expansion chambers to meet these requirements.



FDA compliant



3 chamber sizes available



Certified volume



Special nosepiece adapters are available for the entry port to accommodate the different types of nasal devices



We offers three sizes:



1 L chamber:

to maximise drug deposition below the top stage of the impactor (i.e. for nasal aerosols)



2 L chamber:

to maximise aerosolisation and impactor deposition (i.e. for nasal sprays)



5 L chamber:

for powerful nasal sprays where increased space is required to generate full plume

Glass Expansion Chamber Accessories

Benchtop Holder for Glass Expansion Chamber

For keeping benchtops tidy and glass expansion chambers safe.





Expansion Chamber to Flow Meter Adapter

For ensuring a proper interface between the Glass Expansion Chamber and flow meter when setting flow rate.

Glass Expansion Chambers

Description
1000 mL Glass Expansion Chamber
2000 mL Glass Expansion Chamber
5000 mL Glass Expansion Chamber
Volume Verification Certificate for Expansion Chamber
Adapter & Clamp for ACI/FSA*
Adapter & Clamp for NGI/FSI*
Set of 10 O-Rings for Expansion Chamber Adapter
'Quick Clamp' for ACI
Benchtop Holder for Glass Expansion Chamber

^{*} Please specify Aluminium (A), 316 Stainless Steel (S) or Titanium (T) when placing your order.





Mouthpiece & Nosepiece Adapters

Ensure a proper seal is maintained between the device under test and the sampling apparatus with our range of Mouthpiece and Nosepiece Adapters.

Specially moulded from high quality silicone rubber to ensure superior performance, adapters are available for the more common devices on the market, or can be custom-made for your specific device type.

The adapters are generally transferable between different product test systems, however, there are cases where the inlet diameters may differ between apparatus. Please specify the intended testing system when ordering to ensure the correct size adapter is supplied.

Mouthpiece Adapters


Suffix the letter below to the Cat. No. for listed Mouthpiece Adapters, e.g. 5003C							
С	Easyhaler®	D	Cyclohaler®	Е	Handihaler®	F	Diskus®
G	Novolizer®	Н	Rotahaler®	1	Turbuhaler®	J	Diskhaler®
К	Respimat®	L	Evohaler®	М	Pari LC Plus®	N	Trudell AeroChamber®
0	Tobi Podhaler®	Р	Ellipta®	Q	Rapihaler®	R	Nexthaler®
S	Qvar® Autohaler®	Т	K-haler®	U	Airomir® Inhaler	٧	PowdAir Plus®

Bespoke Design Available

For any device types not listed above, we offer a custom mouthpiece adapter design service. Simply supply a sample of the inhaler to be tested so that a 'cast' can be taken. This is used to create a moulding tool, which is used to make the mouthpiece adapter.

The tool is then supplied along with the mouthpiece adapter(s) to the user so that it can be reused should any additional mouthpiece adapters be required of that design, in the future.





Mouthpiece Adapter Accessories

Inhaler Support Accessory

For devices that require extra support, the Inhaler Support Accessory holds the device under test in the correct position throughout testing.





Mouthpiece Adapter Rack

To keep benchtops tidy and mouthpiece adapters organised.

Mouthpiece Adapters

Cat. No.	Description
5003	Custom Mouthpiece Adapter for Induction Port, DUSA, WSC2, Filter Holder and Child Alberta Idealised Throat
5004	Tooling Charge for Custom Mouthpiece Adapter
5237	Custom Mouthpiece Adapter for Glass Twin Impinger and FP Induction Port
8515	Custom Mouthpiece Adapter for Adult Alberta Idealised Throat and Albuterol SCA
9013	Custom Mouthpiece Adapter for PTT 1000

Accessories

Cat. No.	Description
5003X	Inhaler Support Accessory
5003Y	Mouthpiece Adapter Engraving (per Mouthpiece Adapter)
5004	Tooling Charge for Custom Mouthpiece Adpater
5005	Mouthpiece Adapter Storage Rack
5022	Certificate of Conformance for Mouthpiece
	Adapter Material



AINI nosepiece adapter

Nosepiece Adapters

We offer nosepiece adapters that create a perfect fit between a nasal device and a DUSA, an induction port, and Glass Expansion Chambers.

A nosepiece adapter for the Alberta Idealised Nasal Inlet (AINI) is also available. Custom-built, the adapater creates an airtight seal between the AINI and test device enabling passive nasal devices to be used under air flow rate (predominantly single dose nasal powders).



Nasal Adapters

Cat. No.	Description
5006	Custom Nosepiece Adapter (for Ind Port and/or DUSA)
8544	Nasal Device Nosepiece Adapter for AINI
8957	Nasal Aerosol Nosepiece Adapter for Expansion
	Chamber Inlet
8958	Tooling Charge (per nasal aerosol device)
8959	Nasal Spray Nosepiece Adapter for Expansion
	Chamber Inlet
8960	Tooling Charge (per nasal spray device)
8956	Expansion Chamber to Flow Meter Adapter
5022	Certificate of Conformance for Noisepiece
	Adapter Material



ACI 28.3 L/min



USP Chapter <601> and Ph.Eur. Chapter 2.9.18 and USP Chapter <1604> specify various types of multi-stage cascade impactor that can be used for measuring the drug-specific aerodynamic particle size distribution (APSD) of orally inhaled and nasal drug products (OINDPs).

This process involves quantitative recovery and chemical analysis of the size-fractionated aerosol, typically by High Pressure Liquid Chromatography (HPLC). From the resulting assay a number of important

metrics can be derived that are used to characterise the APSD, in accordance with pharmacopoeial specifications and various FDA and EMA guidance.




Inhalytix data analysis software is a flexible and fully validated solution for the entry, analysis and reporting of the APSD of drug output from all OINDPs. It also serves as a database for laboratory-based cascade impactor inventory and provides for the setting up and running of detailed test methods. User-configurable,

the software will accept data from standard and customised cascade impactors, including the Andersen Cascade Impactor (ACI), Next Generation Impactor (NGI), Fast Screening Impactor (FSI), Fast Screening Andersen (FSA), Glass Twin Impinger (GTI) and Multi-Stage Liquid Impinger (MSLI).

Licensing

Inhalytix is available as a three user licence software package, based on named users that can be added or removed by the system administrator. The software is available via PC, server and cloud-based installations, with digital licence keys supplied by email. Additional packages of three users are available and can be added to the system at any time.


System Characteristics




Quick and easy to install, Inhalytix is 21 CFR Part 11 ready, enabling the creation of users, assignment of multiple roles (typically admin, supervisor and analyst) and access to audit logs, assisting in data monitoring and ensuring data integrity. The software will operate on Windows 7, 8, 10 and 11 operating systems.

System Operation (Configure > Test > Report)

Dashboard: On entering the software the user is presented with a dashboard providing useful information about how the software is being used. This contains information such as the number of analysts and supervisors set up on the system, as well as the total number of tests prepared, executed and completed. It also summarises the number of tests, equipment and report configurations, as well as details of the equipment inventory, databased by type.




Equipment Types



The software is pre-populated with the most commonly used impactor types for immediate use. However, it is not uncommon for custom versions of cascade impactors to be used in some laboratories. In these circumstances, users can generate bespoke impactor types that can then be stored and recalled for use later. This function may, for example, allow a user to add or remove certain stages from an impactor or add special components to the software, such as modified induction ports.

Equipment Inventory

Keeping track of equipment inventory and associating it with the corresponding inhaler testing data can be a burden. For this reason, the **Inhalytix®** equipment asset library allows users to keep their equipment databased and include equipment-specific data in their testing reports. Not only does this allow users to keep track of equipment, it ensures full traceability by keeping comprehensive records of which specific pieces of equipment were used for each test. Furthermore, the software provides the user with the option to enter impactor-specific mensuration data, allowing the precise calculation of stage cut-off diameters, thereby enhancing the precision and accuracy of test results.




Configuration

Testing of different drug products requires different methods to be in place, different equipment to be used and different metrics to be calculated. This configuration takes place in three easy steps:


Reports • Equipment • Methods


1. Reports

The Reports configuration screen allows users to create tailored report templates, which are then stored and can be paired with different test methods, allowing data to be reported as required.









(1)


1

0

The software allows a high degree of customisation, including both a "Summary" or "Detailed" report template and toggles to turn on or off the reporting of a broad range of metrics. Company logos can be added to the report header if required.

2. Equipment

The equipment configuration screen allows users to generate specific combinations of impactor/impinger and components to match the equipment configuration described in the testing protocol. Users simply drag and drop the impactor and components of their choice into the equipment configurator. This, for example, could see the combination of an NGI, with external filter holder, NGI preseparator, NGI induction port, mouthpiece adapter and inhaler. The software automatically sorts the components into the correct order and ensures that only viable combinations can be created.



(9)

(









Creating a test method allows the user to combine detailed product information, such as drug components and device details, with equipment and report configurations. Users have the opportunity to define for example stage groupings and fine particle dose (FPD) specifications and to select whether delivered dose (when testing MDIs, DPIs, SMIs etc.) or drug substance delivery rate (when testing

nebulisers) is recorded. Data for up to 6 different active pharmaceutical ingredients (APIs) can be added per test method.


Configuring the product specific method is the final step before a user can run a test and analyse their results.

Tests

Once the necessary report, equipment and test method configurations are in place, the user is ready to enter the data and complete the analysis. This function can be found under the 'Tests' tab. Tests are completed in three steps:








All tests are databased and their current status can be monitored to see if they are at the prepared stage, whether results have been entered or whether they are complete.

1. Prepare


To prepare for a test, users are required to recall the test method relating to the product to be tested. During this step, users will have the opportunity to enter test specific information, including the number of runs to be performed.



2. Execute

The user then executes the test by entering the number of doses actuated and drug deposition values for each stage of the impactor, as well as any additional components included in the equipment configuration. This process is then repeated for all additional runs. Alternatively, data can be automatically imported from a CSV or XLSX file.

All values are easily displayed in a scrollable table and can be edited at any point prior to analysis, for example when importing data from HPLC software or exporting data for report writing.





3. Analyse

Once all data has been entered or imported the software analyses the data and presents it to the user in the form of:

- **Results Summary** provides all the key metrics for all test runs in a scrollable table for immediate review.
- Groups Results (where used) displays the drug fractions for each stage or size grouping defined in the method.
- Graphs allows viewing of log-probit plot, drug deposition (by impactor stage/component) and cumulative drug distribution for each run. Also allows the comparison of up to 3 runs from the same test or other tests, so long as the same equipment configuration and data analysis specifications have been set previously.
- Reports allows viewing and printing of standard and customised reports.

Summary of Key Features

- · Standardised approach to the analysis of impactor data
- Ph. Eur. 2.9.18 and USP <1604> compliant
- · 21 CFR Part 11 compliant
- Fully validated with in-built auto-validation protocols
- · Supports PC, server and cloud-based installations
- · Equipment inventory and test-related database
- · Impactor-specific mensuration data log
- Bespoke configurations, methods and reports
- Quick 3-step results analysis: Prepare Execute Analyse
- · Runs and/or Tests comparison capabilities

Inhalytix


mnarytix	
Cat. No.	Description
8260C	Inhalytix Data Analysis Software (3 user licences) - Cloud
8260P	Inhalytix Data Analysis Software (1 user licence) - PC
8260S	Inhalytix Data Analysis Software (3 user licences) - Server
8261	Additional 3 User Licences for Inhalytix (Cloud & Server)
8263	Annual Support Package (per user)



Improving IVIVCs


Predicting the pharmacokinetic and pharmacodynamic (PK/PD) properties of orally inhaled and nasal drug products (OINDPs) using methods such as *in vitro* lung deposition modelling and *in silico* PK modelling can be problematic, given the dynamic nature and complex geometry of the lungs, not to mention the need to consider different lung deposition mechanisms (diffusion, sedimentation, impaction etc.) and patient-to-patient variability.

Making a relatively small investment in systems that enhance the clinical realism of standard pharmacopoeial *in-vitro* test set-ups for the delivered dose uniformity (DDU) testing and aerodynamic particle size distribution (APSD) measurement may help bridge the gap between data collected during quality control (QC) testing and *in vivo* performance helping to accelerate and improve research and development (R&D).



Assessing Drug Efficacy

The core *in vitro* tests for OINDPs, for DDU testing and APSD measurement are highly repeatable and validated methods relied upon for product QC. However, in R&D, the requirement is to understand product behaviour better and optimise performance to deliver targeted *in vivo* drug deposition.



*Fine Particle Dose

In this environment, accuracy and sensitivity alone do not maximise the utility of *in vitro* testing. Due to the complex interactions between formulation and device and the impact of patient-to-patient variability, identifying robust relationships between product characteristics and clinical efficacy can be challenging - very few good IVIVCs exist for OINDPs.

Demonstrating Bioequivalence (BE)

One way to assess *in vivo* performance is to compare the characteristics of a test (T) OINDP, typically a generic, relative to those of a reference (R) product. Demonstrating bioequivalence between T and R reduces the need for clinical testing providing the *in vitro* tests capture variability in *in vivo* behaviour. Better IVIVCs are therefore important for the robust demonstration of BE, a prerequisite for regulatory submissions for generics.

In a similar way better IVIVCs also support Quality by Design (QbD) which calls for the systemic identification and control of all parameters that have an impact on the clinical efficacy of a drug product. *In vitro* methods are therefore far more useful in QbD studies if they accurately reflect *in vivo* behaviour.

For OINDPs it is possible to identify Critical Quality attributes (CQAs) relating to the Patient, Device and Formulation. The impact of variability in all of these parameters is necessarily a focus in product development and more easily studied if the clinical realism of *in vitro* test methods is improved.

Patient
CQAs

Flow resistance
Device
CQAs

Flow resistance
Dose delivery method
(e.g. blister)
Formulation dispersion
method (CFD)

Dose
Dissolution rate
Morphology/particle
interaction

T

Throat geometry
Breath profile
Delivered dose
Aerodynamic particle
size distribution

Flow resistance
Dose delivery method
(e.g. blister)
Formulation dispersion
method (CFD)

Dose
Dissolution rate
Morphology/particle
interaction

R

By Grouping the Critical Quality Attributes (CQAs) based on 'Patient', 'Device' and 'Formulation', a greater understanding of the relative difference between the Test (T) and Reference (R) formulations can be ascertained, accelerating the commercialisation of efficacious products and in the case of generics, a more reliable demonstration of bioequivalence.

Not only this, but this 'sameness' method provides a deeper understanding of the performance between different formulations under test. With this additional data, the most promising candidates can be put forward for clinical trials, potentially reducing the risk of clinical trial failure.

Regulatory Guidance

Enhancing the clinical relevance requirements of *in vitro* testing safeguards data quality, patient safety and clinical efficacy.

Despite the slow uptake of a QbD approach to OINDP development, regulators are now beginning to take a more defined position regarding its implementation.

Improving the clinical relevance of *in vitro* tests and *in silico* models is an important area of focus for both the industry and for regulators, largely because of demand for generic OINDPs. This is reflected in the recent investments made by the FDA for the identification, development and validation of clinically relevant *in vitro* testing methods.

Beclomethasone Dipropionate Inhalation Aerosol Draft Guidance (2019)

The FDA has released product specific draft guidance highlighting the use of novel *in vitro* testing approaches for the assessment of Beclomethasone Dipropionate aerosol as an alternate to a comparative clinical endpoint BE study.

The guidance lists additional supportive *in vitro* studies that can be conducted to support and enhance clinical realism and improve IVIVCs.

These studies include the use of representative mouth-throat models and breathing profiles; the characterisation of aerosol velocity profiles and evaporation rate; drug dissolution testing; and a full assessment of particle morphology.

Designed to bridge the knowledge gap between *in vitro* and *in vivo* OINDP performance, our range of IVIVC test equipment provides analysts with the tools required to assess test products under conditions that more closely replicate *in vivo* performance for the most representative testing. There are a number of ways to adapt the existing regulatory standard systems to improve clinical realism for all inhaled drug types, as shown opposite.

Methods for Improving IVIVCs

DDU and APSD Testing

Realistic Breathing Profiles

Most OINDPs are routinely assessed using constant air flow rate conditions, which are not representative of the inhalation/exhalation profiles of human subjects. Different patients exhibit different breathing profiles, which may affect the efficiency of drug delivery especially for passive devices such as dry powder inhalers (DPIs).



See page 226.

Realistic Throat and Nasal Models

The standard Ph.Eur./USP Induction Port is known to poorly represent aerosol transport through the upper respiratory tract. Using more realistic throat and nasal models enables a more representative assessment of drug delivery to the target site.



See page 228.

Dissolution Testing

In vitro dissolution testing is becoming more widely used for optimising efficacy during drug development, ensuring batch-to-batch consistency and in some cases to predict bioavailability *in vivo* and and help demonstrate BE.



See page 238.

Facemask Testing

In situations where the user lacks the capability of using a mouthpiece (e.g. small children, the elderly), it is commonplace to use a facemask for inhaled drug delivery. The amount of inhaled drug available to the patient is dependent upon the interface between the facemask and the patient and must be rigorously quantified under representative conditions.



See page 244.

Morphology

Profiling the morphological properties e.g. particle size and shape of an inhaled drug formulation may be useful for comparative assessment against a reference drug product notably to assess aerosolisation performance and the extent of deagglomeration.



See page 254.

Cold Freon® Effect

Users of MDIs and nasal sprays may well be familiar with the "cold Freon®" effect - the inadvertent reaction, such as a cough, to the chilling sensation at the back of the throat following actuation of the device. Caused by impaction of the delivered dose and the rapid evaporation of any remaining propellant, the cold Freon® effect strongly influences the efficiency of drug delivery.



See page 255.



Improving IVIVCs

DDU and APSD Testing

Two factors that have been identified as being critical to improving the clinical relevance of DDU testing and APSD measurement are:

Realistic Breathing Profiles



Replacing the existing constant air flow rate conditions used in testing with breathing profiles more representative of the conditions applied by specific patient populations.

Realistic Throat and Nasal Models




In the case of APSD measurement, replacing the existing Ph.Eur./USP Induction Port with an age-appropriate mouth/throat or nasal model with a more realistic human-like geometry.



Realistic Breathing Profiles

Human beings do not breathe at a constant flow rate. Rather the breathing cycles generated by patients produce a continually varying flow rate - very different to the fixed, steady-state flow rates used during *in vitro* testing. Applying more representative breathing profiles can therefore help to achieve better IVIVCs.



Whilst the use of breathing simulators is currently only specified by regulators for the dose uniformity assessment of MDIs with spacers/VHCs and also for nebulisers, they can be applied to the assessment of other OINDPs in order to improve clinical realism of the impactor-sized mass obtained during APSD measurement.

Furthermore, the dose delivery and aerosol generation/ dispersion characteristics of many inhaled products (especially passive devices) are known to be sensitive to flow rate properties, such as acceleration, peak flow and inhaled volume creating an additional incentive for use.



Using data acquired from the clinical use of spirometers, breathing simulators are used to generate representative breathing profiles, offering the chance to more closely assess how factors such as the strength of inhalation and lung capacity can affect the performance for passive devices such as DPIs.

See page 156 for more information about our range of Breathing Simulators.

Mixing Inlet

Applying more representative breathing profiles using a breathing simulator during APSD measurement is complicated by two key issues:

- 1 The impactors used to measure APSD must operate at a constant flow rate.
- The test flow rate applied to the inhaler may need to be lower than the minimum calibrated flow rate of the impactor. For example in paediatric studies a representative flow rate may be 10 L/min but the impactor may have a minimum calibrated operating flow rate of 28.3 L/min.



Mixing Inlet (NGI), Mixing Inlet (ACI)

Our Mixing Inlets are designed to allow the cascade impactor to operate at a constant flow rate, whilst permitting a lower fixed or variable rate to pass through the inhaler. Positioned between the induction port/throat/nasal inlet and cascade impactor, Mixing Inlets decouple the flow rate through the device from the air flow drawn through the impactor, enabling more representative testing.

Mixing Inlet

Cat. No.	Description
8328A	Mixing Inlet for ACI, FSA and MSLI (316 Stainless Steel)
8326	ACI to NGI Outlet Adapter
8327	NGI to ACI Outlet Adapter
8329A	Mixing Inlet for NGI and FSI (316 Stainless Steel)
8324	Set of 2 O-Rings for ACI Mixing Inlet
9160	Compressed Air Flow Controller for Mixing Inlet
9164	Air Compressor for Mixing Inlet
9165	Compressed Air Flow Controller Re-Calibration Certificate
9166	Maintenance Kit for Air Compressor



Real-Time Breath Verification Profile Chamber (BVC)

Breathing Simulator Qualification Tools

We offer an extensive range of qualification tools for our range of Breathing Simulators, including a Real-Time Breath Profile Verification Chamber (BVC) to measure and record the breathing profile generated. See page 156 further information.



Realistic Throat and Nasal Models

The drug mass sized by the cascade impactor (impactor sized mass) should ideally be representative of the dose that would actually enter the lungs. To achieve this, the induction port or other accessory used to interface the device to the impactor must capture a representative fraction of the dose. Knowledge of the portion of the

dose captured in the throat or nasal airway is essential to understand the dosage delivery characteristics of a given OINDP. In many cases, the portion of the dose collected in the throat or nasal airway represents a significant proportion of the delivered dose.

Unlike standard induction ports, our models are optimised for a diverse range of patient profiles. Validated via extensive research, our throat and nasal models are designed to be representative of typical patient populations. Each were developed using insights from CT and MRI scans, direct observations of living subjects, and data from archival literature, each model has a standardised internal geometry that closely mimics *in vivo* physiology.


For further information, please see the following papers:

Grgic, B et al., 2004. Regional aerosol deposition and flow measurements in an idealized mouth and throat. Journal of Aerosol Science. 35; 21-32.

Chen, J et al., 2022. In Vitro Regional Deposition of Nasal Sprays in an Idealized Nasal Inlet: Comparison with In Vivo Gamma Scintigraphy. Pharmaceutical Research. 35; 3021-3028.

Or contact us at info@copleyscientific.co.uk

Alberta Idealised Throat AIT



For orally inhaled products (OIPs), the AIT provides analysts with data more representative of measured *in vivo* behaviour, by ensuring that the ISM corresponds with the portion of the aerosol that would likely enter the lungs.

With a standardised, highly reproducible, humanlike geometry, the AIT offers robust performance independent of flow rate and is designed to make drug recovery quick and simple.

Two versions of the AIT are available:



Both come complete with mensuration and leak test certificates.

Key Features:



8511 Adult Alberta Idealised Throat (AIT) in Aluminium

Accessories

8512	Alberta Idealised Throat to ACI/FSA Adapter
8517	FRS Flow Meter Adapter for Adult Alberta Throat
8513	Alberta Idealised Throat to NGI/FSI Adapter
8514	DFM to Adult Alberta Idealised Throat Adapter
8516	Spare Silicone Seal for Adult AIT
8518	Leak Test Inlet Cap and Outlet Adapter for Adult AIT

Child Alberta Idealised Throat (AIT)

8530 Child Alberta Idealised Throat (AIT) in Aluminium


Accessories

8512	Alberta Idealised Throat to ACI/FSA Adapter
8513	Alberta Idealised Throat to NGI/FSI Adapter
8531	DFM to Child Alberta Idealised Throat Adapter
5239	FRS Flow Meter Adapter
8532	Spare Silicone Seal for Alberta Idealised Throat (Child)
8533	Leak Test Inlet Cap and Outlet Adapter for Child AIT



Outlet adapters available to connect the AIT to NGI, ACI, FSI and FSA

Alberta Idealised Nasal Inlet AINI



Understanding and optimising regional deposition is essential to maximise the fraction of drug absorbed via the target pathway and to minimise drug transit to the lungs. For nasally inhaled products, the AINI enables representative testing of the deposition of drug within the nasal airways.

Made up of 4 separate parts: vestibule, turbinates, olfactory region and nasopharynx, the AINI enables representative testing of drug deposition within the nasal airways. The AINI accurately mimics deposition behaviour in each region, allowing the collection of drug samples that reflect the corresponding fraction of the dose for analysis.

The AINI is easily separated into its component parts to enable drug recovery and assay for each individual area.

Key Features:




Alberta Idealised Nasal Inlet (AINI)

		190	4.10
Cat. No.	Description	49	
8540	Alberta Idealised Nasal Inlet (AINI) for NGI/FSI	76. (II)	1965
8541	Alberta Idealised Nasal Inlet (AINI) for ACI/	***	
8544	Nasal Device Nosepiece Adapter for AINI		3 44 5
8326	ACI to NGI Outlet Adapter		
8327	NGI to ACI Outlet Adapter		
8543	Alberta Idealised Nasal Inlet Leak Test Cap and Inlet Ad	apter	
8546	DFM 2000 to AINI Adapter	Different outlet adapters are available	for a
8547	FRS to AINI Adapter	range of applications	



Improving IVIVCs: Example Test System for Realistic DDU




IVIVC System for DDU Testing of MDIs

- A Dose Uniformity Sampling Apparatus (DUSA) for MDIs
- B Inhaler Testing workstation (ITW)
- DUSA Holder
- C Mouthpiece Adapter

- D Breathing Simulator
- Alternative dose collection device:
- F Alternative dose collection device: Filter Holder

Improving IVIVCs - Realistic DDU Testing System Components



Breathing Simulator BRS

With an intuitive touchscreen interface for easy operation, our range of Breathing Simulators are designed to produce breath profiles across a range of ages (paediatric to geriatric) and patient conditions (mild to severe lung impairment).

For further information about the range, see page 156.













In addition to the Breathing Simulator, the following is needed to complete a fully-operational IVIVC test system for DDU testing:

Dose Collection Device

DUSA for MDIs, SMIs and Nasal Aerosols. See page 20.



DUSA for DPIs and Nasal Powders. See page 22.

Required for:





Filter Holder for MDIs with Spacers/VHCs and Nebulisers. See page 26.









Mouthpiece Adapters

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus.







Nosepiece Adapters

Our Nosepiece Adapters interface the nasal device with the test system.

Required for:







See page 211 for further information.






















Improving IVIVCs:


Example Test System for Realistic APSD



IVIVC System for APSD Measurement of DPIs

- A Compressed Air Source
- B Compressed Air Flow Controller
- C Alberta Idealised Throat
- D Mouthpiece Adapter
- E Mixing Inlet
- F Preseparator

- G Cascade Impactor
- H Breathing Simulator
- Flow Rate Sensor
- J Vacuum Pump
- Critical Flow Controller



IVIVC System for APSD Measurement of Nebulisers

- A Breathing Simulator
- B Compressed Air Source
- C Mouthpiece Adapter
- D Induction Port
- E Mixing Inlet
- F Cascade Impactor

- G NGI Cooler™
- H NGI Cooler Stand for BRS 200i
- Compressed Air Flow Controller
- J Vacuum Pump
- Breath Actuation

Improving IVIVCs - Realistic APSD Testing System Components



Breathing Simulator BRS

With an intuitive touchscreen interface for easy operation, our range of Breathing Simulators are designed to produce breath profiles across a range of ages (paediatric to geriatric) and patient conditions (mild to severe lung impairment).

For further information about the range, see page 156.













Mixing Inlet

Decoupling the flow rate through the device from the air flow drawn through the impactor, the Mixing Inlets are needed to enable the cascade impactor to continue to operate at a constant flow rate, whilst allowing a lower fixed or variable rate to pass through the inhaler.

















Alberta Idealised Throat AIT

With a standardised, highly reproducible, human-like geometry, the AIT offers robust performance independent of flow rate and is designed to make drug recovery guick and simple. Adult and child (6-14 years) versions are available.









Alberta Idealised Nasal Inlet AINI

Mimicking nasal drug deposition behaviour in the vestibule, turbinates, olfactory region and nasopharynx, the AINI helps users to identify the fraction of the drug absorbed via the target pathway and realistically evaluate any unintended drug transit to the lungs.

Required for:























In addition to the Breathing Simulator, Mixing Inlet and a realistic throat/nasal model, the following is needed to complete a fully-operational IVIVC test system for APSD measurement:



Cascade Impactor

Forming the basis of most systems used to measure APSD, a choice of cascade impactors is available depending on device type and application. See page 84 for further information about our range of Cascade Impactors.













Vacuum Pump

Our Vacuum Pump range represents the latest in high performance, low maintenance, technology and is specifically designed for use in the testing of OINDPs. See page 188 for further information about our Vacuum Pump range.



















Critical Flow Controller TPK

Positioned between the cascade impactor and vacuum pump, the Critical Flow Controller TPK 100i ensures critical (sonic) flow conditions during IVIVC testing. This ensures changes to balancing flow from the compressed air supply do not affect the cascade impactor flow rate. See page 172 for further information about our Flow Controller range.














Used for establishing accurate and consistent inlet flow rate during testing, the FRS measures flow within method specification. See page 184 for further information about flow rate measurement.



















NGI Cooler™

Accommodating the NGI both open and closed, the NGI Cooler maintains a temperature-controlled environment throughout testing. Additional space allows for the cooling of extra sets of collection cups, so that multiple tests can be undertaken in quick succession. The NGI Cooler is also suitable for cooling of the Andersen Cascade Impactor ACI. See page 202 for further information.

Required for:







NGI Cooler™ Stand for BRS 200i

Supports interfacing of the NGI Cooler with the BRS 200i, whilst saving precious benchtop space.

See page 202 for further information.

Required for:





Compressed Air Flow Controller

Designed to balance the steady state flow rate entering the impactor, the Compressed Air Flow Controller ensures that the flow rate at the inlet to the induction port is zero prior to starting the test.



















Air Compressor for Mixing Inlet

To provide supplementary air to the inlet port of the Mixing Inlet via the Compressed Air Controller.















Mouthpiece Adapters

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus.











Nosepiece Adapter for AINI

Creating an airtight seal between the AINI and test device, the Nosepiece Adapter for AINI enables passive nasal devices to be used under air flow rate (predominantly single dose nasal powders).

See page 211 for further information.





Qualification

GMP regulations require that

- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.





Improving IVIVCs

Cat. No.	Description
8328A	Mixing Inlet for ACI, FSA and MSLI (316 Stainless Steel)
8326	ACI to NGI Outlet Adapter
8327	NGI to ACI Inlet Adapter
8329A	Mixing Inlet for NGI and FSI (316 Stainless Steel)
8324	Set of 2 O-Rings for ACI Mixing Inlet
9160	Compressed Air Flow Controller for Mixing Inlet
9161	Compressed Air Inlet Manifold for Mixing Inlet
9162	Compressed Air Inlet Manifold for Mixing Inlet & BRS 100i
9163	Compressed Air Inlet Manifold for Mixing Inlet & BRS 200i/300i
9164	Air Compressor for Mixing Inlet
9165	Re-calibration of Compressed Air Flow Controller
9166	Maintenance Kit for Air Compressor







Due to the small size of inhaled drug particles and their typically highly soluble nature, dissolution has always assumed to be very rapid at the site of action. However, the dissolution of inhaled drugs is complicated by a number of issues and is becoming an area of increasing interest for regulators. For example, there is concern that variability between patient groups in the amount and composition of lung and nasal fluid may affect drug uptake. It is important to highlight the value of inhaled dissolution as a BE tool, with the potential to discriminate between formulations of the same drug(s).


Designing a standardised dissolution test method relevant to the lungs is not easy because of the small amount of aqueous fluid involved and the presence of endogenous surfactants. Currently, there are no official dissolution test methods specifically for inhaled products.


One of the main problems facing the developers of such methods is the identification and segregation of that part of the total emitted dose actually reaching the target site (as opposed to the whole dose) in a form readily adaptable to conventional dissolution testing techniques.



The small amount of aqueous fluid and surfactant found in the lung make it extremely difficult to mimic inhaled dissolution testing *in vitro*. Marques, Loebenberg and Almukainzi (2011) list five of the most simulated lung fluids in Table 11 of their article 'Simulated Biological Fluids with Possible Application in Dissolution Testing'. Read it to find out more.

We offer a range of equipment designed for particle selection, dose collection and dissolution testing, to help analysts identify, segregate and assess the dissolution characteristics of inhaled drug products.







1 & 2. Particle Selection & Dose Collection

Next Generation Impactor (NGI)

A modification of the standard NGI Collection Cup, the NGI Dissolution Cup and Membrane Holder enables size-fractionated particles from an aerosol cloud to be collected and tested using a conventional tablet dissolution tester.

NGI Dissolution Cups

The NGI Dissolution Cup differs from the standard cup in that it has a 50 mm removable insert in the impaction area.

- 1 Particle sizing is carried out in the conventional manner.
- 2 Following collection, the insert is carefully removed from
- The insert is covered with a pre-punched 55 mm diameter polycarbonate membrane and secured in position in a Membrane Holder, using a ring, to form a sealed "disc" or "sandwich".
- The Membrane Holder is then placed in a conventional Dissolution Tester, such as Copley's DIS 800i and tested in a manner similar to the 'Paddle Over Disc' method described in the Pharmacopoeias.





NGI Dissolution Cup and Membrane Holder



Andersen Cascade Impactor (ACI)

Following a similar technique to that used for the NGI, with the ACI the drug is instead captured directly onto the membrane prior to analysis.

- A 76 mm polycarbonate membrane is applied to the collection plate prior to particle sizing.
- 2 Particle sizing is carried out in the conventional manner.
- The membrane is inverted and sandwiched between the glass and mesh surfaces of the Watchglass/PTFE Assembly (traditionally used for transdermal patches).



Watchglass/PTFE Assembly for use with ACI

NGI Dissolution Cups

Cat. No.	Description
----------	-------------

6001 NGI Dissolution Cup and Membrane Holder (each)

6002 55 mm Punch

6004 Pack of 100 Polycarbonate Filters

(0.1 micron x 76 mm diameter)

6005 Spare Set of O-Rings

ACI with Membrane

Cat. No. Description

6003 Watchglass/PTFE Assembly for use with ACI (each)

Pack of 100 Polycarbonate Filters

(0.1 micron x 76 mm diameter)

3. Dissolution Testing

We offer USP Method 2 dissolution testers for use with the NGI and ACI Membrane Holders.

Further details about our range of dissolution testers can be found in our sister brochure:

"Driving Results in Pharmaceutical Testing".



The following is needed to complete a fully-operational test system for inhaled dissolution dose collection:



Cascade Impactor

Use of a cascade impactor allows size fractionated particles from an aerosol cloud to be collected for testing.

For further information about our range of Cascade Impactors, please see page 84.

Vacuum Pump

Our Vacuum pump range represents the latest in high performance, low maintenance technology and is, specifically designed for use in the testing of OINDPs.

See page 188 for further information about our Vacuum Pump range.





Flow Controller

Suitable for controlling air flow rate across the range required for OINDP testing reproducibility and the ease of method transfer, reducing potential sources of data variability.

See page 172 for further information about our Flow Controller range.

Flow Rate Sensor FRS

Used for establishing accurate and consistent inlet air flow rate during testing, the FRS measures flow within method specification.

See page 184 for further information about flow rate measurement.





Mouthpiece Adapters

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus.







Nosepiece Adapters

Our Nosepiece Adapters interface the nasal device with the test system.

Required for:







Qualification

GMP regulations require that

- · The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing



Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.





Improving IVIVCs

Facemask Testing

In many cases, inhaled drug products may be administered using a facemask instead of a mouthpiece. This is often the case for infants and small children and in other situations where the user lacks the capability to use a mouthpiece.

A key factor in determining the amount of inhaled drug available to the patient is the interface between the facemask and the patient. A properly sized mask, firmly placed against the face, for example, will provide the user with far more drug than a poorly fitting equivalent where much of the drug is lost to the environment through leakage.

Due to the important role that a facemask has in transporting the drug aerosol from the device to the patient, further assessment is required in addition to the standard DDU testing and APSD measurement methods routinely applied.

Relevant for two types of devices:



MDIs used with a spacer/VHC and a facemask



Nebulisers used with a facemask

Face Models

A critical component of the test apparatus used for facemask testing is the face model. This should be appropriate to the age group for which the product is intended, e.g. infant, child or adult. Face models should:





Achieve realistic dead space within the mask and at the same time ensure the absence of leaks between the mask and model.



Have physiologically accurate soft facial tissue to simulate *in vivo* conditions.



Provide a means of mounting the spacer/VHC or nebuliser such that the facemask is in correct alignment with the face model as in "real-life" conditions.

We offer a range of facemask testing systems for different devices, which seek to address the above requirements, whilst also providing sufficient flexibility to allow users to utilise their own validated models, if desired. All models are fitted with replaceable face skins.



Filter Holder and Adapter located in a cavity behind the face model's lips

Face Model Products


Cat. No.	Description
9142	FMA/FMS Filter Holder and Adapter for BRS 100i
9143	FMA/FMS Filter Holder and Adapter for BRS 200i/300i
9103	Pack of 100 Filters for Filter Holder
9144	Adult Head and Adapter for FMA/FMS
9145	Child Head and Adapter for FMA/FMS
9146	Infant Head and Adapter for FMA/FMS
9149	Replacement Face Skins for Adult Head (Pack of 6)
9150	Replacement Face Skins for Child Head (Pack of 6)
9151	Replacement Face Skins for Infant Head (Pack of 6)



Test Systems for Assessing Facemask Performance

Two types of apparatus are available, each providing standardised test methods to quantify the effect of using a facemask on drug delivery from the device under test.

1. Facemask Testing Apparatus FMA for MDIs with a Spacer/VHC




Facemask Testing Apparatus FMA

Cat. No.	Description
9141	Facemask Test Apparatus for Spacers & VHCs Model FMA
9142	FMA/FMS Filter Holder and Adapter for BRS 100i
9143	FMA/FMS Filter Holder and Adapter for BRS 200i/300i

FMA: DDU Testing

- (A) Breathing Simulator BRS
- B Filter Holder and Adapter
- C Face Model
- D Facemask Apparatus FMA



Products Featured in this System



Facemask Testing Apparatus FMA

The FMA is designed to meet all the critical requirements for assessing the impact of facemasks on performance of MDIs with a spacer/VHC.

In addition to the above, the following is needed to complete a fully-operational DDU test system for assessing the impact of facemasks on the performance of MDIs with a Spacer/VHC:

Face Model

Models are available for all age groups - adult, child and infant. All models are fitted with replaceable face skins which provide flexibility and elasticity similar to real life tissue.





Filter Holder & Adapter

Positioned in the cavity behind the lips of the face model, the holder contains a filter to capture the active drug from the device under test.

See page 25 for further information.


Breathing Simulator BRS

Providing breathing profiles that are more clinically representative than a constant flow rate, the Breathing Simulator BRS 200i is ideal for assessing the impact of a facemask on the DDU of MDIs with a spacer/VHC. Alternatively, a basic entry-level model, the Breathing Simulator BRS 100i, is also available.

Find out more about our range of Breathing Simulators on page 156.



FMA: APSD Measurement



Products Featured in this System



Facemask Testing Apparatus FMA

The FMA is designed to meet all the critical requirements for assessing the impact of facemasks on the performance MDIs with a spacer/VHC.

In addition to the above, the following is needed to complete a fully-operational APSD measurement set-up for testing the performance of MDIs with a Spacer/VHC when used with a facemask.

Face Model

Models are available for all age groups - adult, child and infant. All models are fitted with replaceable face skins which provide flexibility and elasticity similar to real life tissue.



Filter Holder & Adapter

Positioned in the cavity behind the lips of the face model, the holder contains a filter to capture the active drug from the device under test.

See page 25 for further information.





Next Generation Impactor NGI

The APSD characterisation of facemask performance should be conducted using an NGI.

See page 84 for further information.

FMA to NGI Interface Accessory

Provides a direct connection between the FMA and Face Model which is mounted onto the inlet of the NGI Induction Port.





Flow Controller

Suitable for setting flow rate and sampling time delays, as well as controlling inhaled volume, our range of Flow Controllers improve testing reproducibility and the ease of method transfer, reducing potential sources of data variability.

See page 172 for further information about Flow Controller Range.

Flow Rate Sensor FRS

Used for establishing accurate and consistent inlet flow rate during testing, the FRS measures flow within method specification.

See page 184 for further information about flow rate measurement.





Vacuum Pump

Our Vacuum Pump range represents the latest in high performance, low maintenance, technology, and is specifically designed for use in the testing of OINDPs.

See page 188 for further information about our Vacuum Pump range.

Qualification

GMP regulations require that


- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.



2. Facemask Testing Stand FMS for Nebulisers




Facemask Testing Stand FMS

Cat. No.	Description
9156	Facemask Stand for Nebulisers Model FMS
9142	FMA/FMS Filter Holder and Adapter for BRS 100i
9143	FMA/FMS Filter Holder and Adapter for BRS 200i/300i

FMS: DDU Testing

- (A) Breathing Simulator
- B Face Model
- c Facemask Stand (FMS)



Products Featured in this System



Facemask Stand FMS

The FMS is designed to meet all the critical requirements for assessing the effect of facemasks on the use of nebulisers.

In addition to the above, the following is needed to complete a fully-operational DDU test system for assessing the impact of facemasks on nebuliser performance:

Face Model

Models are available for all age groups - adult, child and infant. All models are fitted with replaceable face skins which provide flexibility and elasticity similar to real life tissue.





Filter Holder & Adapter

Positioned in the cavity behind the lips of the face model, the holder contains a filter to capture the active drug from the device under test.

See page 25 for further information.


Breathing Simulator BRS

Providing breathing profiles that are more clinically representative than a constant flow rate, the Breathing Simulator BRS 200i is ideal for assessing the impact of a facemask on the DDU of nebuliers. Alternatively, a basic entry-level model, the Breathing Simulator BRS 100i, is also available.

Find out more about our range of Breathing Simulators on page 156.



FMS: APSD Measurement



Products Featured in this System



Facemask Stand FMS

The FMS is designed to meet all the critical requirements for assessing the effect of facemasks on the use of nebulisers.

In addition to the above, the following is needed to complete a fully-operational APSD measurement system for assessing the impact of facemasks on nebuliser performance:

Face Model

Models are available for all age groups - adult, child and infant. All models are fitted with replaceable face skins which provide flexibility and elasticity similar to real life tissue.





Filter Holder & Adapter

Positioned in the cavity behind the lips of the face model, the holder contains a filter to capture the active drug from the device under test.

See page 25 for further information.

Next Generation Impactor NGI

The APSD characterisation of a nebuliser should be conducted using an NGI, because it has calibrated performance at the 15 L/min test rate specified for nebulisers.





FMS to NGI Interface Accessory

Provides a direct connection between the FMS and Face Model that is mounted onto the inlet of the NGI Induction Port.

Flow Rate Sensor FRS

Used for establishing accurate and consistent inlet flow rate during testing, the FRS measures flow within method specification.







Flow Controller

Suitable for setting flow rate and sampling time delays, as well as controlling inhaled volume, our range of Flow Controllers improve testing reproducibility and the ease of method transfer, reducing potential sources of data variability.

See page 172 for further information about Flow Controller Range.

Vacuum Pump

Our Vacuum Pump range represents the latest in high performance, low maintenance, technology, and is specifically designed for use in the testing of OINDPs.

See page 188 for further information about our Vacuum Pump range.







The NGI Cooler™ can only be used for nebulisers with mouthpieces. For nebulisers with facemasks the NGI will need to be removed from the NGI Cooler for testing, once the required temperature has been reached.

Oualification

GMP regulations require that

- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.



Morphology

Cascade impactors separate the delivered dose from an inhaled product on the basis of particle inertia, producing sized fractions which are then subject to chemical assay to produce an APSD for the active drug.

Whilst this process provides a useful indication of where inhaled drug particles are likely to deposit within the respiratory tract, it does not profile the morphological properties of these particles. Generating component specific particle geometric size and shape data may be helpful in understanding differences between formulations and hence their potential bioavailability,

even when APSDs are equivalent. This can be particularly useful in generic development when trying to replicate the performance of a reference product. The Malvern Glass Disc Cup, allows for collection of particles on a quartz glass disk, which can then be transferred to a Malvern Panalytical Morphologi 4-ID or equivalent system for morphological analysis.



Morphology Sampling Apparatus

Cat. No. Description

5242A Malvern Glass Disc Cup, Small

(for Morphologi 4-ID system)

Cold Freon® Effect

The cold Freon®effect is the inadvertent reaction to the chilling sensation at the back of the throat following the actuation of MDIs, and it can significantly influence the efficiency of drug delivery. For example, the effect may cause the patient to cough, or abort the inhalation manoeuvre, resulting in inconsistent dose delivery.

Spray pattern and plume geometry are common measurement techniques employed by the pharmaceutical industry to characterise the emitted spray from MDIs and nasal sprays. However, the reaction of the user to the impaction force of the spray on the throat or nasal passageways is also of much concern.



Cold Freon® effect assessment is important in switching propellants for MDIs. For example, reformulation of CFC to HFA 134a and HFA 152a

Novel Inhaled Formulations

Assessing the cold Freon® effect of a new MDI or nasal formulation is valuable in evaluating and minimising the potential for any unintended reaction by the patient which may impede drug delivery. Assessing the spray force and plume temperature of a given formulation when actuated as per the manufacturer's instructions can give a good indication of whether either of these parameters may induce an adverse reaction by the patient when used in real life.

Generic Inhaled Formulations

An assessment of the cold Freon® effect of generic formulations can also provide useful supportive evidence for the demonstration of BE. Comparative measures of impaction force and temperature are a good indicator of local delivery equivalence, or otherwise, and help to confirm that in clinical use the generic will be interchangeable with the reference product. Since velocity is directly related to the impaction force and temperature, the latter should be a good indicator of local delivery equivalence for an inhaled drug.

Copley offers two types of test apparatus to assess cold Freon®.



Spray Force Tester



Plume Temperature Tester



Drug A and Drug B demonstrate bioequivalence in vitro, however, differences in their cold Freon® characteristics may cause differences in in vivo performance



Spray Force Tester SFT 1000

Offering high precision impaction force testing for MDIs and nasal sprays, the Spray Force Tester SFT 1000 provides analysts with a simple and reliable way of assessing the effects of cold Freon® on the throat and nasal cavity over the duration of the spray plume.





High sensitivity digital load cell



Pass/Fail alarms for userprogrammable limits (for QC)




Memory capability for up to 100 spray force measurements



Load cell calibration verification easily performed by user

Key Features:



A sample of the inhaler to be tested is required at the time of placing an order so that a customised clamp can be made.

SFT 1000: Technical Specifications

Flow Rate Range	0 to 2500 mN
Accuracy	+/-2.5 mN
Adjustable Distance	The distance of the device relative to the impaction plate can be adjusted between 0 and 200 mm +/-0.03 mm using the precision digital gauge.
Power	Battery or mains powered
Dimesions (L x W x H)	580 mm x 200 mm x 80 mm
Reporting	RS-232 output to computer or printer



Supplied complete with calibration certificates for load cell and gauge

Spray Force Tester SFT 1000

Cat. No.	Description	Cat. No.	Description
9000	Spray Force Tester Model SFT 1000	9005	Digital Mini Processor (Statistical Printer)
9001	Additional Device Clamp	9006	IQ/OQ Documentation for SFT 1000
9002	Re-calibration of Spray Force Load Cell	9007	Qualification Tools for SFT 1000
9003	Re-calibration of Digital Gauge	9008	Re-calibration of SFT 1000 Qualification Tools
9004	Spare Impaction Plate		

Plume Temperature Tester PTT 1000

Providing analysts with a quick and easy method for assessing aerosol plume temperature, the PTT 1000 is ideal for the sensitive profiling of MDIs.

The outlet of the PTT 1000 is normally connected to a waste shot collector and vacuum pump to capture the measured doses at the relevant flow rate. It can, however, easily be connected directly to a DUSA collection tube or Induction Port if preferred, since the outside diameter of all three accessories are identical.



Products Featured in this System



Plume Temperature Tester PTT 1000

The PTT 1000 is supplied together with the data acquisition assembly, sampling manifold assembly, flow meter adapter and software.

In addition to the above, the following is needed to complete a fully-operational plume temperature test system for MDIs:

Vacuum Pump

Our Vacuum Pump range represents the latest in high performance, low maintenance, technology, and is specifically designed for use in the testing of OINDPs.

See page 188 for further information about our Vacuum Pump range.



Flow Meter DFM 2000

Used for establishing accurate and consistent inlet flow rate during testing, the DFM 2000 measures and controls flow within method specification.

See page 184 for further information.





Waste Shot Collector WSC2

A compact vacuum filtration system, the Waste Shot Collector WSC2 safely captures aerosols emitted from repeated actuations of the inhaler.

See page 24 for further information.

Mouthpiece Adapters

Moulded from high quality silicone rubber, our Mouthpiece Adapters guarantee an airtight seal between the inhaler under test and the test apparatus.

See page 211 for further information.



Qualification

GMP regulations require that

- The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

Copley provides a range of qualification documentation, services and tools to meet these requirements.

See page 310 for further information.



Plume Temperature Tester PTT 1000

Cat. No.	Description
9010	Plume Temperature Tester Model PTT 1000 (incl. Software)
5001	Waste Shot Collector WSC2
9013	Shortened Mouthpeice Adapter
9011	IQ/OQ Documentation for PTT 1000
9012	Re-calibration of 4 Thermocouples

Special Applications

We offer a range of specialised test equipment for specific applications relating to the performance assessment of orally inhaled and nasal drug products (OINDPs).



Abbreviated Impactor Measurement AIM

The drive for greater efficiency is stimulating debate as to whether full-resolution, multiple-stage cascade impaction can be supplemented with AIM as part of a Quality by Design (QbD) process.

Once the full APSD profile of a product has been established, AIM may be useful as a rapid screening tool in R&D and, with the use of appropriate metrics, in QC applications also.

We also offer a tool that enables analysts to match the flow resistance and flow rate rise-time profiles between a full resolution impactor and its abbreviated counterpart to ensure comparable conditions for aerosol generation, supporting improved equivalence in aerodynamic particle sizing measurements.

See page 261 for further information.

Generic Drug Development

There is growing interest in the development of generic orally inhaled products (OIPs) as the patents on the original products expire. This has led to the reintroduction into the pharmacopoeias of some of the test methods employed in the development of the original drug products.

See page 270 for further information.

Device Robustness/Inhaler Misuse

Device mishandling and poor technique are widely recognised issues associated with the use of inhalers, resulting in inadequately controlled respiratory disease and an over-reliance on emergency remedies.

We offer solutions to aid those developing inhaled drug devices and products in understanding the impact of poor patient technique on the critical quality attributes (CQAs) of inhalers to help optimise inhaler designs for more robust drug delivery.

See page 276 for more information.

Abbreviated Impactor Method AIM

Background

Due to the unique nature of their part device/part formulation, the practical application of QbD principles to OINDPs is not easy.

The preferred and current instrument of choice for measuring the aerodynamic particle size distribution (APSD) of OIPs for both regulators and pharmacopoeias is the cascade impactor (see page 84). Whilst providing a detailed size classification of the aerosol cloud concerned, recent QbD initiatives have highlighted that full resolution multi-stage cascade impaction

methods may not only be time-consuming but also require a high degree of skill and consistency on the part of the analyst if error is to be avoided.

For these reasons and with the adoption of QbD potentially increasing demands for analytical data, attention has turned to the concept of AIM.


AIM in QC

For OIP product batch release testing and QC applications, it is possible to use simpler but highly sensitive metrics to determine if the product is fit for purpose once a full APSD profile has been established using a full-resolution cascade impactor. This is known as Efficient Data Analysis (EDA).

Typically, the APSDs of inhaled products exhibit a Normal (or Gaussian) Distribution centred around the Mass Median Aerodynamic Diameter (MMAD). It is therefore possible to determine even subtle changes in the APSD by measuring the following:

1. Impactor Sized Mass (ISM): the sum of the drug mass deposited on the filter and all impactor stages where the upper-bound size of entering particles is known. This metric indicates any shift in the amplitude of the APSD.

2. Ratio of Large Particle Mass to Small Particle Mass (LPM/SPM): determined by splitting the ISM into two fractions on either side of the MMAD: LPM greater than the MMAD and SPM smaller than the MMAD. This ratio indicates any shift in the central tendency of the APSD.






Although EDA can be applied to full-resolution impactor testing, its true value comes from combining it with AIM, which uses only a reduced number of impactor stages, speeding up throughput and further reducing analytical error. Full-resolution impactor testing is then reserved for out-of-specification (OOS) investigations.

In this diagram, the AIM-QC model shows how abbreviating the ACI to just 2 stages and a filter, with the central stage (Stage X) selected to have a cut-off diameter close to the product MMAD allows the EDA metrics of ISM and LPM/SPM to be easily determined.

The table on page 92 indicates which stage can be used for Stage X.

Adapted from: Mitchell, J.P. et al. Relative Precision of Inhaler Aerodynamic Particle Size Distributon (APSD) Metrics by Full Resolution and Abbreviated Andersen Cascade Impactors (ACIs): Part 1., AAPS PharmSciTechnol., 2010, 11(2): 843-851



AIM in R&D


AIM has also been suggested as a useful tool in R&D for the fast screening of new formulations in product development.

An important aim is to establish how to generate clinically representative data to reduce the dependence on time-consuming and expensive clinical trials.

This is not easy; as has been mentioned before, a cascade impactor is not analogous to the lung. The lung is a complex organ, with high humidity, decreasing velocity with each bifurcation and complex deposition mechanisms (diffusion and sedimentation, as well as impaction). This makes correlation between *in vitro* cascade impactor measurements and deposition in the Human Respiratory Tract (HRT) highly complex.

There is some evidence to suggest that abbreviated versions of full stack cascade impactors can be used to broadly indicate *in vivo* lung deposition based on two or three size bands (or fractions):

- 1. Coarse Particle Mass (CPM) That portion of the aerosol considered to be too large to be inhaled (usually considered to be >5 microns).
- 2. Fine Particle Mass (FPM) That portion between 5 and 1 micron, usually considered likely to deposit deep into the lung and hence be therapeutically effective.
- **3. Extra-fine Particle Mass (EPM)** That portion below 1 micron, usually considered to be too small to deposit in the lung and potentially exhaled.



Adapted from: Mitchell, J.P. et al. Relative Precision of Inhaler Aerodynamic Particle Size Distributon (APSD) Metrics by Full Resolution and Abbreviated Andersen Cascade Impactors (ACIs): Part 1., AAPS PharmSciTechnol., 2010, 11(2): 843-851

AIM - The Future

To meet these various demands and to provide a basis for the proof-of-concept work necessary to validate them, Copley has introduced a number of different versions of abbreviated impactor for use in both QC (QC Models) and R&D (HRT Models). These are based on stage versions of the popular Andersen Cascade Impactor (ACI) and Next Generation Impactor (NGI).

If validated and implemented, these impactors could help to speed up formulation screening, prior to full resolution impactor studies being performed on the most promising candidates and then subsequent used for product release in QC.

Fast Screening Andersen FSA

FSA is an AIM version of the standard ACI suitably modified to provide a reduced stack plus filter (F) suitable for either:



Quality Control (FSA-QC)

Stages 0 (or -1, or -2A) and F are used in conjunction with a Stage X, with a cut-off diameter as close as possible to the MMAD of the aerosol, as determined during full resolution cascade impactor testing.

Product Development (FSA-HRT) with Realistic Throat and Nasal Models

Stages with cut-off diameters are available at 5.0 and 1.0 microns for metered-dose inhaler (MDI) applications at 28.3 L/min. Also, for this and higher flow rates (60 and 90 L/min) stages having traditional ACI cut points of 4.7 and 1.1 microns are available, primarily for dry powder inhaler (DPI) applications.

Find out more about the our Realistic Throat and Nasal products on page 228.



In addition to the FSA, the following ancillaries are required to complete a fully operational test set-up for determining the CPM, FPM, ePM, or LPM/SPM ratio:



FSA-QC with Stage X cut-off diameter close to product MMAD

Cat. No.	Description
8341	FSA-QC - 28.3 L/min (Stages 0, X and F)*
8342	FSA-QC - 60.0 L/min (Stages -1, X and F)*
8343	FSA-QC - 90.0 L/min (Stages -2A, X and F)*

FSA-HRT with cut-off diameters of 5.0 and 1.0 or 4.7 and 1.1 microns

8344	FSA-HRT - 28.3 L/min (Spacer, Stages 5.0 and 1.0 micron, and F)*
8345	FSA-HRT - 28.3 L/min (Spacer, Stages 2, 5 and F)*
8346	FSA-HRT - 60.0 L/min (Spacer, Stages 1, 4 and F)*
8347	FSA-HRT - 90.0 L/min (Spacer, Stages -0, 3 and F)*

Induction Ports

8501	USP Induction Port*
8510	USP Induction Port (One-piece 316 Stainless Steel)
8060	Flow Meter to Induction Port/WSC2 Adapter
5239	FRS Flow Meter Adapter
5238	DFM Flow Meter Adapter

Preseparators for testing DPIs

8401	28.3 L/min Preseparator*
8420	60 L/min Preseparator*
8420-90	90 L/min Preseparator*

Spare Parts

8367-I	Stage 5.0 micron cut-off @ 28.3 L/min*
8368	Stage 1.0 micron cut-off @ 28.3 L/min*
8371	FSA Spacer Stage*
8334	Complete Set of 7 Silicone Rubber O-Rings
8335	Set of 2 Stainless Steel Collection Plates (28.3 L/min)
8336	Set of 2 Stainless Steel Collection Plates (60 or 90 L/min)
8316	Box of 100 Glass Fibre Filters
8308A	Set of 3 Shortened Spring Clamps - 4 Stage
8308B	Set of 3 Shortened Spring Clamps - 3 Stage

^{*}Please specify Aluminium (A), 316 Stainless Steel (S) or Titanium (T) when placing your order.

Reduced NGI rNGI

The individual stages of the NGI are fixed within the seal body, such that they cannot be removed. However, the NGI can be used in an abbreviated form, the rNGI, for both AIM-QC and AIM-HRT applications.


As with the FSA, and depending on the flow rate to be used, a stage between 2 and 4 (see blue highlights in the table below) of the NGI can be selected with a cut-off diameter close to the product's MMAD (AIM-QC application) or close to 5 microns (in the case of an AIM-HRT application).

The rNGI Filter Holder Assembly is placed in the stage immediately after the cut-off stage selected.

It consists of a filter support mesh which is placed on top of the stage nozzles and a split ring used to hold the filter in position on top of the filter support mesh.

When operating the rNGI, particles smaller than the cut-off diameter of the stage preceding the rNGI Filter Holder Assembly will be captured on the paper filter of the rNGI, whilst particles larger than the cut-off diameter will impact as normal in the collection cups of those stages upstream.

Note: when using the rNGI Filter Holder Assembly, it is not possible to have a second stage representing the Extra-fine Particle Mass (EPM).



rNGI

5259 52504 rNGI Filter Holder Assembly

5259A Pack of 100 Filters

	Stage Cut-off Diameters for the NGI at Different Flow Rates									
	Flow Rate (L/min)									
		15	30	40	50	60	70	80	90	100
	1	14.10	11.72	10.03	8.89	8.06	7.42	6.90	6.48	6.12
	2	8.61	6.40	5.51	4.90	4.46	4.12	3.84	3.61	3.42
Stage	3	5.39	3.99	3.45	3.09	2.82	2.61	2.44	2.30	2.18
Sta	4	3.30	2.30	2.01	1.81	1.66	1.54	1.45	1.37	1.31
	5	2.08	1.36	1.17	1.04	0.94	0.87	0.81	0.76	0.72
	6	1.36	0.83	0.70	0.61	0.55	0.50	0.46	0.43	0.40
	7	0.98	0.54	0.45	0.38	0.34	0.31	0.28	0.26	0.24



The rNGI Filter Holder Assembly should be placed in the stage immediately after the stage with the desired cut-off diameter.

Fast Screening Impactor FSI

Based on proven NGI Preseparator technology, the FSI represents a purpose-made approach to AIM that separates the dose into CPM and FPM making it suitable for AIM-HRT applications (i.e. FSI-HRT) for MDIs, DPIs and nasal sprays.

A range of inserts are available, to generate a 5 micron cut-off diameter within the flow rate range of 30-100 L/min at 5 L/min intervals. This makes the FSI ideal for DPIs tested at a flow rate that equates to a 4 kPa pressure drop over the inhaler.

The FSI uses the same induction port as the NGI. It employs a two-stage separation process in which first large non-inhalable boluses are captured in a liquid trap followed by a fine-cut impaction stage at 5 microns. This gives unparalleled accuracy, high capacity, low internal losses and low carryover.

The fine particle dose is collected on a glass fibre filter located in an external filter holder with quick-release catches for easy access.

An additional insert is available for generating a 10 micron cut-off diameter at 30 L/min. When used with a Glass Expansion Chamber (see page 208) this makes the FSI ideal for the fast screening of nasal aerosols and sprays.



Fast Screening Impactor FSI



Filter Holder





Interchangeable Inserts

In addition to the FSI, the following ancillaries are required to complete a fully operational test set-up for determining the CPM, FPM, or LPM/SPM ratio:



Fast Screening Impactor FSI complete

Cat. No.	Description
5260	FSI complete with one insert (please specify flow rate - see below)
5261	Additional Inserts - 5 microns @ 30, 35, 40, 45, 50, 55, 60,
	65, 70, 75, 80, 85, 90, or 100 L/min for MDIs or DPIs
	(please specify flow rate)
5240	Box of 100 Filters (for Fine Fraction Collector)

Fine Fraction Collector for users that already have NGI Preseparator

Fine Fraction Collector only

Note: For a complete system, users must also purchase an insert (see 5261) to replace the existing insert in their preseparator

Accessories for MDIs and DPIs

5203	NGI Induction Port
5239	FRS Flow Meter Adapter
5238	DFM Flow Meter Adapter
5204	NGI Presenarator

Accessories for Nasal

5263 Additional Insert - 10 microns @ 30 L/min for Nasal Sprays

Volume and Resistance Compensator VRC

Patent Pending



Abbreviated impactors are designed to help reduce the burden of full resolution cascade impaction studies, following proper aerodynamic particle size distribution (APSD) profiling.

However, differences in total volume and flow resistance between a full resolution cascade impactor and its abbreviated counterpart is known to cause variability in the flow rate rise-time profiles between the two test set-ups. This difference reduces parity between test conditions, especially for passive, dry powder inhalers (DPIs) where start-up kinetics can be important.

The Volume and Resistance Compensator (VRC) enables analysts to match the flow resistance and flow rate rise-time profiles between the two test set-ups to ensure comparable conditions for aerosol generation, supporting improved equivalence in aerodynamic particle sizing measurements.

Key Features:





Suitable for use with all types of abbreviated impactors



Easily adjustable for different set-ups



Designed using Copley expertise




Ensures parity between test methods, improving data comparability



/olume and flow resistance can be independently varied for precise flow ate rise-time matching.



The VRC connects inline between a critical flow controller and an abbreviated impactor. Shown here (L to R): Fast Screening Impactor (FSI), Volume and Resistance Compensator (VRC), Critical Flow Controller (TPK 100i-R) and High Capacity Vacuum Pump (HCP6).



VRC Performance

This graph demonstrates how the NGI rise-time can be closely matched using the VRC with the FSI.

The purple line is the rise-time line with the **FSI only**.

The orange line is the rise-time with the **NGI only**.

The green line is the rise-time with **both the FSI and VRC**.

The VRC was adjusted to first match the flow resistance of the NGI, then the volume was adjusted to match the rise-time.

Volume and Resistance Compensator VRC

Cat. No. Description

5280 VRC - Volume and Resistance Compensator



Special Applications

Generic Drug Development

The success of a generic drug formulation submission relies on the robust demonstration of bioequivalence (BE) to a reference labelled drug (RLD). This normally involves the provision of *in vitro* data to demonstrate that the generic will perform in a clinically identical way to the RLD.

The FDA regularly issue product-specific guidance for several active pharmaceutical ingredients (APIs)that are used globally for the treatment of asthma and COPD and are consequently routine targets for generic development. The USP also have introduced product-specific monographs for Fluticasone Propionate (FP) and Salmeterol.

The original product-specific monographs called for the use of test equipment based on methods used in the original development of these products.

The USP listed four such monographs for FP and FP/ Salmeterol combination products:

- Two relate to the use of the APIs as aerosols delivered by an MDI
- Two are for APIs prepared as inhalation powders for delivery by a DPI

A further monograph for Albuterol Inhalation solution products has been approved.

The product-specific monographs concerned covered both DDU testing and APSD measurements. DDU and APSD are required performance metrics for all OIPs because of their defining influence on the success and consistency of drug delivery.

Fluticasone Propionate/Salmeterol Aerosols & Powders

The inhalation powder monographs require that DDU measurements be conducted for a duration consistent with the withdrawal of 2 litres of air. This volume is generally considered to be representative of a typical patient with asthma or COPD.

APSD measurement was originally conducted using a standard ACI equipped with a specially modified induction port common to both aerosols and powders and a specially modified inlet cone and preseparator for aerosols and powders respectively.

According to the monographs, the 28.3 L/min version of the ACI (Stages 0 to 7 plus filter stage) was used to measure APSD for both aerosols and powders despite the fact that the powder method specifies testing at 60 L/min.

The duration of testing for APSD measurements is adjusted to give the volumetric equivalent of 3 litres of air. This is likely due to the need to achieve adequate volume changes in the ACI.

FP/Salmeterol Aerosols

Apparatus requirements:

Sample Collection Apparatus for FP/Salmeterol Aerosols Aerodo

Aerodynamic Particle Size Distribution


Andersen Cascade Impactor (ACI)

FP/Salmeterol Induction Port

ACI Inlet Cone for FP/Salmeterol Aerosols

FP/Salmeterol Powders

Apparatus requirements:



In addition to the above and previous page, the following are recommended to complete a fully-operational test set-up for the DDU testing and APSD measurement of **FP/Salmeterol Aerosols & Powders**.



Apparatus for DDU testing of FP/Salmeterol Products

Cat. No. Description

8646 Sample Collection Apparatus for FP/Salmeterol Aerosols8640 Sample Collection Apparatus for FP/Salmeterol Powders

Spare Parts for Sample Collection Apparatus for Aerosols

8649 Pack of 500 Cotton Wool Balls

8647 Separating Flask8648 Flow Meter Adapter8650 Vacuum Pump Adapter

Spare Parts for Sample Collection Apparatus for Powders

8641 Pack of 100 Glass Fibre Filters 70 mm

8903 Throat

8642 Upper Chamber8643 Lower Chamber

8610 Stainless Steel Filter Support Disc

8645 Clamp Assembly8909 Flow Meter Adapter8910 Vacuum Pump Adapter

8644 Spare Set of Glassware (complete)

Apparatus for APSD testing of FP/Salmeterol Products

8372 ACI Inlet Cone for FP/Salmeterol Aerosols*8405 ACI Preseparator for FP/Salmeterol Powders*

8406 Set of 2 O-rings for FP/Salmeterol ACI Preseparator (Spare)

8505 FP/Salmeterol Induction Port*

8505SW FP/Salmeterol Induction Port (One-piece 316 Stainless Steel)

8920 FRS Flow Meter Adapter for GTI/FP Ind Port
 8506 Flow Meter Adapter for FP/S Induction Port
 5401A FP/Salmeterol ACI Carrying/Wash Rack

Other

8503 Set of 2 Silicone Rubber Rinsing Caps for FP Induction Port

^{*} Please specify Aluminium (A) or 316 Stainless Steel (S) when placing your order.

Albuterol Inhalation Aerosols

The draft monograph for Albuterol Inhalation Aerosols (Albuterol Inhalation Aerosol In-Process Revision 44(1)) specified a special glass Sample Collection Apparatus to be used for DDU testing (see below).




The original apparatus used a solid plastic firing adapter, instead of a mouthpiece adapter, to accept an inhaler with a circular mouthpiece of corresponding dimensions. Alternatively, a silicone Mouthpiece Adapter (page 211) could be used.





APSD measurement was conducted using a standard ACI equipped with a specially modified induction port. A special Inlet Sleeve is available that slips over the induction port inlet, to enable the induction port to be used with regular mouthpiece adapters used on USP/NGI induction ports.

Delivered Dose Uniformity



Firing Adapter

Aerodynamic Particle Size Distribution



Albuterol Induction Port

Albuterol Induction Port Inlet Sleeve (optional)

In addition to the above, the following are recommended to complete a fully-operational test set-up for the DDU testing and APSD measurement of **Albuterol Inhalation Aerosols**.



Apparatus for DDU testing of Albuterol Aerosol Products

Cat. No. Description

8520 Sample Collection Apparatus for Albuterol Aerosol

8524 Glass Wool (1 m length)

8521 Firing Adapter8522 Flow Meter Adapter

Spare Parts for Sample Collection Apparatus for Albuterol Aerosol

8523 Glassware for Albuterol Aerosol Sample Collection Apparatus

Apparatus for APSD testing of Albuterol Aerosol Products

8509 Albuterol Induction Port*

8509SW Albuterol Induction Port (One-piece stainless steel)

8519 Albuterol Induction Port Inlet Sleeve*8920 FRS Flow Meter Adapter for GTI/FP Ind Port

5238 DFM Flow Meter Adapter

^{*} Please specify Aluminium (A) or 316 Stainless Steel (S) when placing your order.

Special Applications Device Robustness/ Inhaler Misuse

Patient Exhalation Simulator PES

It is estimated that between 14-22% of patients exhale into their DPI mouthpiece prior to the inhalation step*. The consequence of this poor technique may be insufficient drug delivery for effective administration and ultimately, inadequately controlled respiratory disease and/or an overreliance on emergency medication.

The **Patient Exhalation Simulator** (PES) accurately replicates the effects of a patient exhaling into the device mouthpiece prior to the inhalation step. The warm, humid air generated by the PES can be set at flow rates representative of different human exhalation profiles.

The PES enables developers to assess how device misuse impacts the critical quality attributes of the inhaler, empowering device design optimisation to ensure robust drug delivery.



Simple to set-up and easy-to-use



Adjustable air flow temperature and flow rate



Ideal for assessing a range of patient profiles



Low maintenance



Works with existing Copley mouthpiece adapters



Qualification tools are available



Key Features:








Air flow **temperature** and **flow rate** are adjustable via a digital display

Dimension (w x d x h)





PES Technical Specifications

Temperature set-points	31°C, 34°C, 37°C
Relative humidity	Always saturated
Flow rate range	10 - 60 L/min
Water reservoir capacity	850 mL
PES Performance Tested at set-points: • 34°C, 60 L/min Under ambient conditions: • 22°C, 45% RH	Temperature: 37°C ± 1.5°C Relative humidity: 85% RH ± 5% RH

225 x 300 x 1030 mm

Qualification & Maintenance

- Comprehensive IQ/OQ documentation package available
- Extended warranty available

Patient Exhalation Simulator PES

Cat. No.	Description
9120	Patient Exhalation Simulator - Model PES
9126	Qualification Tools for Patient Exhalation Simulator
9130	IQ/OQ Documentation for Patient Exhalation
	Simulator PES
1076	PES Extended Warranty - 1 year
1077	PES Extended Warranty - 2 years

Automation

Delivering up to a four-fold increase in throughput, automation reduces manual handling and operator input, delivering enhanced reproducibility, lowering the risk of repetitive strain injury (RSI) and reducing overall testing costs.

We supply a broad range of automation solutions supporting both sampling and recovery for delivered dose uniformity (DDU) testing and aerodynamic particle size distribution (APSD) measurement. Our off-the-shelf solutions streamline validation and product testing methods and boost test accuracy and productivity in both R&D and QC.



Improve efficiency



Reduce variability



Eliminate handling errors



Increase testing capacity



Automated Shake, Fire & Flow Control for MDIs, Nasal Sprays and Nasal Aerosols

Vertus® III Series

Vertus III and Vertus III+ are fully automated benchtop shake and fire systems for precise, controlled and reproducible MDI, nasal spray and nasal aerosol testing.

Suitable for:







See page 284.



Automated 10-Way Shake and Fire to Waste for MDIs

DecaVertus® III

A high-throughput 10-way shake and fire to waste system for highly reproducible and controlled MDI testing.

Suitable for:



See page 290.



Automated Drug Recovery for DDU Testing

DUSA Shaker™ DTS 100i

For full, fast and repeatable drug recovery from both MDI and DPI DUSA collection tubes.













See page 294.

Automated Cascade Impactor Preparation

Impactor Coater™ IC 200i

Standardises impaction surface coating for both NGI Collection Cups and ACI Collection Plates.







See page 296.

Automated Drug Recovery for APSD Measurement

Gentle Rocker™ GR 200i

Promotes easy and fully repeatable dissolution of active drug present on NGI Collection Cup and ACI Collection Plate surfaces prior to analysis.

Recommended for:







See page 298.

Impactor Genie™ IG 200i An innovative 2-in-1 solution

Combining the coating capabilities of the IC 200i with the drug recovery features of the GR 200i.







See page 300.








Automation

Automated Shake, Fire and Flow Control for MDIs, Nasal Sprays and Aerosols

Due to the nature of metered spray pump technology and propellant-based aerosols, the testing of MDIs, nasal sprays and nasal aerosols is inherently susceptible to variability from a number of different sources.



Identifying issues within the test method and limiting variability between analysts can be challenging, but inadequate control may lead to erroneous data and consequently substantial costs to the company.

Automated shake and fire systems enhance the sensitivity of OINDP testing and, more broadly, boost data integrity by eliminating firing errors, controlling air flow speed and automating leak testing. Such systems

enable precise, controlled, reproducible testing while at the same time boosting productivity. Vertus® III and DecaVertus® III range offers extensive parameter control and monitoring, allowing:

- Precise and easy method validation
- Streamlined routine testing
- Cause of variation identification
- Enhanced data integrity and accuracy

Choose your Automated Shake & Fire System

	Vertus III	Vertus III+	DecaVertus III					
Max. Number of Devices Supported per Run	1	1	10					
Fire to Sample	✓	✓	Х					
Fire to Waste	✓	✓	✓					
Sample Weighing	х	✓	х					
	Devices supported							
MDIs	✓	✓	✓					
Nasal Sprays	✓	✓	Х					
Nasal Aerosols	✓	✓	✓ (canister only)					

Vertus® III Range

Offering precisely controlled and repeatable delivery for Delivered Dose Uniformity (DDU) testing and Aerodynamic Particle Size Distribution (APSD) measurement, the Vertus III and Vertus III+ automate all aspects of MDI, nasal spray and nasal aerosol dose testing.

Compatible with over 40 different collection device combinations including DUSA, Next Generation Impactor (NGI), Alberta Idealised Throat and Nasal Inlets, and Spray Force Tester, the Vertus III range offers total control over the test technique, but the flexibility to apply any industry-standard shake and fire test method.

The Vertus III range offers complete control over all test parameters, including:

Shaking profile

- Speed
- Angle
- Duration

Time between shake and fire

Firing profile

- Force
- Rise time
- Hold time
- · Release time

Air flow through the system

As the Vertus III range is fully compatible with DecaVertus® III, methods can be easily transferred between systems, with DecaVertus III used to alleviate the burden of through-life testing.





Ph. Eur., EMA, USP, FDA, ChP and NMPA compliant



21 CFR Part 11 compliant



Precise control over all test parameters



Compatible with all standard collection devices



Suitable for a wide range of MDIs, nasal sprays and nasal aerosols



Integrated air flow



Ideal for both DDU and APSD



Suitable for both R&D and QC applications



In situ impactor leak testing capability



Extensive reporting options




Stores and recalls methods



Broad shake and fire parameters accommodate a wide scope of methods

Key Features:



Shot Weight Measurement

The Vertus III+ has the additional capability of measuring shot weight (the weight of the dose released during a single actuation) via an integrated analytical balance. Useful for assessing the consistency of drug release from the device, shot weight is an efficient way to detect misfiring and more broadly, for analytical troubleshooting.

Ph. Eur. (monograph 0676) requires the monitoring of Uniformity of Delivered Mass by collecting shot weights of nasal sprays through life. Automation significantly streamlines the testing process when shot weight information is required.





Exhaust Port

The exhaust port supports the efficient extraction of flammable propellants or high potency drugs where additional safety measures are required.



Interface Plates

The Vertus III range is compatible with collection devices for all compendial testing, plus other standard tests for MDIs, nasal sprays and nasal aerosols.

Interface Plates for MDIs



Vertus III shown here with **DUSA Stack** and **Priming & Waste Module**

Priming & Waste Module

The new Priming & Waste Module integrates firing-towaste into automated test methods, enabling compendial entire contents testing with minimal manual input.

Each interface plate can be placed directly on top of the Priming & Waste Module. Vertus III and Vertus III+ can switch automatically between priming and test levels, firing-to-waste or to dose collection as required, without operator intervention. This enables highly efficient testing procedures, most notably to meet through-life test requirements for DDU and APSD. Additionally, the Priming & Waste Module can be used as a standalone interface for waste shot collection.

Compatible Test Interfaces



Priming & Waste Module only



Next Generation Impactor NGI with Induction Port



Andersen Cascade Impactor ACI with Induction Port



Next Generation Impactor NGI with Adult Alberta Idealised Throat



Andersen Cascade Impactor ACI with Child Alberta Idealised Throat



Fast Screening Andersen FSA with Induction Port



Fast Screening Impactor FSI with Adult Alberta Idealised Throat



Glass Twin Impinger GTI



Thin Layer Chromatography (TLC) Plate



Fast Screening Impactor FSI with Induction Port




Spray Force Tester SFT



Plume Temperature Tester PTT

Interface Plates for Nasal Sprays & Nasal Aerosols



Vertus III shown here with DUSA Interface Plate

For DDU testing of nasal sprays and nasal aerosols, USP <601> recommends a 'mechanical actuation procedure' to control actuation force, speed, stroke length and for units to be 'thoroughly shaken' prior to firing the dose.

In addition to this, the 2003 FDA guidance on Bioavailability and Bioequivalence recommends automated actuation systems for BE assessments to decrease variability in drug delivery.

Compatible Test Interfaces



Nasal Spray Dose Collector NSDC



Nasal Spray Waste Collector NSWC



Next Generation Impactor NGI with 2L Glass Expansion Chamber



Andersen Cascade Impactor ACI with Alberta Idealised Nasal Inlet AINI




Next Generation Impactor NGI with Alberta Idealised Nasal Inlet AINI



Andersen Cascade Impactor ACI with 2L Glass Expansion Chamber



Fast Screening Andersen FSA with Alberta Idealised Nasal Inlet AINI



Fast Screening Impactor FSI with 2L Glass Expansion Chamber



Fast Screening Andersen FSA with 2L Glass Expansion Chamber



Fast Screening Impactor FSI with Alberta Idealised Nasal Inlet AINI



Glass Twin Impinger with Nasal Adapter



Thin Layer Chromatography (TLC) Plate

Technical Specifications: Vertus III & Vertus III+

Shaking parameter control in	cludes:							
	Shaking parameter control includes:							
Shake starting angle	✓	Shake speed	✓					
Shake angle	1	Shake duration	/					
Firing parameter control includes:								
Fire force	1	Force release time	/					
Fire rise time	/	Firing angle	1					
Air flow parameter control includes:								
Air flow rate	1	Air flow measurement	1					
'								
Shot weight measurement (Vertus III+ only)								
Weight range:		0.01 mg to 200 g						
		Resolution: 0.01 mg						
Device compatibility								
MDIs:	1	Nasal aerosols:	1					
Nasal sprays:	/							
User interface:								
10.1" colour touchscreen								
Dimensions (w x d x h):								

Qualification & Maintenance

- Comprehensive IQ/OQ documentation
- · Qualification kit available
- Extended warranty available
- Remote support and field servicing available

Vertus II & Vertus Plus: Reporting

Extensive data output options are available as standard:







Reported Parameters:

 \cdot Run report \cdot Audit report \cdot Method report

- Connectivity:
- Ethernet x 4
- USB x 3
- Run Out digital output
- Run In digital input
- RS-232
- Balance for shot weight collection
- Temperature & Relative Humidity probe
- Label Printer

Vertus III / Vertus III+ Shake and Fire System

Cat. No.	Description
9770	Vertus III Shake and Fire System
9790	Vertus III+ Shake and Fire System
9772	Anti-Static option for Vertus III
1078	Vertus III Extended Warranty - 1 year
1079	Vertus III Extended Warranty - 2 years
1080	Vertus III+ Extended Warranty - 1 year
1081	Vertus III+ Extended Warranty - 2 years

Interface plates

9777	DUSA (x4) Interface Plate for MDIs for Vertus III
9749	DUSA (x1) Interface Plate for Nasal Products for
	Vertus II and III
9775	NGI Interface Plate for Vertus III
9715	GTI Interface Plate for Vertus II and III
9744	FSI Interface Plate for Vertus II and III
9776	ACI/FSA Interface Plate for Vertus III
9784	Spray Force Tester Interface Plate for Vertus III
9778	Gripper for USP & Nasal Induction Port for Vertus III
9779	Gripper for FP Induction Port for Vertus III
9781	Gripper for Alberta Idealised Nasal Inlet (AINI)
	for Vertus III
9782	Gripper for Adult Alberta Idealised Throat (AIT)
	for Vertus III
9783	Gripper for Child Alberta Idealised Throat (AIT)
	for Vertus III
9785	TLC Plate Interface Plate for MDIs, Small for Vertus III
9786	TLC Plate Interface Plate for MDIs, Large for Vertus III
9787	TLC Plate Interface Plate for Nasal Sprays, Small
	for Vertus III
9788	TLC Plate Interface Plate for Nasal Sprays, Large
	for Vertus III

Acessories for MDI

9705	MDI Holder (per inhaler design)
9901	Mouthpiece Adapter Mould (per inhaler/inlet design)
9902	Mouthpiece Adapter for ACI/NGI Induction Port and DUSA
9903	Mouthpiece Adapter for Other Inlets (each)

Accessories for Nasal

9735	Nasal Spray Dose Collector (NSDC) for Vertus II and III
9736	Nasal Spray Waste Collector (NSWC) for Vertus II and III
9738	Nasal Spray Holder for use with NSDC and NSWC
9746	Nasal Spray Holder for use with Expansion Chamber
9747	Nasal Spray Holder for use with AINI & DUSA
9748	Nasal Spray Holder for use with GTI
8544	Nasal Device Nosepiece Adapter for AINI
8545	Tooling Charge (per nasal device)
9781	Gripper for Alberta Idealised Nasal Inlet (AINI) for Vertus III
9910	Nosepiece Adapter for ACI/NGI Induction Port and DUSA
9901A	Nosepiece Adapter Mould (per nasal spray design)

General Accessories

Cat. No.	Description
9773	Temperature and Humidity Sensor for Vertus III
9765	Label Printer
9798	Vertus III, Vertus III+, DecaVertus III Qualification Kit
9799	Re-calibration of Vertus III/DecaVertus III
	Qualification Tools
9774	IQ/OQ Documentation for Vertus III/Vertus III+
9714	Air compressor
8791	Foot Switch - compatible with TPK/BAC/BRS/Vertus/
	DecaVertus

Maintenance and Support

1006	Remote Access Support for Vertus/DecaVertus (1yr/10hrs)
9721	Remote Diagnostic Gateway
9724	Direct Connection Setup - Remote Support

Spares

9712 9716	Spare Filter Cartridge for Waste Shot Collector Direct Thermal Printer Labels (12 Rolls of 475
	each) - spares
9719	Thermal Transfer Printer Labels (12 Rolls of 475
	each) - optional
9725	Thermal Transfer Printer Ribbon (6 Cartridges)
	- optional
9792	TLC Pre-Coated Plates Size 5 x 10cm (50 Plates)
9793	TLC Pre-Coated Plates Size 5 x 20cm (100 Plates)
9794	TLC Pre-Coated Plates Size 10 x 20cm (50 Plates)
9795	TLC Pre-Coated Plates Size 20 x 20cm (25 Plates)
9796	Inline Filter for Vacuum Inlet



DecaVertus® III

Automating firing-to-waste for through life testing of up to ten MDIs per test run, the DecaVertus is a high-throughput system for reproducible, controlled testing.

Highly advantageous from the perspective of enhancing test repeatability, conserving analyst time and eliminating the risk of repetitive strain injury (RSI), the DecaVertus ensures firing-to-waste occurs under closely controlled conditions, eliminating potential sources of variability from testing.

DecaVertus III offers complete control over all test parameters, including:

Shaking profile

- Speed
- Angle
- Duration

Time between shake and fire

Firing profile

- Force
- Rise time
- Hold time
- · Release time

Air flow through the system

As DecaVertus III is fully compatible with the Vertus® III range, methods can be easily transferred between systems, enabling the same parameters to be used for dose collection on Vertus III and through life firing-to-waste on DecaVertus.





Ph. Eur., EMA, USP, FDA, ChP and NMPA compliant



21 CFR Part 11 compliant



Precise control over all test parameters



Suitable for a wide range of MDI devices



Improves reproducibility and frees up analyst time



Independent air flow control per channel



Stores and recalls methods




Extensive reporting options



Broad shake and fire parameters accommodate a wide scope of methods

Key Features:



Independent air flow control for each channel: each inhaler has its own dedicated air flow channel to minimise clogging and enable high-volume products to be tested in a single run, reducing the need to stop and clean the filters.



Technical Specifications: DecaVertus III

Shaking parameter contro	l includes:		
Shake starting angle	✓	Shake speed	✓
Shake angle	✓	Shake duration	✓

Firing parameter contro	ol includes:		
Fire force	✓	Force release time	1
Fire rise time	✓	Firing angle	✓

Fire rise time	✓	Firing angle	✓
Air flow parameter con	trol includes:		
Air flow before firing	1	Air flow after firing	✓
Device compatibility			
MDIs:	✓	Can-only:	1

User interface:

10.1" colour touchscreen

Dimensions (w x d x h):

1130 x 630 x 370 mm

Connectivity:

- Ethernet x 2
- Run In digital input
- USB x 3
- RS-232
- Run Out digital output
- Label Printer

Qualification & Maintenance

- Comprehensive IQ/OQ documentation
- · Qualification kit available
- Extended warranty available
- · Remote support and field servicing available

DecaVertus III: Reporting

Extensive data output options are available as standard:







Reported Parameters:

• Run report • Audit report • Method report

DecaVertus III Waste Shot Collection System for 10 MDIs

Cat. No.	Description
9870	DecaVertus III Shake and Fire to Waste System
1082	DecaVertus III Extended Warranty - 1 year
1083	DecaVertus III Extended Warranty - 2 years 701

Accessories (MDIs only)

9805	Carriage for MDI (per inhaler design)
9808	Can Only Carriage for DecaVertus (per can design)
9798	Vertus III, Vertus III+, DecaVertus III Qualification Kit
9799	Re-calibration of Vertus III/DecaVertus III
	Qualification Tools
9871	IQ/OQ Documentation for DecaVertus III
9765	Label Printer
8791	Foot Switch - compatible with TPK/BAC/BRS/Vertus/DecaVertus

Spare parts

9716	Direct Thermal Printer Labels (12 Rolls of 475 each)
9719	Thermal Transfer Printer Labels (12 Rolls of 475 each)
9725	Thermal Transfer Printer Ribbon (6 Cartridges)
9820	Pack of 10 Spare Waste Filter Cartridges
9821	Pack of 100 O-rings for Actuator Pins
9850	Replacement Set of 10 Actuator Pins for DecaVertus Carriage

Maintenance and Support





DUSA Shaker™ DTS 100i

Manual drug recovery from the Dose Uniformity Sampling Apparatus (DUSA) for dose uniformity testing can be time-consuming, prone to variability and may lead to repetitive strain injury (RSI).

The DUSA Shaker™ DTS 100i provides full, fast and repeatable drug recovery from all internal MDI and DPI DUSA collection tube surfaces, for quicker, more efficient regulatory and compendial delivered dose uniformity (DDU) testing.

Eliminating a time-consuming and highly variable manual drug recovery process, the DTS 100i releases analysts for higher value work whilst reducing the risk of exposure to RSI.

For those new to automation, the DTS 100i is a perfect low-cost, first step towards reduced test variability, fewer out of specification results, greater productivity and safer working practice. For established laboratories, the DTS 100i is a cost-efficient, modular solution that will slot easily into existing set-ups and workflows.



Boosts throughput by up to 10 times



Achieves complete drug recovery via multidirectional mixing



Improves data integrity



Accepts both MDI and DPI DUSA collection tubes




3 rinsing actions: Shake, Roll, or Shake and Roll



Enhances productivity

Key Features:



When assessing DDU of DPIs, DUSA collection tubes without a P1 port must be used to enable rotation (Cat No. 8608A, Collection Tube without P1 Port).



DTS 100i Technical Specifications

DUSA Collection Tube Capacity	DUSA for MDIs x 10 DUSA for DPIs x 10	
Shake Speed:	20 - 200 RPM	
Roller Rotational Speed:	Fixed at 11.4 RPM	
Timer Control:	Up to 100 hours	
Connectivity:	USB A, USB B, RS-232	
Dimensions (w x d x h)	410 x 626 x 227 mm	

Qualification & Maintenance

- Comprehensive IQ/OQ documentation packages and toolkits available
- · Qualification Kit available
- · Extended Warranty available

DUSA Shaker™ DTS 100i

Cat. No.	Description
8630	DUSA Shaker DTS 100i (without collection tubes)
8621	IQ/OQ Documentation for DUSA Shaker
8623	Qualification Tools for DUSA Shaker
8624	Re-calibration of DUSA Shaker Qualification Tools
1032	DUSA Shaker Extended Warranty - 1 year
1033	DUSA Shaker Extended Warranty - 2 years



IC 200i with NGI Collection Cup Tray

IC 200i with ACI Collection Plate Tray & Cups ACI Collection Plate Tray & Cups purchased separately. (Cat. No. 5933)

Impactor Coater™ IC 200i

To prevent particle bounce and subsequent re-entrainment within the flowing airstream during aerodynamic particle size distribution (APSD) sampling, regulators recommend the coating of each impactor stage collection surface. However, the manual coating of each collection surface is prone to variability and is labour-intensive.

The Impactor Coater IC 200i reproducibly applies surface coatings to both the NGI Collection Cups and the ACI Collection Plates, eliminating the problem of particle bounce and re-entrainment when using cascade impactors to measure the APSD of OINDPs.

Standardising the application of surface coating to each collection surface, the IC 200i removes the inherent variability associated with the coating process, while boosting laboratory productivity and throughput.



Coats surfaces in as little as 2 minutes



Frees up analysts for other tasks



Enables easy method transfer between sites



Minimises coating solution wastage



A note about impaction surface coating

For OINDPs where the particles are hard and dry, (e.g. Dry Powder Inhalers (DPIs)) or where only a few actuations are delivered to the impactor, such as is the case for Metered Dose Inhaler (MDIs), particle bounce and re-entrainment in the flowing airstream can bias the measured size data to finer sizes.

It is therefore important to assess the potential impact of these phenomena on downstream stages at an early point in development so that corrective action can be taken.

Coating the impaction surfaces with a tacky, viscous material such as glycerol or silicone oil is recommended by the regulators to address this problem. If a surface coating is required, the amount, its uniformity and the method in which it is applied and its potential to affect drug recovery should be assessed during method development.

Key Features:





IC 200i Technical Specifications

Flow Rate	0 - 100%		
Dispense and Reverse Cycle Time	0 - 10 minutes		
Connectivity:	USB A and USB B		
Dimensions (w x d x h)	590 x 320 x 250 mm [IC 200i] 150 x 220 x 130 mm [Pump]		

Qualification & Maintenance

- Comprehensive IQ/OQ/PQ documentation package and toolkit available
- Extended warranty available

Impactor Coater™ IC 200i		Accesso	Accessories		
Cat. No.	Description	Cat. No.	Description		
5940	Impactor Coater Model IC 200i	5943	IQ/OQ Documentation for Impactor Coater ICi Series		
5942	Cover and Tubing Set for NGI Cup Tray & Cups	5926	Qualification Tools for GR, IC, IGi Series		
5941	Cover and Tubing Set for ACI Collection Plate Tray & Cups	5927	Re-calibration of Qualification Tools for GR, IC, IGi Series		
5933	ACI Collection Plate Tray & Cups	1072	IC 200i Extended Warranty - 1 year		
		1073	IC 200i Extended Warranty - 2 years		
Spares		8120	Inhaler Testing Workstation - BasePlate and Upright		
-		8140	ITW Cover Stand Attachment		
5947	Spare Set of 8 Pump Tubing Cassettes	5224	Storage Cabinet for Impactor Collection Trays		
5901	500 mL Solvent Reservoir complete with 8-way Cap	8766	Printer		
5902	1000 mL Solvent Reservoir complete with 8-way Cap				



GR 200i with NGI Collection Cup Tray

GR 200i with ACI Collection Plate Tray Plate & Cups ACI Collection Plate Tray & Cups purchased separately. (Cat. No. 5933)

Gentle Rocker™ GR 200i

Drug recovery from each collection surface is a critical but time-consuming element of aerodynamic particle size distribution (APSD) sampling with a cascade impactor. It involves dispensing a defined aliquot of solvent on to each surface, followed by repeated agitation to ensure complete drug dissolution.

The Gentle Rocker GR 200i promotes easy and fully repeatable dissolution of the active drug present in both the NGI Collection Cups and on the ACI Collection Plate surfaces following sampling.

Gently agitating solvent back and forth across the impaction surface to aid assay sample preparation, the GR 200i enables specific, reproducible drug recovery and easy method transfer, delivering reliable results and a lighter analytical workload.



Quick and easy drug recovery



Suitable for a broad range of drug recovery methods



Frees up analysts for other tasks



Enables easy method transfer between sites



Optional: Low Evaporation Cover with seals to minimise solvent loss where evaporation is a particular problem

Key Features:





GR 200i Technical Specifications

Agitation speed	10 - 60 RPM (± 1 RPM)		
Run Time	Up to 100 hours or Up to 60,000 revolutions		
Connectivity:	USB A, USB B, RS-232		
Dimensions (w x d x h)	590 x 320 x 235 mm		

Qualification & Maintenance

- Comprehensive IQ/OQ/PQ documentation package and toolkit available
- Extended warranty available

Gentle Rocker™ GR 200i

Gentle Rocker™ GR 200i	Accessories		
Cat. No. Description	Cat. No. Description		
5932 Gentle Rocker Model GR 200i	5925 IQ/OQ Documentation for Gentle Rocker GRi Series		
5933 ACI Collection Plate Tray & Cups	5926 Qualification Tools for GR, IC, IGi Series		
	5927 Re-calibration of Qualification Tools for GR, IC, IGi Series		
	1070 GR 200i Extended Warranty - 1 year		
	1071 GR 200i Extended Warranty - 2 years		
	5934 Low Evaporation Cover for ACI Collection Plate Tray & Cups		
	5935 Low Evaporation Cover for NGI Cup Tray & Cups		
	8120 Inhaler Testing Workstation - Baseplate and Upright		
	8140 ITW Cover Stand Attachment		
	5224 Storage Cabinet for Impactor Collection Trays		
	8766 Printer		



Impactor Genie™ IG 200i

Combining the impaction surface coating power of the Impactor Coater™ IC 200i with the drug recovery capabilities of the Gentle Rocker™ GR 200i, the Impactor Genie IG 200i is the ultimate 2-in-1 solution for quicker, more efficient, regulatory and compendial aerodynamic particle size distribution (APSD) impactor preparation and drug recovery.

Compatible with both NGI Collection Cups and ACI Collection Plates, the IG 200i transforms both pre- and post-sampling processing for busy analysts.

Offering double the automation power, the IG 200i enhances sampling repeatability and accuracy, while boosting analytical productivity and enabling easy method transfer between sites. Transform between Impactor Coater and Gentle Rocker in 2 easy steps!

Mode: Impactor Coater IC 200i





Uniform coating in as little as 2 minutes





Innovative tray tilting function eases sample collection post-drug recovery





Samples protected during process by dust cover



Adjustable agitation speed range

ACI Collection Plate Tray & Cups

Required to enable compatibility with the collection plates of the Andersen Cascade Impactor (ACI), the **ACI Collection Plate Tray & Cups** accommodates all 8 stages, plus filter.



Impactor Genie™ IG 200i: Accessories

Cover Stand

Included as standard with purchase of the IG 200i, the **Cover Stand** is ideal for storing the Dust Cover, Impactor Coater Cover and/or Low Evaporation Cover when not in use. Additional attachments holders can be purchased separately.




Low Evaporation Cover

Fitted with seals and retaining clips, the **Low Evaporation Cover** minimises solvent loss where risk of evaporation is a particular problem.



Storage Cabinet for Impactor Collection Trays

Providing storage for up to six NGI Collection Cup Trays and/or ACI Collection Plate Trays, the space-saving **Storage Cabinet** helps keep benchtop spaces tidy and organised (trays not included).



Qualification & Maintenance

- Comprehensive IQ/OQ/PQ documentation package and toolkit available
- · Extended warranty available

Impactor Genie IG 200i		Accessories		
Cat. No.	Description	Cat. No.	Description	
5945	Impactor Genie Model IG 200i	5946	IQ/OQ Documentation for Impactor Genie IGi Series	
5933	ACI Collection Plate Tray & Cups	5926	Qualification Tools for GR, IC, IGi Series	
5941	Cover and Tubing Set for ACI Collection Plate Tray & Cups	5927	Re-calibration of Qualification Tools for GR, IC, IGi Series	
5942	Cover and Tubing Set for NGI Cup Tray & Cups	1074	IG 200i Extended Warranty - 1 year	
		1075	IG 200i Extended Warranty - 2 years	
Spares		5934	Low Evaporation Cover for ACI Collection Plate Tray & Cu	
•		5935	Low Evaporation Cover for NGI Cup Tray & Cups	
5947	Spare Set of 8 Pump Tubing Cassettes	5224	Storage Cabinet for Impactor Collection Trays	
8140	ITW Cover Stand Attachment	8766	Printer	
5901	500 mL Solvent Reservoir complete with 8-way Cap			
5902	1000 mL Solvent Reservoir complete with 8-way Cap			



Sample Preparation Unit SPU 200i

Ensuring full, reproducible drug recovery from the NGI, ACI and FP/Salmeterol Induction Ports and the NGI Preseparator, the Sample Preparation Unit SPU 200i automates repetitive drug recovery procedures, alleviating testing bottlenecks and reducing the unwanted effects of repetitive strain injury (RSI).



Easy to use touchscreen interface



Reproducible sample preparation



Variable speed control for different dissolution applications



Ideal for use with Induction Ports and/or Preseparators

Key Features:







SPU 200i fitted with 2 x ACI Induction Ports



Fixture wth ACI/Albuterol Induction Port



Fixture with NGI Induction Port



Fixture with FP Induction Port

SPU 200i: User Interface



Setting a test parameter



Set v Actual test parameters (before test run)



Set v Actual test parameters (during test run)



Settings menu



Report output settings menu

SPU 200i: Technical Specifications

Speed:	Variable (20 and 60 rpm (+/- 1 rpm))	
Rinsing Cycle Duration:	0 - 120,000 revolutions or 100 hours	
Rotational Direction	Fixtures reverse rotation direction half way through run	
Connectivity:	RS-232 USB A USB B	
Dimensions (w x d x h):	285 x 335 x 295 (with a single Induction Port Fixture) 420 x 335 x 310 (with a single Preseparator Fixture)	

Compliance and Maintenance

- Comprehensive IQ/OQ documentation packages and toolkits available
- · Qualification Kit available
- Extended Warranty available

Sample Preparation Unit 200i

Cat. No.	Description
9222	Sample Preparation Unit Model SPU 200i (without Fixtures)
1038	SPU 2000 Extended Warranty - 1 year
1039	SPU 2000 Extended Warranty - 2 years

Accessories

Cat. No.	Description
9226	Fixture for ACI/NGI/Albuterol & FP Induction Port (each)
8503	Set of 2 Silicone Rubber Rinsing Caps for FP Induction Port
8504	Set of 2 Silicone Rubber Rinsing Caps for ACI/Albuterol
	Induction Port
9227	Fixture for NGI Preseparator (each)
5265	Set of 2 Silicone Rubber Rinsing Caps for NGI Induction Port
5266	Set of 2 Silicone Rubber Rinsing Caps for NGI Preseparator
9223	IQ/OQ Documentation for SPU 200i
9213	SPU 200i Qualification Tools
9214	Re-calibration of SPU 200i Qualification Tools
9765	Label Printer



Impactor Cleaning System

Ensuring the thorough, reproducible and controlled cleaning and drying of cascade impactors, the Impactor Cleaning System has been designed to clean component parts of both the NGI and ACI. Regular cleaning and drying are an essential element of good impactor practice. They ensure that the instrument is free of debris prior to testing and that it remains in optimum condition throughout its life.



Available as a complete system, or as individual components



Consistent, reproducible cleaning



Benchtop system



Suitable for both NGI and ACI cleaning

Clean your impactor in 4 easy steps:



1. Cleaning

Impactor Ultrasonic Cleaning Bath



2. Rinsing

Impactor Rinse Bath



3. Aspiration

Impactor Suction Aspiration



4. Drying

Impactor Drying Oven



Step 1. Ultrasonic Cleaning Bath

Using ultrasound (usually from 15-400 kHz) to promote the effective cleaning of nozzles and other difficult-to-access places, the Impactor Ultrasonic Cleaning Bath is able to efficiently remove sticky, adhering and embedded particles from solid surfaces.

Step 2. Impactor Rinse Bath

Following cleaning, the impactor parts are normally rinsed in clean cold water and left to drain.

Step 3. Impactor Suction Aspirator

Used to remove the small amounts of excess water that collect in the bottom of the impactor stages and preseparator parts following rinsing and prior to drying, the Impactor Suction Aspirator comprises a hand-held probe linked via a water collection jar to a vacuum pump, which provides the necessary suction.

Step 4. Impactor Drying Oven

Following sonication, rinsing and aspiration, the impactor parts should be dried using a heated cabinet. The Impactor Drying Oven has a temperature range of 25 - 70 +/-1 degrees C, ideal for impactor part drying. Designed to accept 3 individual carrying racks, the unit is fitted with an inner glass inspection door together with a wipe-clean, all stainless-steel interior for ease-of use and cleaning.

The 4-speed forced air circulation means that the oven reacts rapidly to change and is ideally suited to impactor drying, where maximum accuracy and warm-up are required and the door is to be opened on a frequent basis.

Impactor Cleaning System Accessories

Carrying/Wash Racks

The impactor parts are normally placed in a rack prior to immersion (a) to segregate them during the cleaning process and (b) to maximise the surface area exposed to the cleaning process. The Impactor Carrying/Wash Racks are constructed from heavy duty polypropylene and fitted with neoprene cushions to prevent scratching to the outer surfaces of the parts.



NGI Rack

The NGI rack has 12 apertures corresponding to the 8 Collection Cups, NGI Induction Port and the three parts of the NGI Preseparator.

ACI Rack

The ACI Rack has 21 apertures corresponding to the 8 stages, the 8 Collection Plates, the Inlet Cone, Induction Port and the 2 parts of the Preseparator of the ACI.





Storage Cabinet for Impactor Collection Trays

Accommodates up to six NGI Collection Cup Trays / ACI Collection Plate Trays (NGI Collection Cup Trays / ACI Collection Plate Trays not included).

FP/Salmeterol ACI Rack

Available to accommodate the special Induction Port and Preseparator used.

Each rack measures 420 mm (w) x 230 mm (d) and is designed to fit inside the basket used in the Impactor Ultrasonic Cleaning Bath. The basket prevents the carrying rack from touching the bottom or sides of the bath.

Impactor Cleaning System

Cat. No. Description

5400 Impactor Cleaning System (excluding Carrying/Wash Rack)

5205 NGI Carrying/Wash Rack5401 ACI Carrying/Wash Rack

5401A FP/Salmeterol ACI Carrying/Wash Rack

Modules Only

5402 Impactor Ultrasonic Cleaning Bath (including basket and lid)

5403 Impactor Rinse Bath
5404 Impactor Suction Aspirator
5405 Impactor Drying Oven
5406 Stainless Steel Drip Tray



Qualification/ Servicing & Training

Good Manufacturing Practices (GMP) regulations require that:

- A. The test methods used to monitor pharmaceuticals must meet proper standards of accuracy and reliability
- B. Companies should establish procedures to ensure the fitness for use of instruments that generate data supporting product testing

However, these GMP regulations do not provide definitive guidance as to how these aims are to be achieved.

The USP has sought to address this problem by the introduction of a series of chapters as follows:

<1058> Analytical Instrument Qualification

<1225> Validation of Compendial Procedures

<1226> Verification of Compendial Procedures

<1603> Good Cascade Impactor Practices

It is interesting to note that the scientific community has used the terms "validation" and "qualification" on an interchangeable basis thus creating a degree of ambiguity as to their use. For this reason, USP have suggested that:

- A. The term "qualification" be applied to instrumentation
- B. The term "validation" be applied to processes and software

The term "Analytical Instrument Qualification" (AIQ) is used for ensuring that an instrument is suitable for its intended application and the term "Analytical Method Validation (AMV)" is used for ensuring that the analytical and software procedures employed are suitable for their intended application.

The USP Chapter <1058> Analytical Instrument Qualification describes in detail the four phase approach to qualification based on design (DQ), installation (IQ), operational (OQ) and performance (PQ) qualification.

It is important to note that the purpose of AIQ and its counterpart, AMV, is to ensure the quality of analysis before conducting the test, whereas system suitability tests and quality control checks ensure the quality of analytical results immediately before or during sample analysis.

The performance of inhaler testing equipment and the methods associated with them can be influenced by factors other than the equipment itself:

- · Analytical (human error)
- · Instrument (errors in instrument and/or ancillary equipment)

If these sources of error can be eliminated then it is fair to assume that any anomalies in results are a product of the device/formulation combination itself.



Analytical Errors (Human Errors)



Training



Use of automation (see page 278)



Instrumentation-Related Errors



Qualified instrumentation



Validated analytical procedures

Ways to Reduce Errors in Orally Inhaled and Nasal Drug Product (OINDP) Testing

Copley recognises the scientific and regulatory importance of these initiatives. Therefore, we have designed a selection of products, services and documentation to assist you through the OINDP testing journey:



Qualification ServicesSee page 310



Product
Protection Plans
See page 319



Support See page 320



TrainingSee page 321



Qualification Services

Impactor Qualification

Stage and Components Mensuration

Both the Ph. Eur. and USP lay down certain criteria which the cascade impaction system and method selected for the inhaler must fulfil prior to and during use.

The performance, repeatability and reproducibility of a cascade impactor are dependent on a number of factors, the most critical being the nozzle dimensions (and their spatial arrangement) on each stage together with the air flow rate passing through it.

Providing these critical parameters are within the quoted specification, then the impactors concerned can be expected to give comparable results.

The process of measuring the nozzle diameters and other critical dimensions of cascade impactors is called impactor mensuration.

Both the Ph.Eur. and USP recommend the stage mensuration of impactors prior to use and periodically thereafter.

In practice, cascade impactors often corrode and wear with use owing to their repeated exposure to formulations and recovery solvents. This is particularly true of aluminium impactors.

This can lead to full or partial nozzle occlusions and/or deterioration to the condition of the nozzles, causing changes in the impactor aerodynamics and hence particle collection characteristics.

Stage mensuration, is used to ensure that cascade impactors conform to the critical dimensions stated in USP Chapters <601> and <1603> and Ph.Eur. Chapter 2.9.18 and are therefore fit for use.

Stage mensuration replaces the need for repetitive calibration using standardised aerosols.



Copley provides a one-stop, quick turn-around mensuration service for all types of Ph.Eur. and USP specified impactors, including induction ports and preseparators.



Mensuration Certificate



Mensuration of ACI Stages using the Mitutoyo QV404 Vision Inspection System





Stage Mensuration Certificate with Histogram Option

Mensuration certificates are supplied as standard with all new impactors, preseparators and induction ports, detailing how each component conforms to the pharmacopoeial requirements.

As impactors and ancillaries are put into use, regular re-mensurations (at least annually) should be performed to monitor and confirm their "in-use" compliance.

Data Interpretation

Copley adopts Effective Diameter and In-Use Margin as recognised by the European Pharmaceutical Aerosol Group (EPAG) as a means of determining the suitability of cascade impactors for use.


Derived from the area-mean and area-median diameters of multi-nozzle impactor stages, ED is a useful parameter that can be used to monitor "drift" in the D50 of impactor stages (median nozzle diameter).

The In-Use Margin is calculated as the % of USP/Ph.Eur. tolerance that remains, relative to the ED. If the ED is equal to the stage nominal diameter then the In-Use Margin would be 100%. If, however, the ED is equal to the diameter defined by the upper or lower USP/Ph.Eur. tolerance then the In-Use Margin would be 0%. It follows that if the ED falls outside the compendia tolerance then the In-Use Margin would be a negative value.

Successive mensuration reports allow the tracking and monitoring of any deterioration of In-Use Margin, a useful way of investigating how an impactor is wearing with time. This approach allows the likelihood of an out-of-specification (OOS) stage occurring within the next calibration cycle to be predicted, indicating when remedial work will be required.



Effects of improvements in the NGI manufacturing processes relating to Stage 5 of the NGI with serial number. Every nozzle on the NGI has always met pharmacopoeial specifications (heavy black lines). Now though, every NGI has an ED within just half the range of the pharmacopoeial specification (heavy blue lines). These data therefore provide evidence of our commitment to continuous quality improvement.



Impactor Performance Restoration

Following impactor mensuration there are three possible results; ED within specification, ED in excess of an upper limit and ED below the lower limit for the stage:



ED within specification

No restoration is required when mensuration shows ED within specification.



ED in excess of an upper limit

This is a sign that the nozzles have worn, either as a result of corrosion from the solvents used to dissolve the active drug or erosion from the constant passage of particles through the nozzles concerned. In this case the restoration is feasible as it is not practical to reapply metal to impactor nozzles. Replacement of the stage will be required.



ED below the lower limit

The vast majority of impactors tend to drift out of specification because ED decreases below the lower limit for the stage. This can be caused by a build-up of hardened particulates or, more likely, because corrosion produces metal salts that occlude the nozzle. The formation of oxidised impurities at the nozzle exit is a commonly encountered cause of occlusion, particularly for aluminium impactors, which is why materials such as stainless steel and titanium are often also used.

In this case of ED below the lower limit, performance can sometimes be improved or restored.

Rigorous cleaning and ultrasonics (see page 306 for the Impactor Cleaning System) can be used to remove deposits and restore performance.

Stage Pinning can also be attempted as a secondary option. Pushing stainless steel "go" pins with a diameter between the nominal diameter and the lower tolerance limit for the stage through each nozzle can serve to clear accumulated debris.

Stage Replacement is recommended in cases that the restoration of the impactor stage is not achievable via stage pinning.









Pinning various stages of the ACI

Pinning Kit with close-up of Pin

Impactor Mensuration Services

Cat. No.	Description	Cat. No.	Description
8590	Induction Port Mensuration	5290	NGI Stage Mensuration
8390	ACI Stage Mensuration	5291	NGI Preseparator Mensuration
8990	60 L/min Conversion Kit Mensuration	8591	Alberta Idealised Throat Mensuration
5236	90 L/min Conversion Kit Mensuration	8340	FSA Stage Mensuration
8490	ACI Preseparator Mensuration	5270	FSI Insert Mensuration
8311	ACI Stage Mensuration Histogram (per stage)	8917	GTI Mensuration
8890	MSLI Stage Mensuration and Leak Test		

Mensuration 'Returns' Boxes

5292	NGI Seal Body Mensuration 'Returns' Box	5233	ACI or NGI Leak Test Certificat
5297	Replacement NGI Carton	5234	ACI or NGI Delta-P Certificate

Pinning Kits and Services

5430	ACI Pinning Service (per stage)
5431	ACI Pinning Kit
5432	NGI Pinning Service (per stage)
5433	NGI Pinning Kit

Leak Testing

5233	ACI or NGI Leak Test Certificate
5234	ACI or NGI Delta-P Certificate
5251A	Re-calibration of LTK2 Leak Test Kit tools

In-House and On-Site Equipment Servicing and Calibration

Copley offers a comprehensive range of servicing, maintenance and qualification options, tailored to individual customer needs, providing quality maintenance and calibration procedures at competitive prices:

- In-house equipment servicing
- · On-site equipment servicing
- In-house equipment calibration
- On-site equipment calibration
- On-site equipment IQ/OQ

What is included?



Tailored services for your needs



Qualified engineers and technicians trained to a high standard



Choose between:

- Service contract
- One-off offering



Documentation supplied and completed to GxP standards as per regulatory requirements




Single account manager contact to ensure excellent service

We will be pleased to discuss your individual requirements and quote accordingly.

Qualification Tools and Documents

IQ/OQ Documentation





According to USP Chapter <1058>, Analytical Instrument Qualification is "the collection of documented evidence that an instrument performs suitably for its intended purpose".

It is important to note that the stage mensuration process described on previous pages is intended to replace the need for repetitive impactor calibration based on standard aerosols. It ensures that only impactors that conform to specification are used in testing. Whilst mensuration or calibration is an important part of the qualification process, it does not in itself qualify the whole inhaler testing for use.

This is a separate process. The Installation Qualification/ Operation Qualification Documentation (IQ/OQ) Documentation provided by Copley guides the user through this important process and confirms that the system is fully qualified for use.

It includes:

Master Plan

• Defines the aim and scope of the qualification

• Installation Qualification

 Outlines the test plan, the standard operating procedures and test protocols necessary to perform the IQ for the system concerned

Operation Qualification

 Outlines the test plan and the standard operation procedures and test protocols to perform the OQ of the system concerned

Qualification Documents

Cat. No. Description
 8000 IQ/OQ Documentation for Inhaler Testing Systems
 9500 Respiratory Drug Delivery Essential Theory & Practice Book

Individual ancillaries and automation IQ/OQ documentation can be found in the relevant sections

Qualification Tools



Inhaler Testing Qualification Kit ITQK2

Includes all the tools required to perform IQ/OQ Qualification procedures and can also be used for calibration of the Flow Controllers TPK 100i/R and BAC 100i/R.

In addition to the Inhaler Testing Qualification Kit, the following qualification kits are available for the following products:



Breathing Simulator BRSi Series - Page 156



Facemask Apparatus FMA - Page 244



NGI Cooler™ Page 202



EnviroMate™ Page 196



Patient Exhalation
Simulator PES - Page 276



Spray Force Tester SFT 1000 - Page 256



DUSA Shaker™ DTS 100i Qualification Tools - Page 294



Impactor Coater™ IC 200i - Page 296



Gentle Rocker™ GR 100i Page 298



Impactor Genie™ IG 200i Page 300



Sample Preparation Unit SPU 200i - Page 302



Vertus® III Page 284



DecaVertus® III Page 290

Qualification Tools

Cat. No.	Description
5440	Inhaler Testing Qualification Kit Model ITQK2
5445	Re-calibration of ITQK2 Kit tools
5207	NGI Leak Tester



Product Protection Plans

Standard 12 Months Warranty

Copley offers a 12 months supplier's warranty as standard with our entire product range.

Extended Product Protection

For selected items, Copley offers the option to obtain extended product protection for a further period of 12 or 24 months after the standard warranty expires. We have confidence in our excellent product quality but an extended protection plan provides the peace of mind that comes with an added layer of assurance.

Products that extended product protection plans are available for:



Flow Controllers See page 172



Vacuum Pumps See page 188



Breathing Simulators
See page 156



Automation Tools See page 278

Support

Buy with confidence from Copley. When you purchase equipment from us, you not only get outstanding instrumentation but also a complete customer care package which extends from the start of the sales process through to installation, training, after-sales support and beyond. With a global network of experienced and knowledgeable distributors you can rest assured that, wherever you may be, there is support every step of the way.

Design Support

Our design team has many years' experience working closely with the inhaler testing community in helping to develop ideas for solving particular problems.

Whether you have a longstanding problem, or one that has been created by the introduction of a new process, an idea for a new product, or even a bespoke design that you need manufacturing, we would be delighted to hear from you.



Training Services



As a world leader in the provision of equipment for testing OINDPs, Copley offers a range of tailored training packages for both analysts and lab managers of OINDP researchers, developers, manufacturers and testing labs.

Bespoke Customer Training

Utilise standard Copley packages

Routine induction training for new staff

Customise content to specific requirements

Example training topics:

- · *In-vitro* testing methods for OINDPs (MDIs, DPIs, nebulisers, SMIs, nasal products)
- · Improving the clinical relevance of in-vitro test methods
- · IQ/OQ and maintenance of inhaler testing systems

Book your training course.

- ✓ Highly experienced trainers
- Bespoke training programs
- ✓ On-site training available
- ✓ Certification provided

Please contact us to find out more about our range of training packages.

Contact us at: sales@copleyscientific.co.uk or call: +44 (0)115 961 6229

Multi-company Group Training

Inhaler Testing Academy®
Peer-to-Peer learning
Industry Networking

Inhaler Testing Academy



Our intensive group training program offers attendees a grounding in the regulatory requirements, testing fundamentals and the application of core methods, for all types of OINDPs.

Designed for both beginners and experienced analysts, the **Inhaler Testing Academy** training offers attendees a unique learning experience via a combination of engaging lectures and practical demonstrations.



To find out more, view our brochure or contact us at sales@copleyscientific.co.uk

Index

A	
Abbreviated Impactor	
Measurement (AIM)	260-267
ACI	79, 90-95
ACI Carrying/Wash Rack	94, 308
ACI, Cleaning	306-309
ACI Cleaning System	306-309
ACI Collection Plate Rack	93,95
ACI Drying Oven	307, 309
ACI Pinning Kit	315
ACI Pinning Service	315
ACI Rinse Bath	307, 309
ACI Sample Preparation Unit	302-305
ACI Suction Aspirator	307, 309
ACI Ultrasonic Cleaning Bath 30	07, 308, 309
Active Pharmaceutical Ingredient	(API) 18,80
Active Substance Delivery	
Rate (Nebulisers)	52, 53
Actuation Sensor for MDIs	179
Adapters, Mouthpiece	211-213
Adapters, Nosepiece	211, 213
Add-on Devices 7, 32, 4	0-44, 111-116
Aerodynamic Particle Size (APSD)	13, 78-153
AIM - HRT Model	263-266
AIM - QC Model	261-264
AINI	230
Alberta Idealised Nasal Inlet (AINI) 230
Alberta Idealised Throat (AIT)	228-229
Albuterol Aerosols, Testing	30, 274-275
Analytical Instrument Qualification (AIQ)	310, 311
Analytical Method Validation (AM	V) 310, 311
Ancillaries	154-213
Andersen Cascade/Impactor (ACI)	90-95, 241
Andersen Fast Screening Impactor (FSA)	263-264
Antistatic Grounding Kit	200
APSD	13, 78-153
	30, 298, 300
Aqueous Droplet Inhalers 6, 10, 58	
Auto Flow Control	02, 130 133
279, 282-283, 284, 288, 2	90, 291, 292
	80, 296-300
Automated Drug Recovery	
279, 280, 281, 294, 29	
	79, 282-289
Automated Flow Control	172-183
Automated Nasal Testing System 26, 28, 29, 64,	127 1/12 227
	79, 282-289
	•
Automated Snake and Fire 2 Automation	278-309
Automation (MDI)	282-293
Automation (NGI)	296-309

Book, Reference	317
Breath Actuation	
Controller BAC 100i	176-179
Breath-Actuated MDIs	7, 20, 21, 104, 172
Breathing	
Simulators	25, 154, 156-171
Bounce, Particle	96, 104, 118, 296
BP Content Uniformity Apparatus	

Caps, Rinsing	303,305	
Carrying/Wash Rack ACI	94,308	
Carrying/Wash		
Rack, NGI	88, 89, 308, 309	
Cascade Impactors	80-83, 84-103	
Cascade Impactor		
Preparation	280, 296, 300	
CDER	12, 16, 17	
CDRH	12	
Chinese Pharmacopoeia (Ch	P) 16,17	
Cleaning Systems, ACI/NGI	306-309	
Coarse Particle Mass (CPM)	262, 264, 266	
Coating, Cups		
104, 118, 122, 280, 296-297, 300		
Cold Freon® Effect	255-259	
Collection Tubes for DPIs	22-23	
Collection Tubes for MDIs	21	
Connectors, Quick-Release	207	
Conversion Kits, 60 & 90 L/m	nin 95	
Cooler, NGI	202-203	
Critical Quality Attributes (CQA	As) 15, 18, 78, 224	
Critical (Sonic) Flow	174, 180, 191	
Cup Coater	296-297	
Cut-Off Diameter		
86, 92, 98, 101, 216, 263, 265, 266		
Cup Tray Storage Cabinet, NO	299, 301	

Data Analysis Software (Inhalytix) 214-220

D

DecaVertus [®] III	290-293
Delivered Dose Sampling Apparatus	for:
- Nebulisers with a Mouthpiece	52-54

repulsers with a Mouthpiece	32 3-
- Nebulisers with a Facemask	250-25
- Spacers & VHCs	40-43
Delivered Dose Uniformity (DDU)	18-76

Design 320 Device Misuse 276 Device Resistance 46, 92, 118, 172-173 Device Robustness 276-277 Dissolution Cup 240, 241

Dissolution Inhaled 238-243
Dissolution Testing 298-299
- Dose Collection 239-241

Dosage Unit Sampling Apparatus
(DUSA) for DPIs

Dosage Unit Sampling Apparatus

21

26

262, 265

(DUSA) for MDIs

Dosage Unit Sampling Apparatus
(DUSA) for Nasal Sprays, Aerosols
and Powders

Drug Delivery Devices 6-11

Drug Delivery, PES 276

Drug Dissolution 298 - 299, 300 - 301

Drug Losses, Inter-Stage 85, 96, 98
Drug Recovery 298, 300

 Drying Oven, ACI
 307,309

 Dry Powder Inhalers (DPIs)
 8

 DUSA Shaker™
 279,294-295

E

Effective Cut-off Diameter (ECD) 92 Efficient Data Analysis (EDA) 261 **Electrostatic Effects** 200 Electrostatic Eliminator 200 Emitted Dose 18, 238 Entire Contents (DDU) 19, 20, 33, 41, 47 **Enviromental Control Chamber** 196-200 **Enviromental Variability** 196, 198 European Medicines 12, 14, 17 Agency (EMA) European Pharmaceutical 16, 17, 203, 313 Aerosol Group (EPAG) European Pharmacopoeia (Ph.Eur.) 15 87, 89, 217, 266 External Filter Holder, NGI

F

•	
Facemask-based products	225, 244-253
Facemask Stand for Nebulisers (FMS)	250-253
Facemask Test Apparatus for Spacers and VHCs (FMA)	246-249
Fast Screening Andersen (FSA)	263-264
Fast Screening Impactor (FSI)	266-267
Fast Screening Impactors	263-267
Fast Screening NGI (rNGI)	265
Flow Rate Sensor FRS	185, 186

Extra-fine Particle Mass (EPM)

F

Fine Particle Dose (FPD) 78, 82, 100, 141, 147, 218, 223 Fine Particle Mass (FPM) 78, 262, 264, 266 Flow Controllers: - TPK-100i 180-183 - TPK 100i-R 180-183 - BAC 100i 176-179 - BAC 100i-R 176-179 Flow Rate Measurment 184-187 Flow Rate 173, 184, 185 Flow Rate Sensor 185,186 Flow Rate Stability 174 Flow Resistance 22, 46, 48, 173, 224, 265 Fluticasone Propionate Aerosols - Testing 271, 273 Fluticasone Propionate Powders - Testing 271, 272, 273

G

Food and Drug Administration (FDA)

Generic Drug Development 270-275
Glass Expansion Chambers 208-210
Glass Impinger 100-102
Glass Twin Impinger 100-102
Gentle Rocker 298-299
Geometric Standard Deviation (GSD) 82, 84
Global Harmonisation Task Force (GHTF) 14
Gravimetric Cup, NGI 87, 89

Н

Relative Humidity, EnviroMate $^{\text{TM}}$ 196, 197, 198

I

ICH Guidelines	14
Idealised Nasal	230
Idealised Throat	228-229
Impactor Choice	103
Impactor Cleaning System	306-309
Impactor Performance Testing	304-307
Impactor Preperation	296-300
Impactor Qualification	312-315
Impactor Sized Mass (ISM)	226, 228, 261
Impinger, Glass, Twin	100-102
Inhaled Dissolution	238-243
Inhaled Drug Products	6-11
Inhaler Misuse	276-277
Inhaler Testing Academy®	321
Inhaler Testing Software (Inhaly	tix®) 214-221
Inhaler Testing Workstation (ITV	V) 204-207
Inhalytix Software	214-221
Inlet, Mixing	227, 234
Inspiration Volumes	173, 174

- 1

Installation Qualifica	ation 316, 317-318
Internal Filter Holde	r, NGI 87, 89
International Pharm Consortium on Regu Science (IPAC-RS)	
International Standa Organisation (ISO)	ards 2,14,16,1 7
Inter-Stage Losses	85, 96, 100
<i>In vitro - in vivo</i> correlation	14, 81, 156, 157, 222-225
In-Use Margin	313
IPAC-RS	16,17
IQ/OQ Documentati	on 317
ISO 20072	16, 17
ISO 27427:2013	17, 158, 162, 167
ISO 9001:2015 Quali	ty
Management Syster	n 2
ITW	204-207
IVIVC	14, 81, 156, 157, 222-225

J

12, 13, 14, 16, 17

Japanese Pharmacopoeia (JP)	16, 17
Jet Nebulisers	9

L

Label Printer	179, 183, 285,	288, 291, 305
Large Particle Ma	ass (LPM)	261, 262
Leak Testing	83, 176	180, 284, 315
Leak Test Kit		315
Losses, Inter-Sta	ge	85, 96, 100
Losses, Wall		84

М

Mass Balance	84
MDI Actuation Sensor	158, 179, 183
MDI Angle of Rotation	288,292
MDI Automation	282-293
MDI Shake and Fire	279, 282-293
Measurement, Flow Rate	184-187
Mensuration (impactor)	83, 312-315
Mesh Nebulisers	9
Meter, Static	200
Metered-Dose Inhalers (MDIs)	7, 32-44, 104-116
Metered Nasal Spray Pump	10
MHLW (Japan)	12,14
Mitutoyo QV404 Vision	
Inspection System	313
Mixing Inlet	227, 234
Mouthpiece Adapters	211-213
Multi-Stage Liquid	06.00
Impinger (MSLI)	96-98

N

Nasal Aerosols	10, 68-71, 141-146
Nasal Delivery Systems	10
Nasal, Idealised	230
Nasal Powders	10, 72-76, 147-152
Nasal Spray Dose Collector	27-29
Nasal Spray Modification,	
Twin Impinger	100, 10
Nebulisers	9, 52-56, 124-129
Nebulisers, Delivered Dose	52-56
Nebulisers, Particle Size	124-129
Next Generation Impactor (NGI) 84-8 9
NGI	84-89
NGI +	85, 89
NGI Carrying/Wash Rack	88, 308
NGI Cooler™	202-203
NGI Cup Coater	296-297
NGI Cup Tray Storage Cabin	et 30
NGI Gentle Rocker	298-299
NGI Gravimetric Cup	87, 89
NGI Leak Tester	318
NGI Pinning Kit	315
NGI Pinning Service	315
NGI Plus	85, 89
NGI, Reduced (rNGI)	265
NGI Sample Preparation Un	it 280, 302-30 5
Nosepiece Adapters	211, 213
NSDC	27, 28, 29

0

OINDPs	6-11	
Operation Qualification	312	
Organisations	12-17	

Ρ

Institute (PQRI)

P1 Measurement Adapter	120
Particle Bounce	96, 104, 118, 296
Particle Size	78-83
Patient Breath Profile	156, 157, 158-171
Patient Exhalation Simulator	; PES 276-277
PES	276-277
Pharmacopeial Forum	15
Pharmeuropa	15
Ph.Eur.	15
Pinning Kits	315
Pinning Service	315
Plume Temperature Testing	258-259
Powder-Based Nasal Devices	10, 72-76, 147-152
Pressure Drop	
22, 42, 46, 72	2, 92, 118, 120, 173
Process Analytical	
Technologies (PAT)	14
Product Quality Research	

16

Р

Propellant-Based Nasal Aerosols 6, 10
PTT 1000 Plume Temperature Tester 258-259
Pumps, Vacuum 155, 188-192
PVC Tubing 207

Q

Qualification	310	
Qualification, Servicing & Training	310-321	
Qualification Documentation IQ/OQ	317	
Qualification Tool Kit	318	
Quality by Design (QbD)		
2, 13, 14, 156, 224, 260, 261		
Quality Control Checks	311	
Quick Clamp 91	1, 93, 202	
Quick-Release Connectors	207	

R

Reduced NGI (rNGI) 265 Re-entrainment 96, 104, 118, 296 Reference Book 317 **Regulatory Bodies** 12-17 Repetitive Strain Injury 278,302 Resistance to Flow 22, 46, 48, 173, 224, 265 Reverse Valved Holding Chambers 40 88, 89, 94, 95, 273, 303, 305 Rinsing Caps **Robust Drug Delivery**

s

Salmeterol Aerosols, Testing 270, 271-273 Salmeterol Powders, Testing 270, 271-273 Sample Preparation Unit 281, 302-305 Servicing (Qualification) 310-316 279, 282-293 Shake and Fire Shaker, DUSA 279, 294-295 25, 154, 156-171 Simulators, Breathing Small Particle Mass (SPM) 261, 262 Soft Mist Inhalers 10 Software (Inhalytix®) 214-211 Sonic Flow 23, 174, 180, 189, 191 Spacers 7, 40-41, 111-116 256-257 Spray Force Tester SFT 1000 Spray Pattern 255 SPU Sample Preparation Unit 281, 302-305 Stage Cleaning 83,306-309 Stage Mensuration 83, 312-315 Static Meter 200 Storage Cabinet, NGI Cup Tray 299, 301 Suction Aspirator 307, 309 System Suitability Tests 311-321 System Qualification 312-318

T.

Target Delivered Dose (TDD)	33, 47
Temperature, $EnviroMate^{TM}$	196, 197, 198
Throats, Idealised	228-229
Total Active Substance Delivered (Nebulisers)	52, 53
TPK 100i Flow Controller	180-183
Training	321
Tubing	207
Twin Impinger, Glass	100-102

ш

Ultrasonic Cleaning Bath, ACI	307, 309
Ultrasonic Nebulisers	6, 9
United States Pharmacopeia (USP)	15, 17

٧

Vacuum Pumps	155, 188-192
Validation	310
Valved Holding Chambers (VHCs)	7, 40-44, 111-116
Verification	310
Vertus® III	283, 284-289
Vertus III+	283, 284-289
Vision Inspection System,	
Mitutoyo QV 404	312
Volume, Inspiration	173, 174

W

Wall losses	84
Waste Shot Collector	28, 258, 279
Wash Rack, ACI	92-309
Wash Rack, NGI	88,308
Watchglass/PTFF Assembly	239, 241

-

21	CFR Part 11	157, 215, 220, 28	34, 290
60	L/min Conversion	Kit	92,95
90	L/min Conversion	Kit	92, 95

Notes		

Inh	naler Testing

Ir	haler Testing

Inhaler Testing			



COPLEY

Copley Scientific Limited

Colwick Quays Business Park, Road No.2 Nottingham, NG4 2JY United Kingdom

- & +44 (0)115 961 6229
- ı +44 (0)115 961 7637
- ⊕ copleyscientific.com

